

Cite this: *Nanoscale Adv.*, 2023, 5, 3664

# Lanthanide doped nanoparticles for reliable and precise luminescence nanothermometry in the third biological window

Ana C. C. Soares, <sup>a</sup> Tasso O. Sales, <sup>a</sup> Erving C. Ximendes, <sup>bc</sup> Daniel Jaque <sup>\*bc</sup> and Carlos Jacinto <sup>\*a</sup>

In recent years, infrared emitting luminescent nanothermometers have attracted significant attention because their potential for the development of new diagnosis and therapy procedures. Despite their promising applications, concerns have been raised about their reliability due to the spectral distortions induced by tissues that are present even in the commonly used second biological window (1000–1370 nm). In this work, we present an innovative solution to this issue by demonstrating the effectiveness of shifting the operation range of these nanothermometers to the third biological window (1550–1850 nm). Through experimental evidence using ytterbium, erbium, and thulium tri-doped CaF<sub>2</sub> nanoparticles, we demonstrate that luminescence spectra acquired in the third biological window are minimally distorted by the presence of tissue, opening the way to reliable luminescence thermometry. In addition, advanced analysis (singular value decomposition) of emission spectra allows sub-degree thermal uncertainties to be achieved.

Received 23rd December 2022  
Accepted 12th May 2023

DOI: 10.1039/d2na00941b

rsc.li/nanoscale-advances

## Introduction

Temperature governs and controls physiological processes.<sup>1,2</sup> In recent years, different nanomaterials have been proposed as thermal reporters,<sup>3–5</sup> and, despite their applicability in a diversity of fields, they have found some limitations when being applied in biomedicine. For instance, thermal measurements in living systems should be achieved in a remote way in order to interfere minimally with the normal behaviour of the system. In order to achieve such remote thermal monitoring luminescent nanothermometers (LNThs) have emerged as a potential solution.<sup>5,6</sup> LNThs are luminescent nanoparticles whose emission is temperature dependent so that thermal readout can be achieved from an adequate analysis of their emission. Thermal readout can be achieved from the analysis of one or more spectroscopic parameters such as the peak position, linewidth, intensity ratio or polarisation state.<sup>7–9</sup>

LNThs have been able to provide thermal readouts inside cells and living animals.<sup>7,8</sup> Indeed, LNThs have already been used for sub-degree thermal reading in tissues, for early detection of incipient diseases such as ischemia and tumours,

for tissue diagnosis and for providing thermal control in photothermal therapies at the *in vivo* level.<sup>7,9,10</sup> In spite of this recent success, the *in vivo* application of luminescence thermometry needs to overcome some inherent limitations that are related to the optical extinction of tissues.

On one hand, the strong attenuation of visible light in tissues makes it difficult to access internal organs optically.<sup>11</sup> This limits the potential of visible-emitting LNThs. Indeed, *in vivo* luminescence thermometry requires working in the near-infrared (NIR) spectral range due to the lower absorption and scattering of light in this spectral range. Traditionally, NIR is divided into three biological windows (BWs), where optical extinction is minimum: the first (BW-I) extending from 650 up to 950 nm, the second (BW-II) ranging from 1000 up to 1370 nm and the third (BW-III) from 1550 up to 1870 nm.<sup>12,13</sup> BW-I has the drawback of interference with autofluorescence;<sup>14,15</sup> however, the absence of water absorption in this spectral range makes it ideal for laser excitation.<sup>16,17</sup> Operating within BW-II results in an improved signal-to-noise ratio, a reduced (but not negligible) autofluorescence background and availability of cost-effective fluorescence cameras.<sup>18,19</sup> All these together have made BW-III very popular for pre-clinical fluorescence imaging. Finally, BW-III offers advantages such as the complete absence of autofluorescence and minimum tissue extinction but it is not extensively explored due to the absence of accessible fluorescence cameras working in this spectral range.<sup>13,20,21</sup> As a consequence of all these pros and cons most of the work on *in vivo* fluorescence imaging and thermometry has been traditionally developed by using probes working in BW-I,<sup>22–24</sup> although

<sup>a</sup>Group of Nano-Photonics and Imaging, Instituto de Física, Universidade Federal de Alagoas, 57072-900, Maceió-AL, Brazil. E-mail: cjacinto@fis.ufal.br

<sup>b</sup>Nanomaterials for Bioimaging Group (nanoBIG), Departamento de Física de Materiales, Facultad de Ciencias, Universidad Autónoma de Madrid, Madrid 28049, Spain. E-mail: daniel.jaque@uam.es

<sup>c</sup>Nanomaterials for Bioimaging Group (nanoBIG), Instituto Ramón y Cajal de Investigación Sanitaria, Hospital Ramón y Cajal, Madrid 28034, Spain



during the last few years the use of optical probes working in BW-II has become more and more popular. A reduced number of studies have demonstrated the suitability of optical probes working in BW-III for high penetration *in vivo* imaging.<sup>21,25–30</sup>

Tissue-induced light extinction not only affects the depth achievable by LNTs but also impacts their reliability.<sup>31–39</sup> When a given tissue has a flat extinction spectrum, meaning it lacks any significant peaks or valleys within a certain wavelength range, the light that passes through it will retain its spectral shape (it will retain the relative intensities between emission lines). Hence, while there will be an overall reduction in intensity, the shape of the spectrum will remain the same. This property becomes particularly useful in luminescence thermometry. To grasp this concept, let's consider a scenario where a LNT is located within a medium that has an extinction coefficient with thermal and wavelength dependencies described using  $\mu(\lambda, T)$ . According to the Beer–Lambert law, the luminescence intensity detected at wavelength  $\lambda$  after passing through the medium at a temperature  $T$  is given by:

$$I_{\text{det}}(\lambda, T) = I_0(\lambda, T)e^{-\mu(\lambda, T)L} \quad (1)$$

Thus, the ratio between the detected intensities at  $\lambda_i$  and  $\lambda_f$  is given by:

$$R_{\text{det}}(\lambda_i, \lambda_f, T) = \frac{I_0(\lambda_i, T)e^{-\mu(\lambda_i, T)L}}{I_0(\lambda_f, T)e^{-\mu(\lambda_f, T)L}} = R_{\text{emi}}(\lambda_i, \lambda_f, T) \frac{e^{-\mu(\lambda_i, T)L}}{e^{-\mu(\lambda_f, T)L}} \quad (2)$$

which means that if  $\lambda_i$  and  $\lambda_f$  are within an interval in which the extinction coefficient of the surrounding medium varies with the wavelength (range  $[\lambda_1, \lambda_2]$  schematically represented in Fig. 1), it is mandatory to know the value of  $\mu$  at both  $\lambda_i$  and  $\lambda_f$  to properly convert a calibration made only with the ratio between the emitted intensities ( $R_{\text{emi}}$ ) into the one that is valid even under the presence of the propagating medium. This limits the

potential application of luminescence thermometry as it requires the precise knowledge of how the extinction of a medium depends on both wavelength and temperature. In the case of tissues, this is far from being simple.

However, if  $\lambda_i$  and  $\lambda_f$  are within the spectral range in which the extinction coefficient of the surrounding medium is flat (range  $[\lambda_2, \lambda_3]$  schematically represented in Fig. 1), the detected ratio becomes:

$$R_{\text{det}}(\lambda_i, \lambda_f, T) = \frac{I_0(\lambda_i, T)e^{-\mu(\lambda_i, T)L}}{I_0(\lambda_f, T)e^{-\mu(\lambda_f, T)L}} = R_{\text{emi}}(\lambda_i, \lambda_f, T) \quad (3)$$

*i.e.*, the insertion of a luminescent thermometer into the medium does not alter the calibration of the thermometer based on the ratio between emitted intensities. Thus, if we could find a spectral region where that was generally valid for biological tissues, the issue of tissue-induced spectral distortions would be avoided and the thermal readout provided by ratiometric nanothermometers would also become reliable.

Luckily, when analysing the extinction coefficient,  $\mu$ , of representative human tissues (Fig. 2(a), as obtained from ref. 40) in the 600–2250 nm spectral range, it is found that in the 1590–1860 nm region,  $\mu$  is approximately constant. And this range is found within what is generally described as BW-III.<sup>12</sup> To further demonstrate this, Fig. 2(b) shows the derivative of the extinction coefficient as calculated from Fig. 2(a) showing how this remains close to zero in a great portion of BW-III. As explained above, this means that thermometry based on the luminescence intensity ratio can be made reliable. Despite that there are already some studies reporting on LNTs operating in




Fig. 1 The light that propagates through a tissue is affected by absorption and scattering processes that result in an effective reduction of its intensity. This reduction is best described by what is usually called the extinction (or attenuation) spectrum. It is observed that, for thermometry based on fluorescence intensity ratios, the performance is improved when the ratio is calculated between intensities found within wavelength ranges where the attenuation spectrum is flat.




Fig. 2 (a) Extinction coefficient of human (prostate) tissue. The spectral extension of the three biological windows is indicated. Data obtained from ref. 40. (b) Derivative of the absorption coefficient of human tissue with respect to the wavelength as calculated from (a).



BW-III,<sup>25,29,41–44</sup> none of them have explored the possible presence/absence of tissue-induced spectral distortions.

In this work we have investigated the use of lanthanide doped nanoparticles for thermal sensing in BW-III. In particular we have evaluated the performance of  $\text{Yb}^{3+}$ ,  $\text{Er}^{3+}$  and  $\text{Tm}^{3+}$  doped  $\text{CaF}_2$  nanoparticles (NPs) as LNTHs in BW-III.  $\text{CaF}_2$  NPs have been selected as host materials due to their low effective phonon energy ( $280\text{ cm}^{-1}$  that ensures low nonradiative decay probabilities), broad transparency window (from UV to IR), simple synthesis route, and low toxicity in biological systems.<sup>45–48</sup> Experiments have been conducted to demonstrate the suitability of  $\text{Tm}^{3+}$  emissions in BW-III for reliable (distortion-free) luminescence thermometry. In addition, we have also explored the potential use of advanced analysis to improve the thermal uncertainties of the thermal readouts provided by  $\text{CaF}_2$  NPs.

## Experiments

### Synthesis of nanoparticles

For the synthesis of the nanoparticle ( $\text{CaF}_2:10\text{Yb}^{3+}/6\text{Er}^{3+}/1\text{Tm}^{3+}$ ), an amount of 0.83 mmol of  $\text{CaCl}_2$ , 0.1 mmol of  $\text{YbCl}_3$ , 0.06 mmol of  $\text{ErCl}_3$  and 0.01 mmol of  $\text{TmCl}_3$  were added in a 100 ml round bottom single neck flask containing 20 ml of Milli-Q water and the solution was heated at  $90\text{ }^\circ\text{C}$ , approximately, under constant stirring. Subsequently, 2 mmol of a  $\text{NH}_4\text{F}$  solution was injected into the mixture dropwise and kept at  $90\text{ }^\circ\text{C}$  for 1 h under vigorous stirring. After this time and after being cooled down to room temperature naturally, white precipitates were collected by centrifugation and washed with Milli-Q water three times. The resulting product was dried at  $60\text{ }^\circ\text{C}$  in an ambient atmosphere for 48 h and finally heat treated at  $500\text{ }^\circ\text{C}$  for 3 h. We did not simply choose these rare-earth ion concentrations, but a detailed study was carried out in which we concluded that this sample is the best in terms of luminescence relative to the investigated bands.

### Experimental apparatus

The luminescence measurements were performed by using a fluorometer (NanoLog, Horiba) coupled to an R928P or R5509-73 photomultiplier tube or a liquid nitrogen refrigerated solid state extended InGaAs detector. The  $10\text{Yb}^{3+}/6\text{Er}^{3+}/1\text{Tm}^{3+}$  co-doped  $\text{CaF}_2$  NPs were optically excited using a Lumics Laser operating at  $940\text{ nm}$ , which was focused on the sample. For thermal analysis, Peltier plates connected to a voltage source were used as heating systems, and the electrical voltage was manually adjusted at the source to work within the physiological temperature range. The measurements for biological applications were made using *ex vivo* chicken breast tissue. Tissue-induced spectral distortions were investigated by placing the chicken breast in front of the cuvette containing the NPs; the fluorescence signal was collected using an optical fiber and dispersed to a monochromator, as shown in Fig. 3(a). The luminescence spectra/images were collected using a confocal fluorescence microscope (LabRam Evolution UV-VIS-NIR) coupled with a Symphony II detector and focused with a  $20\times$  long working distance microscope. In this case, we placed the




Fig. 3 Schematic representations of the experimental setup used for spectral distortions induced by tissue. (a) Representation of *ex vivo* tissue disposition and used nanoparticles. (b) Illustration of the experimental system used for spectral distortion of the Yb–Er–Tm  $\text{CaF}_2$  NPs induced by *ex vivo* tissue.

biological tissue under the mobile platform and scanned it, as illustrated in Fig. 3(b).

## Results and discussion

The inset of Fig. 4(a) shows a characteristic transmission electron microscopy (TEM) image of the  $\text{CaF}_2:\text{Yb}/\text{Tm}/\text{Er}$  NPs utilized in this work. The size histogram that has been calculated from the statistical analysis of the TEM image is also included in Fig. 4(a), from which fitting to lognormal distributions resulted in an average diameter of  $20.8 \pm 0.3\text{ nm}$  with a width of  $7.5\text{ nm}$  ( $R$ -square = 0.93 was obtained). A typical XRD




Fig. 4 (a) Size distribution for the  $\text{CaF}_2:\text{Yb}^{3+}/\text{Tm}^{3+}/\text{Er}^{3+}$  nanoparticles. The inset shows a TEM image of these nanoparticles. The bar size is  $60\text{ nm}$ . (b) X-ray powder diffraction patterns of  $\text{CaF}_2:\text{Yb}^{3+}/\text{Tm}^{3+}/\text{Er}^{3+}$  nanoparticles (top) and the standard XRD pattern (bottom).



pattern of  $\text{CaF}_2\text{:Yb/Tm/Er}$  NPs is shown in Fig. 4(b)-top. The sharp diffraction peaks in the pattern in Fig. 4(b)-bottom can be indexed as pure cubic-phase  $\text{CaF}_2$ , which is in good agreement with the standard XRD pattern JCPDS 35-0816 (Fig. 4(b)-bottom). No other impurity peaks were identified, which indicates formation of pure cubic-phase  $\text{CaF}_2$  structures with a highly crystalline nature. Compared with the standard peaks, all the diffraction peaks of pure cubic-phase  $\text{CaF}_2$  are barely shifted, because the ionic radius of  $\text{Ca}^{2+}$  is close to the dopant ionic radius of the ions  $\text{Yb}^{3+}$  and  $\text{Er}^{3+}$ .

At room temperature and under 940 nm excitation, the Yb–Er–Tm triply doped  $\text{CaF}_2$  NPs present four characteristic emission bands at around 660, 1230, 1550 and 1800 nm, as shown in Fig. 5(a). These luminescence bands lie within the three BWs as defined previously. In addition, the laser excitation at 940 nm (in BW-I) is sufficiently far from the water absorption peak ( $\sim 980$  nm) so that heating is avoided. The  $\text{Yb}^{3+}$  ions act as sensitizers absorbing the 940 nm radiation through the  ${}^2\text{F}_{7/2} \rightarrow {}^2\text{F}_{5/2}$  transition and by means of energy transfer (ET) both  $\text{Er}^{3+}$  ( ${}^4\text{I}_{11/2}$ ) and  $\text{Tm}^{3+}$  ( ${}^3\text{H}_5$ ) ions are excited. Multiphonon non-radiative decays populate the  ${}^4\text{I}_{13/2}$  energy level of  $\text{Er}^{3+}$  ions from which the emission at around 1550 nm is produced ( ${}^4\text{I}_{13/2} \rightarrow {}^4\text{I}_{15/2}$  transition). A second photon excites the  $\text{Er}^{3+}$  ions to the  ${}^4\text{S}_{3/2}$  level through a second ET from the  $\text{Yb}^{3+}$  to  $\text{Er}^{3+}$  ions, from which multiphonon decay occurs to the  ${}^4\text{F}_{9/2}$  level, promoting the emission at around 660 nm ( ${}^4\text{F}_{9/2} \rightarrow {}^4\text{I}_{15/2}$  transition). The  ${}^3\text{H}_5 \rightarrow {}^3\text{H}_6$  radiative decay of  $\text{Tm}^{3+}$  ions gives rise to the emission around 1230 nm. Multiphonon decay from the  ${}^3\text{H}_5$  level

populates the  ${}^3\text{F}_4$  state, from which the emission at around 1800 nm is generated ( ${}^3\text{F}_4 \rightarrow {}^3\text{H}_6$  transition). Excitation of  $\text{Tm}^{3+}$  ions from the  ${}^3\text{H}_5$  level to the  ${}^3\text{F}_2$  state occurs thanks to energy transfer from  $\text{Er}^{3+}$  ions. Nonradiative transitions from the  ${}^3\text{F}_2$  state populate the  ${}^3\text{H}_4$  energy level from which the emission at around 800 nm is generated (see Fig. 5(b)).

For the sake of evaluating the performance of  $\text{CaF}_2\text{:Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$  NPs for subtissue applications, we have investigated the spectral distortions induced in the different emission bands of  $\text{CaF}_2\text{:Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$  NPs by chicken breast tissues with thicknesses of 1.0 and 2.0 mm (red and blue curves in Fig. 6(a), (c) and (e)). It is possible to notice that the emission bands lying within BW-I and BW-II present larger distortions when compared to those induced in the III-BW. Indeed, the  $\text{Tm}^{3+}$  emission at around 1700 nm practically maintains its spectral shape even in the presence of a 2 mm thick tissue. To quantify the tissue-induced spectral distortions, the percentage difference ( $\sigma$ ) in the spectral shape was evaluated according to a parameter previously defined by Ximendes *et al.*<sup>49</sup> as:

$$\sigma_{\text{tissue}} = \int |I_{\text{tissue}}^{\text{norm}}(\lambda) - I_{\text{nonatt}}^{\text{norm}}(\lambda)| d\lambda \quad (4)$$

where  $I_{\text{tissue}}$  and  $I_{\text{nonatt}}$  denote respectively the spectra obtained with and without tissue, and the index norm indicates that the spectra are normalized to the area (*i.e.*,  $I^{\text{norm}}(\lambda) = \frac{I(\lambda)}{\int I(\lambda) d\lambda}$ ). The results included in Fig. 6(b), (d) and (f) reveal that the tissue-induced percentage change in the shape of the spectra in BW-I and BW-II are 20.5% and 21.6%, respectively (for a tissue thickness of 2 mm). On the other hand, the tissue-induced spectral distortion obtained in BW-III was significantly




Fig. 5 (a) Emission spectrum of  $\text{Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$  tri-doped  $\text{CaF}_2$  NPs under 940 nm excitation at room temperature in the I-BW (red), II-BW (left peak of the green spectrum) and III-BW (right peak of the green spectrum). (b) Simplified energy scheme of the  $\text{Yb}^{3+}$ ,  $\text{Er}^{3+}$  and  $\text{Tm}^{3+}$  ions, showing excitation and emission decays (full lines), nonradiative decays (wavy lines), and the energy transfer process (curved lines).




Fig. 6 Emission spectra of  $\text{Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$  tri-doped  $\text{CaF}_2$  NPs under 940 nm excitation at room temperature without and with 1.0 and 2.0 mm of tissue in the (a) III-BW, (c) II-BW, and (e) I-BW. The percentage difference in the spectral shape for tissues' thickness with 1.0 mm (red) and 2.0 mm (blue) in the (b) III-BW, (d) II-BW, and (f) I-BW.



reduced down to 4%. When it comes to subcutaneous sensing based on ratiometric luminescence thermometry, the reduction of  $\sigma$  is of utmost importance, as it has been explained previously. Hence, we will hereafter focus on the emission of our  $\text{CaF}_2:\text{Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$  NPs lying within BW-III for ratiometric thermal sensing.

Fig. 7(a) shows the normalized (to the intensity at 1812 nm) emission spectra of  $\text{CaF}_2:\text{Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$  NPs as obtained at 33 and 57 °C in BW-III. A relative reduction in the emission of  $\text{Tm}^{3+}$  ions at around 1800 nm and an increment in the emission of  $\text{Er}^{3+}$  at around 1550 nm are observed. Due to the different thermal quenching of  $\text{Er}^{3+}$  and  $\text{Tm}^{3+}$  ions and to the population distribution between the  $\text{Tm}^{3+}$  levels, the shape of the emission spectra changes with temperature. This opens the possibility of using the BW-III band of  $\text{CaF}_2:\text{Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$  NPs for thermal sensing. As it has been discussed previously, reliable ratiometric thermal sensing would be possible only by considering the emission lines that are within the spectral range in which the extinction coefficient is flat. Then, according to the data included in Fig. 2, reliable thermal sensing can be achieved by using the temperature dependence of the intensities of the emission lines lying within the 1590–1860 nm range. This means that for thermal sensing purposes we can use the temperature dependence of the emission lines at 1618, 1675, 1725 and 1812 nm normalized to integrated emitted intensity ( $I_1$ ,  $I_2$ ,  $I_3$  and  $I_4$ , respectively). The temperature variations of these intensities (normalized to the integrated emission in the whole band) are shown in Fig. 7(b). The traditional approach for ratiometric thermal sensing consists of recording the temperature variation of these normalized intensities so that a proper analysis of emission shape can give the thermal readout. Alternatively, as

recently demonstrated, multiple linear regression can also be applied to obtain a larger thermal sensitivity.<sup>50</sup> Nevertheless, computing the thermal dependence of each of these parameters and building a regression model based on them can be a burdensome task. As a matter of fact, it can be a statistically complex process due to the possibility of the dependent variables being collinear. Thus, to avoid these issues and make the analysis simpler, we applied single value decomposition (SVD) to the dataset containing the temperature dependence of relative emitted intensities at these wavelengths. SVD is a computational method that falls into the category known as dimensionality reduction (DR), which, simply put,<sup>51,52</sup> is a way of representing a dataset in a space of lower dimensionality. It is a matrix factorization technique widely used in scientific research and a powerful tool that can extract important information from large datasets, such as patterns and relationships between variables. SVD decomposes a matrix dataset into three other matrices: a left singular matrix, a diagonal matrix, and a right singular matrix. The diagonal matrix contains the singular values of the original matrix, which represent the relative importance of each column and row in the original matrix. These singular values can be used to reduce the dimensionality of the original matrix, allowing for easier computation and visualization of data. It has a wide range of applications in various fields, such as computer vision, natural language processing, and recommendation systems. It has several applications in spectroscopy, particularly in processing and analysing spectroscopic data. One of the main applications is noise reduction. Spectroscopic data often contain a significant amount of noise, which can affect the accuracy and reliability of the analysis. SVD can be used to remove noise from the data by identifying the principal components of the data and eliminating the components that are due to noise. This results in a cleaner and more accurate spectrum, which can be used for further analysis. For our purposes, SVD is a very handy technique since visualizing the simultaneous dependence of four different parameters is not so straightforward. Additionally, DR techniques have recently been shown to best summarize the calibration of a luminescent thermometer and to provide better precision in thermal readouts.<sup>33</sup>

When applying SVD to the dataset defined using all the relative intensities, more than 95% of the variance (computed from the relative contribution of the eigenvalues Fig. 7(b)) can be explained using a single vector (hereafter  $\text{RSV}_1$ ). Both the values of the level of variance explained and the coordinates of  $\text{RSV}_1$  are found through the factorization of the data matrix. The thermal dependence of  $\text{RSV}_1$  is included in Fig. 7(b) and represents a very linear trend. When estimating the thermal resolution achieved by the readouts given by the four relative intensities and the readout provided by  $\text{RSV}_1$ , it has been shown that the uncertainty is minimized with the latter although the performance improvement with respect to that achieved by analysing only  $I_4$  is reduced (bar graph included in Fig. 7(b)).




Fig. 7 (a) Emission spectra of  $\text{Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$   $\text{CaF}_2$  NPs at 33 °C (blue) and 57 °C (red) under 940 nm excitation in the III-BW. (b) Thermal dependence of the different normalized intensities (as defined in the main text) of  $\text{RSV}_1$  and the thermal uncertainties obtained with the different parameters.

## Conclusions

In conclusion, we investigated the potential of a luminescent nanothermometer based on  $\text{Yb}^{3+}/\text{Er}^{3+}/\text{Tm}^{3+}$  tri-doped calcium



fluoride (CaF<sub>2</sub>) nanoparticles operating in three biological windows. It was shown that their emission in the third biological window was minimally affected by the presence of a biological tissue. Hence, it could be optimal for subcutaneous luminescence sensing. The thermal dependence of the luminescence of the nanoparticles was then studied and several thermometric parameters were identified. The application of singular value decomposition to the set of parameters provided a single parameter that could account for more than 95% of the variance in the calibration. Such a parameter, in turn, was shown to be linearly dependent on temperature and provided a minimum thermal uncertainty below 1 °C. In short, the use of BW-III is a great solution to the problem recently presented by the scientific community, recovering the reliability of luminescent thermometers for bio-applications and the tri-doped CaF<sub>2</sub> matrix with Yb<sup>3+</sup>, Tm<sup>3+</sup> and Er<sup>3+</sup> is a luminescent nanothermometer with high relative thermal sensitivity.

## Author contributions

A. C. C. Soares and T. O. Sales contributed to conceptualization, data curation, formal analysis, investigation, visualization, and writing the original draft. E. C. Ximendes was responsible for data curation, formal analysis, methodology, validation, and writing – review & editing. Daniel Jaque contributed to funding acquisition, methodology and writing – review & editing. C. Jacinto was responsible for project administration, resources, supervision, funding acquisition and writing – review & editing.

## Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

We acknowledge the financial support from Brazilian Agencies: FINEP (Financiadora de Estudos e Projetos) through the grants INFRAPESQ-11 and INFRAPESQ-12 in using the Fluorometer NanoLog (Horiba) and Microscope LabRAM HR Evolution (Horiba); CNPq (Conselho Nacional de Desenvolvimento Científico e Tecnológico) through the scholarship in Research Productivity 1C under No. 304967/2018-1 (C. J.) and Doctoral scholarship (A. C. C. S.); CAPES (Coordenação de Aperfeiçoamento de Pessoal de Nível Superior) by means of the PNP/D/CAPES postdoctoral scholarship (T. O. S.). This work was also financed by the Spanish Ministerio de Innovación y Ciencias under project NANONERV PID2019-106211RB-100. E. C. X. is grateful for a Juan de la Cierva Incorporación scholarship (IJC2020-045229-I).

## References

- W. J. Youden, Index for rating diagnostic tests, *Cancer*, 1950, **3**, 32–35.
- P. R. N. Childs, J. R. Greenwood and C. A. Long, Review of temperature measurement, *Rev. Sci. Instrum.*, 2000, **71**, 2959–2978.
- L. C. Kennedy, *et al.*, A New Era for Cancer Treatment: gold-Nanoparticle-Mediated Thermal Therapies, *Small*, 2011, **7**, 169–183.
- D. Jaque, *et al.*, Nanoparticles for photothermal therapies, *Nanoscale*, 2014, **6**, 9494–9530.
- B. del Rosal, *et al.*, Neodymium-Based Stoichiometric Ultrasmall Nanoparticles for Multifunctional Deep-Tissue Photothermal Therapy, *Adv. Opt. Mater.*, 2016, **4**, 782–789.
- D. Jaque and F. Vetrone, Luminescence nanothermometry, *Nanoscale*, 2012, **4**, 4301.
- E. Carrasco, *et al.*, Intratumoral Thermal Reading During Photo-Thermal Therapy by Multifunctional Fluorescent Nanoparticles, *Adv. Funct. Mater.*, 2015, **25**, 615–626.
- E. C. Ximendes, *et al.*, Unveiling *in vivo* Subcutaneous Thermal Dynamics by Infrared Luminescent Nanothermometers, *Nano Lett.*, 2016, **16**, 1695–1703.
- E. C. Ximendes, *et al.*, *In vivo* Ischemia Detection by Luminescent Nanothermometers, *Adv. Healthcare Mater.*, 2017, **6**, 1601195.
- H. D. A. Santos, *et al.*, *In vivo* Early Tumor Detection and Diagnosis by Infrared Luminescence Transient Nanothermometry, *Adv. Funct. Mater.*, 2018, **28**, 1803924.
- D. Jaque, *et al.*, Inorganic nanoparticles for optical bioimaging, *Adv. Opt. Photonics*, 2016, **8**, 1–103.
- A. M. Smith, M. C. Mancini and S. Nie, Second window for *in vivo* imaging, *Nat. Nanotechnol.*, 2009, **4**, 710–711.
- E. Hemmer, A. Benayas, F. Légaré and F. Vetrone, Exploiting the biological windows: current perspectives on fluorescent bioprobes emitting above 1000 nm, *Nanoscale Horiz.*, 2016, **1**, 168–184.
- B. del Rosal, I. Villa, D. Jaque and F. Sanz-Rodríguez, *In vivo* autofluorescence in the biological windows: the role of pigmentation, *J. Biophotonics*, 2016, **9**, 1059–1067.
- B. del Rosal, *et al.*, Overcoming Autofluorescence: long-Lifetime Infrared Nanoparticles for Time-Gated *in vivo* Imaging, *Adv. Mater.*, 2016, **28**, 10188–10193.
- I. Villa, *et al.*, 1.3 μm emitting SrF<sub>2</sub>:Nd<sup>3+</sup> nanoparticles for high contrast *in vivo* imaging in the second biological window, *Nano Res.*, 2015, **8**, 649–665.
- D. Kobat, *et al.*, Deep tissue multiphoton microscopy using longer wavelength excitation, *Opt. Express*, 2009, **17**, 13354.
- B. del Rosal, I. Villa, D. Jaque and F. Sanz-Rodríguez, *In vivo* autofluorescence in the biological windows: the role of pigmentation, *J. Biophotonics*, 2016, **9**, 1059–1067.
- A. F. Pereira, J. F. Silva, A. S. Gouveia-Neto and C. Jacinto, 1.319 μm excited thulium doped nanoparticles for subtissue thermal sensing with deep penetration and high contrast imaging, *Sens. Actuators, B*, 2017, **238**, 525–531.
- J. Xu, D. Murata, J. Ueda, B. Viana and S. Tanabe, Toward Rechargeable Persistent Luminescence for the First and Third Biological Windows *via* Persistent Energy Transfer and Electron Trap Redistribution, *Inorg. Chem.*, 2018, **57**, 5194–5203.
- E. C. Ximendes, *et al.*, *In vivo* Subcutaneous Thermal Video Recording by Supersensitive Infrared Nanothermometers, *Adv. Funct. Mater.*, 2017, **27**, 1702249.



- 22 M. Lin, *et al.*, Facile synthesis of mono-disperse sub-20 nm NaY(WO<sub>4</sub>)<sub>2</sub>:Er<sup>3+</sup>,Yb<sup>3+</sup> upconversion nanoparticles: a new choice for nanothermometry, *J. Mater. Chem. C*, 2019, 7, 2971–2977.
- 23 F. Vetrone, *et al.*, Temperature Sensing Using Fluorescent Nanothermometers, *ACS Nano*, 2010, 4, 3254–3258.
- 24 S. Zhou, *et al.*, Strategy for thermometry via Tm<sup>3+</sup>-doped NaYF<sub>4</sub> core-shell nanoparticles, *Opt. Lett.*, 2014, 39, 6687.
- 25 A. Skripka, *et al.*, Double rare-earth nanothermometer in aqueous media: opening the third optical transparency window to temperature sensing, *Nanoscale*, 2017, 9, 3079–3085.
- 26 A. Zhang, *et al.*, Simultaneous luminescence in I, II and III biological windows realized by using the energy transfer of Yb<sup>3+</sup> → Er<sup>3+</sup>/Ho<sup>3+</sup> → Cr<sup>3+</sup>, *Chem. Eng. J.*, 2019, 365, 400–404.
- 27 K. Maciejewska, A. Bednarkiewicz and L. Marciniak, The influence of the Er<sup>3+</sup> dopant concentration in LaPO<sub>4</sub>:Nd<sup>3+</sup>, Er<sup>3+</sup> on thermometric properties of ratiometric and kinetic-based luminescent thermometers operating in NIR II and NIR III optical windows, *Phys. B*, 2021, 620, 413247.
- 28 M. Kamimura, Recent Progress of Near-Infrared Fluorescence *in vivo* Bioimaging in the Second and Third Biological Window, *Anal. Sci.*, 2021, 37, 691–697.
- 29 J. Hu, *et al.*, Quantum Dots Emitting in the Third Biological Window as Bimodal Contrast Agents for Cardiovascular Imaging, *Adv. Funct. Mater.*, 2017, 27, 1703276.
- 30 C. Hazra, A. Skripka, S. J. L. Ribeiro and F. Vetrone, Erbium Single-Band Nanothermometry in the Third Biological Imaging Window: Potential and Limitations, *Adv. Opt. Mater.*, 2020, 8, 2001178.
- 31 M. Quintanilla, M. Henriksen-Lacey, C. Renero-Lecuna and L. M. Liz-Marzán, Challenges for optical nanothermometry in biological environments, *Chem. Soc. Rev.*, 2022, 51, 4223–4242.
- 32 V. B. Koman, *et al.*, A wavelength-induced frequency filtering method for fluorescent nanosensors *in vivo*, *Nat. Nanotechnol.*, 2022, 17, 643–652.
- 33 E. Ximendes, R. Marin, L. D. Carlos and D. Jaque, Less is more: dimensionality reduction as a general strategy for more precise luminescence thermometry, *Light: Sci. Appl.*, 2022, 11, 237.
- 34 F. E. Maturi, *et al.*, Going Above and Beyond: A Tenfold Gain in the Performance of Luminescence Thermometers Joining Multiparametric Sensing and Multiple Regression, *Laser Photonics Rev.*, 2021, 15, 2100301.
- 35 Y. Shen, J. Lifante, N. Fernández, D. Jaque and E. Ximendes, *In vivo* Spectral Distortions of Infrared Luminescent Nanothermometers Compromise Their Reliability, *ACS Nano*, 2020, 14, 4122–4133.
- 36 A. Bednarkiewicz, L. Marciniak, L. D. Carlos and D. Jaque, Standardizing luminescence nanothermometry for biomedical applications, *Nanoscale*, 2020, 12, 14405–14421.
- 37 Y. Shen, *et al.*, Ag<sub>2</sub>S Nanoheaters with Multi parameter Sensing for Reliable Thermal Feedback during *in vivo* Tumor Therapy, *Adv. Funct. Mater.*, 2020, 30, 2002730.
- 38 M. Tan, *et al.*, Accurate *in vivo* Nanothermometry through NIR-II Lanthanide Luminescence Lifetime, *Small*, 2020, 16, 2004118.
- 39 J. Zhou, B. del Rosal, D. Jaque, S. Uchiyama and D. Jin, Advances and challenges for fluorescence nanothermometry, *Nat. Methods*, 2020, 17, 967–980.
- 40 L. A. Sordillo, Y. Pu, S. Pratavieira, Y. Budansky and R. R. Alfano, Deep optical imaging of tissue using the second and third near-infrared spectral windows, *J. Biomed. Opt.*, 2014, 19, 056004.
- 41 M. Runowski, *et al.*, Sr<sub>2</sub>LuF<sub>7</sub>:Yb<sup>3+</sup>-Ho<sup>3+</sup>-Er<sup>3+</sup> Upconverting Nanoparticles as Luminescent Thermometers in the First, Second and Third Biological Windows, *ACS Appl. Nano Mater.*, 2020, 3, 6406–6415.
- 42 E. D. Onal and K. Guven, Plasmonic Photothermal Therapy in Third and Fourth Biological Windows, *J. Phys. Chem. C*, 2017, 121, 684–690.
- 43 S. Sekiyama, *et al.*, Temperature Sensing of Deep Abdominal Region in Mice by Using Over-1000 nm Near-Infrared Luminescence of Rare-Earth-Doped NaYF<sub>4</sub> Nanothermometer, *Sci. Rep.*, 2018, 8, 16979.
- 44 C. Hazra, A. Skripka, S. J. L. Ribeiro and F. Vetrone, Erbium Single-Band Nanothermometry in the Third Biological Imaging Window: Potential and Limitations, *Adv. Opt. Mater.*, 2020, 8, 2001178.
- 45 S. Hao, *et al.*, Heterogeneous core/shell fluoride nanocrystals with enhanced upconversion photoluminescence for *in vivo* bioimaging, *Nanoscale*, 2015, 7, 10775–10780.
- 46 B. Ritter, T. Krahl, K. Rurack and E. Kemnitz, Nanoscale CaF<sub>2</sub> doped with Eu<sup>3+</sup> and Tb<sup>3+</sup> through fluorolytic sol-gel synthesis, *J. Mater. Chem. C*, 2014, 2, 8607–8613.
- 47 J. Ding, *et al.*, Optical Spectra Properties and Continuous-Wave Laser Performance of Tm,Y:CaF<sub>2</sub> Single Crystals, *Int. J. Opt.*, 2018, 2018, 1–7.
- 48 G. Lakshminarayana, R. Yang, M. Mao, J. Qiu and I. V. Kityk, Photoluminescence of Sm<sup>3+</sup>, Dy<sup>3+</sup>, and Tm<sup>3+</sup>-doped transparent glass ceramics containing CaF<sub>2</sub> nanocrystals, *J. Non-Cryst. Solids*, 2009, 355, 2668–2673.
- 49 Y. Shen, J. Lifante, N. Fernández, D. Jaque and E. Ximendes, *In vivo* Spectral Distortions of Infrared Luminescent Nanothermometers Compromise Their Reliability, *ACS Nano*, 2020, 14, 4122–4133.
- 50 F. E. Maturi, *et al.*, Going Above and Beyond: A Tenfold Gain in the Performance of Luminescence Thermometers Joining Multiparametric Sensing and Multiple Regression, *Laser Photonics Rev.*, 2021, 15, 2100301.
- 51 S. L. Brunton and J. N. Kutz, *Data-Driven Science and Engineering*, Cambridge University Press, 2019, DOI: [10.1017/9781108380690](https://doi.org/10.1017/9781108380690).
- 52 S. L. Brunton and J. N. Kutz, Singular Value Decomposition (SVD), in *Data-Driven Science and Engineering*, Cambridge University Press, 2019, pp. 3–46, DOI: [10.1017/9781108380690.002](https://doi.org/10.1017/9781108380690.002).

