

PAPER

View Article Online
View Journal | View IssueCite this: *RSC Adv.*, 2019, 9, 1147Received 30th October 2018
Accepted 23rd December 2018

DOI: 10.1039/c8ra08967a

rsc.li/rsc-advances

A FRET based ratiometric fluorescent probe for detection of sulfite in food†

Gongxiao Zhang,^{‡a} Ruixue Ji,^{‡a} Xiangyu Kong,^{‡a} Fujiao Ning,^{‡a} Aikun Liu,^a Jichun Cui^b and Yanqing Ge^{†a*}

A new fluorophore pyrido[1,2-*a*]benzimidazole based ratiometric fluorescent probe for the selective detection of sulfite ions in water was investigated. It shows large (pseudo) Stokes shifts (260 nm), high FRET efficiency, high selectivity and sensitivity. A distinct color change from red to colorless was observed and importantly, it proves to be a convenient and efficient tool to detect the sulfite levels in sugar samples.

Introduction

Sulfites are widely applied in the production of foods, pharmaceutical products and beverages as antimicrobial agents, antioxidants and enzyme inhibitors.¹ However, the levels of sulfites have been strictly controlled because of the harmful effects towards the human body, such as allergic reactions and food intolerance symptoms.^{2–4} Therefore, for food quality control and quality assurance, it is important to develop a low-cost, selective and sensitive method for sulfite determination.

Compared with several traditional methods such as electrochemistry and chromatography, fluorescent probes are particularly attractive due to their simplicity, high selectivity and high sensitivity.⁵ Many fluorescent probes are reported to detect sulfites based on the changes of emission intensity.⁶ However, the signal output of these intensity-based fluorescent probes can be affected by instrumental factors, environmental effects and probe concentrations.⁷ Therefore, ratiometric fluorescent probes which are independent of the environmental effects are more desirable.^{8–11} Also, they have large Stokes shifts which can prevent serious self-quenching and fluorescence detection errors. So far, only a few well-behaved ratiometric probes for sulfite have been reported.^{12–26} Herein, a new FRET based ratiometric fluorescent probe was designed (Scheme 1). A pyrido[1,2-*a*]benzimidazole fluorophore, due to its favorable photophysical properties, was selected as the donor. While a hemicyanine fluorophore of which the absorption band

overlapped well the emission band of the donor was chosen as the acceptor (Fig. S1†).

Experimental

Reagents and instrumentation

RF-5301PC luminescence spectrophotometer (Shimadzu) and UV-2600 spectrometer (Shimadzu) were used to measure the fluorescence and UV-vis absorption spectra, respectively. NMR spectra were recorded on a Bruker Avance 400 (400 MHz) spectrometer (in CDCl₃, TMS as an internal standard). All commercial reagents and solvents were used without further purification.

Preparation of the probe PBI-S


Compound 1 and compound 2 were synthesized according to the literature.^{27,28} To a solution of compound 1 (580 mg, 2 mmol) in 20 mL CH₂Cl₂ was added DMAP (60 mg, 0.4 mmol) and EDC (576 mg, 3.0 mmol). Then, compound 2 (744 mg, 12 mmol) was added. The mixture was stirred for 12 h at room temperature. After removal of the solvent under reduced pressure, the crude probe was obtained, which was purified flash silica gel chromatography (CH₃OH : CH₂Cl₂ = 1 : 100) to afford pure probe **PBI-S** (purity: 98.6%, 456 mg, yield: 71%). ¹H NMR (400 MHz, CDCl₃) δ 8.55 (d, *J* = 8.0 Hz, 1H), 8.00 (d, *J* = 8.0 Hz, 1H), 7.96 (d, *J* = 8.0 Hz, 1H), 7.87 (d, *J* = 8.0 Hz, 2H), 7.54 (m, 6H), 7.44 (m, 2H), 6.95 (d, *J* = 8.0 Hz, 1H), 6.79 (d, *J* = 16.0 Hz, 1H), 6.74 (d, *J* = 8.0 Hz, 2H), 3.92 (m, 1H), 3.52 (m, 2H), 3.24 (m,

^aSchool of Chemistry and Pharmaceutical Engineering, Taishan Medical University, Taian, Shandong 271016, P. R. China. E-mail: geyanqing2016@126.com; Fax: +86-538-6229741; Tel: +86-538-6229741

^bSchool of Chemistry and Chemical Engineering, Liaocheng University, Liaocheng, Shandong 252000, P. R. China

† Electronic supplementary information (ESI) available: ¹H NMR, ¹³C NMR and MS spectra of probe, and additional cell images. See DOI: 10.1039/c8ra08967a

‡ Equal contribution.



Scheme 1 Synthesis of probe PBI-S.

1H), 3.12 (m, 2H), 2.98 (m, 1H), 2.18 (m, 1H), 1.75 (s, 6H). ^{13}C NMR (100 MHz, CDCl_3) δ 175.9, 174.2, 167.8, 153.2, 147.5, 146.7, 145.0, 131.7, 130.1, 129.4, 128.7, 126.1, 125.0, 124.2, 122.0, 120.6, 114.4, 111.5, 111.0, 110.7, 110.6, 109.9, 97.2, 96.3, 46.5, 46.2, 45.0, 41.0, 26.7. HRMS: 642.2590 ($[\text{M} + \text{H}]^+$); calcd for $\text{C}_{40}\text{H}_{32}\text{N}_7\text{O}_2$: 642.2617.

Results and discussion

PBI-S was synthesized by the classical condensation of compound **1** with compound **2** in one step in good yield (Scheme 1).

The solvent and effect of water content on fluorescence spectra of the probe was investigated firstly (Fig. S2 and S3 †). Considering its better peak shapes, we chose the solvent ratio (DMF/PBS = 3 : 7) as the *vitro* test solvent system.

The absorption band of **PBI-S**, attributing to hemicyanine and pyrido[1,2-*a*]benzimidazole moiety, centered at 556 nm and 334 nm. When sodium sulfite was added gradually to the solution of **PBI-S** (DMF/PBS = 3 : 7), the absorption peak at 334 nm increased while the peak at 556 nm decreased (Fig. 1). Concomitantly, the solution turned from red to colorless in color, which proves that **PBI-S** could act as a “naked-eye” probe for HSO_3^- . It was calculated that the ratio of the absorbance at 334 and 556 nm displayed an over 300-fold enhancement, which indicates that **PBI-S** is capable for detecting sulfite by UV-vis absorption.

Next, the fluorescence titration was conducted. The free **PBI-S** displayed two obvious fluorescence bands at 465 nm and 640 nm, which attributed to pyrido[1,2-*a*]benzimidazole and hemicyanine unit, respectively. Upon addition of HSO_3^- incrementally, fluorescence emission band at 465 nm increased




Fig. 1 Absorption spectra of **PBI-S** (5 μM) upon addition of HSO_3^- (150 μM) in DMF-PBS buffer solution (3/7, v/v, pH = 7.20).




Fig. 2 Fluorescence spectra of **PBI-S** (5 μM) upon addition of HSO_3^- (150 μM) in DMF-PBS buffer solution (3/7, v/v, pH = 7.20), (λ_{ex} = 380 nm, slit = 10 nm/10 nm).




Fig. 3 UV-vis spectral changes of probe **PBI-S** (5 μM) in the presence of various analytes (150 μM).

gradually while the fluorescence at 640 nm decreased (Fig. 2), which implies that the acceptor moiety was destroyed by the reaction of the probe with HSO_3^- and the FRET between the donor and the acceptor was switched off.


The ratio of emission intensities at 465 and 640 nm (I_{465}/I_{640}) changed from 0.52 in the absence of HSO_3^- to 123.75, a 238-fold variation in the ratios. Furthermore, the emission ratio (I_{465}/I_{640}) was found to increase linearly with the HSO_3^- concentration changed from 3 to 9 μM (Fig. S4 †). The detection limit was determined to be 62 nM which is calculated by $3\sigma/k$ (“ σ ” is the standard deviation of 15 blank measurements, “ k ” is the slope of the fitting line).

Then the selectivity of **PBI-S** toward sulfite over other relevant anions was conducted and it was observed that only by the addition of sulfite could the absorption and fluorescence change. While with other anions, no change in the UV and FL spectra was noted (Fig. 3 and S5 †). However, some spectral changes in GSH, Hcy and Cys were observed due to nucleophilic property of the SH group in them.

Furthermore, we studied the responses of **PBI-S** to the sulfite in the presence of the other competitive anions (including GSH, Hcy and Cys). It was observed that when the other anions were added subsequently, the fluorescence emission ratios (I_{465}/I_{640}) were not influenced (Fig. S6 †).

As shown in Fig. S7 † , the probe was stable and can function well over a wide range of pH (6.0–10.0). Time-dependent fluorescence response to SO_3^{2-} was also carried out (Fig. S8 †). The reaction could be completed within 2 min (50 eq. SO_3^{2-}). Such a rapid response is appealing for real-time detection.

The supposed mechanism is shown in Scheme 2. For the probe alone, FRET between the pyrido[1,2-*a*]benzimidazole and hemicyanine moiety resulted in red emission from hemicyanine. Upon the addition of HSO_3^- , the energy transfer from pyrido[1,2-*a*]benzimidazole donor to hemicyanine acceptor was



Scheme 2 Proposed mechanism of probe.



Table 1 Concentration of the sulfite in sugar samples^a

Sample	HSO ₃ ⁻ level ± SD (μmol L ⁻¹)	Added (μmol L ⁻¹)	Found ± SD (μmol L ⁻¹)	Recovery ± SD (%)
Crystal sugar	0.63 ± 0.048	0.5	1.07 ± 0.031	94.69 ± 2.6%
		1	1.58 ± 0.036	96.93 ± 2.3%
Granulated sugar	1.21 ± 0.049	0.5	1.80 ± 0.026	105.26 ± 0.51%
		1	2.04 ± 0.030	92.31 ± 0.03%
Soft sugar	0.76 ± 0.045	0.5	1.28 ± 0.0140	101.16 ± 3.3%
		1	1.70 ± 0.0127	96.59 ± 3.5%

^a The results were presented as means ± SE with replicates $n = 5$.

interrupted because of the breakage of C=C bond by the addition reaction with HSO₃⁻. The MS spectra of the reaction product also supported the deduction (calcd for C₄₀H₃₄N₇O₅S: 723.2342, found 723.2334, Fig. S9 and S10†).

Finally, **PBI-S** was applied to detect sulfite in sample analysis of commercially purchased granulated sugar, crystal sugar, and soft sugar. The sugar (5.0 g) was dissolved in deionized water and then was diluted to 10 mL. To the solution of the probe **PBI-S** (5 μM) in the DMF/PBS buffer (DMF/PBS = 3 : 7, pH 7.20, 10 mM), aliquots of the sugar solution were added. As shown in Table 1, the HSO₃⁻ concentration in sugar was successfully determined by the probe **PBI-S** with good recovery.

It was calculated that the HSO₃⁻ concentration in the three sugar samples were 0.79, 1.52, and 0.96 mg kg⁻¹, respectively. A traditional titration method was carried out to test the accuracy of this method and Table S1† showed that results of the two methods were equivalent.

Conclusions

In summary, we have developed a new fluorophore pyrido[1,2-*a*]benzimidazole based ratiometric fluorescent probe for selective detection of sulfite ion in water. It shows large (pseudo) Stokes shifts (260 nm), high FRET efficiency, high selectivity and sensitivity. A distinct color change from red to colorless was observed and importantly, it proves to be a convenient and efficient tool to detect the sulfite levels in sugar samples.

Conflicts of interest

There are no conflicts to declare.

Acknowledgements

This work was supported by the Science Fund of Shandong Province for Excellent Young Scholars (ZR2017JL015) and the Natural Science Foundation of China (21602153 and 11574009).

Notes and references

- 1 J. L. Wang, Y. F. Hao, H. Wang, S. X. Yang, H. Y. Tian, B. G. Sun and Y. G. Liu, Rapidly responsive and highly selective fluorescent probe for bisulfite detection in food, *J. Agric. Food Chem.*, 2017, **65**, 2883–2887.

- 2 M. G. Choi, J. Y. Hwang, S. Y. Eor and S. K. Chang, Chromogenic and fluorogenic signaling of sulfite by selective deprotection of resorufin levulinate, *Org. Lett.*, 2010, **24**, 5624–5627.
- 3 X. F. Gu, C. H. Liu, Y. C. Zhu and Y. Z. Zhu, A boron-dipyrromethene-based fluorescent probe for colorimetric and ratiometric detection of sulfite, *J. Agric. Food Chem.*, 2011, **59**, 11935–11939.
- 4 K. Y. Chen, Y. Guo, Z. H. Lu, B. Q. Yang and Z. Shi, Novel coumarin-based fluorescent probe for selective detection of bisulfite anion in water, *Chin. J. Chem.*, 2010, **28**, 55–60.
- 5 Y. M. Yang, Q. Zhao, W. Feng and F. Y. Li, Luminescent chemodosimeters for bioimaging, *Chem. Rev.*, 2013, **113**, 192–270.
- 6 V. S. Lin, W. Chen, M. Xian and C. J. Chang, Chemical probes for molecular imaging and detection of hydrogen sulfide and reactive sulfur species in biological systems, *Chem. Soc. Rev.*, 2014, **44**, 4596–4618.
- 7 L. Yuan, W. Y. Lin, K. B. Zheng and S. S. Zhu, FRET-based small-molecule fluorescent probes: rational design and bioimaging applications, *Acc. Chem. Res.*, 2013, **46**, 1462–1473.
- 8 X. L. Zheng, R. X. Ji, X. Q. Cao and Y. Q. Ge, FRET-based ratiometric fluorescent probe for Cu²⁺ with a new indolizine fluorophore, *Anal. Chim. Acta*, 2017, **978**, 48–54.
- 9 Y. Q. Ge, R. X. Ji, S. L. Shen and X. Q. Cao, A ratiometric fluorescent probe for sensing Cu²⁺ based on new imidazo [1,5-*a*]pyridine fluorescent dye, *Sens. Actuators, B*, 2017, **245**, 875–881.
- 10 Y. Q. Ge, X. J. Xing, A. K. Liu, R. X. Ji, S. L. Shen and X. Q. Cao, A novel imidazo[1,5-*a*]pyridine-rhodamine FRET system as an efficient ratiometric fluorescent probe for Hg²⁺ in living cells, *Dyes Pigm.*, 2017, **146**, 136–142.
- 11 R. X. Ji, A. K. Liu, S. L. Shen, X. Q. Cao, F. Li and Y. Q. Ge, An indolizine-rhodamine based FRET fluorescence sensor for highly sensitive and selective detection of Hg²⁺ in living cells, *RSC Adv.*, 2017, **7**, 40829–40833.
- 12 X. F. Yang, M. L. Zhao and G. Wang, A rhodamine-based fluorescent probe selective for bisulfite anion in aqueous ethanol media, *Sens. Actuators, B*, 2011, **152**, 8–13.
- 13 Y. Q. Sun, P. Wang, J. Liu, J. Y. Zhang and W. Guo, A fluorescent turn-on probe for bisulfite based on hydrogen bond-inhibited C=N isomerization mechanism, *Analyst*, 2012, **137**, 3430–3433.



- 14 H. Y. Tian, J. H. Qian, Q. Sun, H. Y. Bai and W. B. Zhang, Colorimetric and ratiometric fluorescent detection of sulfite in water *via* cationic surfactant-promoted addition of sulfite to α,β -unsaturated ketone, *Anal. Chim. Acta*, 2013, **788**, 165–170.
- 15 M. Y. Wu, K. Li, C. Y. Li, J. T. Hou and X. Q. Yu, A water-soluble near-infrared probe for colorimetric and ratiometric sensing of SO_2 derivatives in living cells, *Chem. Commun.*, 2014, **50**, 183–185.
- 16 Z. X. Liu, S. J. Guo, J. Y. Piao, X. Zhou and X. Wu, A reversible fluorescent probe for circulatory detection of sulfites through a redox-based tandem reaction, *RSC Adv.*, 2014, **4**, 54554–54557.
- 17 W. Xu, C. L. Teoh, J. J. Peng, D. D. Su, L. Yuan and Y. T. Chang, A mitochondria-targeted ratiometric fluorescent probe to monitor endogenously generated sulfur dioxide derivatives in living cells, *Biomaterials*, 2015, **56**, 1–9.
- 18 Y. Liu, K. Li, M. Y. Wu, Y. H. Liu, Y. M. Xie and X. Q. Yu, A mitochondria-targeted colorimetric and ratiometric fluorescent probe for biological SO_2 derivatives in living cells, *Chem. Commun.*, 2015, **51**, 10236–10239.
- 19 X. J. Liu, Q. W. Yang, W. Q. Chen, L. N. Mo, S. Chen, J. Kang and X. Z. Song, A ratiometric fluorescent probe for rapid, sensitive and selective detection of sulfur dioxide with large Stokes shifts by single wavelength excitation, *Org. Biomol. Chem.*, 2015, **32**, 8663–8668.
- 20 B. Roubinet, L. Bailly, E. Petit, P. Y. Renard and A. Romieu, A FRET-based probe for fluorescence sensing of sulfide/sulfite analytes, using a novel long-wavelength water-soluble 7-hydroxycoumarin as reporter fluorophore, *Tetrahedron Lett.*, 2015, **56**, 1015–1019.
- 21 W. L. Wu, H. L. Ma, M. F. Huang, J. Y. Miao and B. X. Zhao, Mitochondria-targeted ratiometric fluorescent probe based on FRET for bisulfite, *Sens. Actuators, B*, 2017, **241**, 239–244.
- 22 Z. Chen, F. Z. Chen, Y. C. Sun, H. Liu, H. P. He, X. H. Zhang and S. F. Wang, A novel ratiometric fluorescent probe for selective detection of bisulfite in living cells, *RSC Adv.*, 2017, **7**, 2573–2577.
- 23 Y. Yao, Q. Sun, Z. Chena, R. Huang, W. Zhang and J. Qian, A mitochondria-targeted near infrared ratiometric fluorescent probe for the detection of sulfite in aqueous and in living cells, *Talanta*, 2018, **189**, 429–436.
- 24 J. Xu, D. J. Zheng, M. M. Su, Y. C. Chen, Q. C. Jiao, Y. S. Yang and H. L. Zhu, A rapid cell-permeating turn-on probe for sensitive and selective detection of sulfite in living cells, *Org. Biomol. Chem.*, 2018, **16**, 6940–6946.
- 25 K. Dou, G. Chen, F. Yu, Z. Sun, G. Li, X. Zhao, L. Chen and J. You, A two-photon ratiometric fluorescent probe for the synergistic detection of the mitochondrial SO_2/HClO crosstalk in cells and *in vivo*, *J. Mater. Chem. B*, 2017, **5**, 8389–8398.
- 26 K. Dou, Q. Fu, G. Chen, F. Yu, Y. Liu, Z. Cao, G. Li, X. Zhao, L. Xia, L. Chen, H. Wang and J. You, A novel dual-ratiometric response fluorescent probe for SO_2/ClO^- detection in cells and *in vivo* and its application in exploring the dichotomous role of SO_2 under the ClO^- induced oxidative stress, *Biomaterials*, 2017, **133**, 82–93.
- 27 Y. Q. Ge, X. L. Zheng, R. X. Ji, S. L. Shen and X. Q. Cao, A new pyrido[1,2-a]benzimidazole-rhodamine FRET system as an efficient ratiometric fluorescent probe for Cu^{2+} in living cells, *Anal. Chim. Acta*, 2017, **965**, 103–110.
- 28 Y. Q. Ge, A. K. Liu, R. X. Ji, S. L. Shen and X. Q. Cao, Detection of Hg^{2+} by a FRET ratiometric fluorescent probe based on a novel pyrido[1,2-a]benzimidazole rhodamine system, *Sens. Actuators, B*, 2017, **251**, 410–415.

