



Cite this: *RSC Adv.*, 2019, 9, 4963

A review of polystyrene bead manipulation by dielectrophoresis

Qiaoying Chen and Yong J. Yuan *

Exploitation of the intrinsic electrical properties of particles has recently emerged as an appealing approach for trapping and separating various scaled particles. Initiative particle manipulation by dielectrophoresis (DEP) showed remarkable advantages including high speed, ease of handling, high precision and being label-free. Herein, we provide a general overview of the manipulation of polystyrene (PS) beads and related particles *via* DEP; especially, the wide applications of these manipulated PS beads in the quantitative evaluation of device performance for model validation and standardization have been discussed. The motion and polarizability of the PS beads induced by DEP were analyzed and classified into two categories as positive and negative DEP within the time and space domains. The DEP techniques used for bioparticle manipulation were demonstrated, and their applications were conducted in four fields: trapping of single-sized PS beads, separation of multiple-sized PS beads by size, separation of PS beads and non-bioparticles, and separation of PS beads and bioparticles. Finally, future perspectives on DEP-on-a-chip have been proposed to discriminate bio-targets in the network of microfluidic channels.

Received 31st October 2018

Accepted 14th January 2019

DOI: 10.1039/c8ra09017c

rsc.li/rsc-advances

Laboratory of Biosensing and MicroMechatronics, School of Materials Science and Engineering, Southwest Jiaotong University, Chengdu, Sichuan 610031, China.
 E-mail: yongyuan@swjtu.edu.cn

1. Introduction

The manipulation of various scaled particles is of significant interest in biomedical applications. The manipulation techniques mostly include electrophoresis (EP),¹ dielectrophoresis (DEP),^{2,3} optical tweezing,⁴ Raman spectroscopy,^{5,6} impedance (IM) spectroscopy,^{7,8} hydrodynamic flows^{9,10} and acoustic



Qiaoying Chen has been pursuing a PhD degree in Materials Science at Southwest Jiaotong University, Chengdu, China since 2011. She received her Bachelor's degree from Chongqing University in 2010. Her research interests include dielectrophoresis systems, biosensors, and microbiological detection. Publication list: Y. J. Yuan, Q. Chen, J. Li, QCM-based immunosensor for determining immunoglobulin G, *RSC Adv.*, 2016, 6(46), 40336–40342.



Yong J. Yuan received his PhD degree in biosensors from the University of Western Sydney, Australia, in 1997. He is a Professor of BioMEMS in the School of Materials Science and Engineering at Southwest Jiaotong University (SWJTU), China. Prior to joining SWJTU in 2008, he was a Research Chemist at CSIRO (1997–1999) working on oxygen scavenging materials, a Guest Researcher of the Agency

of Industrial Science and Technology in Japan (1999–2000) conducting research on the metabolic oscillation of yeast, and a Research Program Leader at the Industrial Research Limited of New Zealand (2000–2008) to develop biochips by the integration of novel materials, structures, devices, and processes. His current research interests include biomedical sensors, MEMS, nano/micro devices and fabrication.






Fig. 1 Studies reported on DEP from 2000 to 2018.¹⁴

traps;¹¹ DEP has attracted significant attention to control particles due to its high selectivity and rapidity in manipulation.^{12,13} DEP was extensively used in 2005 and then reached a culmination in 2016.^{16,17} There are nearly 6k reports on DEP since 2000, as shown in Fig. 1.¹⁴ The number of studies reported on DEP has significantly increased due to the development of micro-fabrication and related techniques. There was an explosion in the number of studies reported on DEP in 2003–2008, and a steady plateau was achieved in 2008–2018.

DEP is based on the fact that particles with different electrical characteristics behave differently in a nonuniform electric field. The DEP force provides an efficient and convenient way to control particles, especially for selective manipulation and rapid separation. It has been used in industrial food safety, clinical cell sorting, infection diagnosis, and enrichment of particle populations for drug development. DEP has been used in the trapping process,¹⁵ alignment,^{16,17} and isolation and separation of various sized particles,¹⁸ which are applied in a wide range of clinical, biological and environmental applications. Some particles, such as cells, nanoparticles (NPs),¹⁹ nanotubes (NTs),²⁰ nanowires,²¹ nanorods,²² DNA²³ and viruses,²⁴ are typically used in the DEP manipulation, as shown in Fig. 2. As a standard particle, polystyrene (PS) beads are widely used in the performance and quantitative evaluation of DEP systems to demonstrate the feasibility of the proposed design.




Fig. 2 Studies reported on DEP over 2000–2017 for various types of DEP particles.^{20–25}

Considering the various types of biomedical and biotechnological applications, it is apparent that *in situ* diagnostics for cells and bacteria are highly significant in DEP manipulation. Mansor *et al.* presented a historical review of single-cell electrical property analysis and its development from classical techniques to recent advances in microfluidic techniques.²⁶ Kasetsirikul *et al.* further reviewed DEP as a malaria diagnostic method for both detection of malaria and separation of infected erythrocytes, especially the malaria parasite-infected red blood cells (RBCs).²⁷ Adekanmbi's group provided a much better and deeper understanding on how DEP could be utilized to manipulate diseased cells such as those associated with malaria, cancer, dengue, anthrax and human African trypanosomiasis.²⁸ Páez-Avilés *et al.* summarized both the DEP and impedance analysis (IA)-combined approaches and the latest relevant improvements, such as those in terms of selectivity, sensitivity, detection time, and conductivity variation enhancements, focusing on bacteria concentration and detection.²⁹ Viefhues's group focused on DNA manipulated by DEP and presented an important theory in DEP applications, a recent review of DNA applications and discussion of the current challenges and future tasks.³⁰

Herein, we reviewed studies mainly reported between 2000 and 2018 on the DEP manipulation of particles. This review is distinct from the previous reviews since we have solely focused on the DEP manipulation of PS beads. A brief background on the DEP theory and simulation of PS beads has been discussed. The detailed analysis of the PS beads experimentally manipulated by DEP has been divided into two sections: trapping of single-sized PS beads and separation of the PS beads and mixed particles (*e.g.* non-bioparticles, DNA, cells, bacteria, and viruses). Finally, the DEP techniques have been proposed for biosensing applications in the future.

2. Theoretical background and simulation

2.1. Motion of the PS beads

DEP is the motion of dielectrically polarized particles in non-uniform electric fields. Unlike direct current (DC) electric fields,³¹ alternating current (AC) electric fields can be used to manipulate many types of particles in different media by simply adjusting the AC electric field parameters (*e.g.*, magnitude, frequency, wave shape, wave symmetry and phase) and eliminating the influence of electrostatic forces. Manipulation of nano/micro-scaled particles within a DEP chip can generally be categorized based on the types of force fields used for the manipulation. When suspended in a fluid medium in the microchannel, an electrically polarizable particle experiences F_{DEP} , which makes the cell move to the local field direction.¹⁶ If the dielectric constant of the particle is different from that of the surrounding medium, the particle moves toward the high intensity (pDEP) or low intensity (nDEP) part of the electric field. In addition to F_{DEP} , the particle is subjected to other forces, such as Stokes' drag force, F_{Drag} ,^{32,33} gravity force, F_{Gra} ,³⁴ buoyancy force, F_{Buoy} , AC electroosmosis³⁵ (ACEO) force, F_{EO} ,



electrophoretic (EP) force, F_{EP} ,³⁵ and Brownian force, F_B ,^{36,37} as listed in Table 1. The Clausius–Mossotti (CM) factor $K(\omega)$ is related to the electric field frequency ($\omega = 2\pi f$); r_p is the radius of a particle, σ is the conductivity and ϵ^* is the complex permittivity of particles and the suspending medium ($\epsilon^* = \epsilon - j\sigma/\omega$, where the subscripts p and m represent the particles and suspending medium, respectively). c is a non-dimensional factor accounting for the wall effects ($c = 1$ for spherical), η is the dynamic viscosity of the fluid,³⁷ v is fluid velocity,³³ ρ is density, V is volume, q is the effective charge, and E is the electric field intensity. ξ is a Gaussian random vector, k_B is the Boltzmann constant, T is the absolute temperature, and Δt is the time step used in the numerical integration scheme. C_b is the coefficient that depends on the distance among particles and the length of the particle chains ($3 < C_b < 1000$).

2.1.1. Gravity and buoyancy vs. Brownian motion. The Stokes' drag force F_{Drag} depends on the flow rates in microfluids; if the flow rate ranges from 10 nL min^{-1} to $10 \text{ }\mu\text{L min}^{-1}$, it provides drag forces in the range from 0.07 to 70 pN on a $1 \text{ }\mu\text{m}$ PS bead.³⁸ The drag force can be apparently neglected considering that particles move slowly in a droplet. If a $10 \text{ }\mu\text{m}$ PS bead falls into a DEP area by gravity and buoyancy, it experiences a force of approximately 50 fN. At a low frequency ($<100 \text{ kHz}$), the ACEO exerts a significant force on particles, resulting in convection of the trapped particles to the centre of electrodes. This effect is strongest when the electric double layer (EDL) is presumed to be in quasi-equilibrium, corresponding to a frequency significantly below the charge relaxation time of an electrolyte.³⁹ The use of larger voltages prevents the hindering effects of Brownian motion; however, excessive voltages may cause unwanted effects, including electrolysis, electrochemical reactions or Joule heating, which can be overcome using a high frequency ($>1 \text{ MHz}$).⁴⁰

As abovementioned, gravity can be ignored if the particle density is close to that of the fluid and the diameter of the particle is small.⁴¹ Moreover, Brownian motion becomes important as the particle size decreases to the nano-scale. The relative magnitude of Brownian force increases with a decrease in particle size. This is due to thermal fluctuations at the molecular-scale in the liquid suspending medium that induce random forces on the suspending particles.

2.1.2. Particle interaction vs. Brownian motion. Not only the equivalent dipole moment of a single particle in the AC electric field should be considered, but also the interaction between neighboring particles should be taken into account when the concentration of the sample liquid increases.⁴² PS beads with similar and dissimilar sizes and stable physico-chemical properties are often used as numerical analysis objects in this area due to their particle–particle interaction and relative motion.^{43,44} Mathematically, modelling is feasible to study the interactions between dielectric spheres employing multipole expansion. Moreover, the numerical methods used for solving coupling flow-electricity-particle motion can correctly reveal the interactive motion behaviour of the DEP particles in a uniform electrical field. The particle velocity and time behaviour of interactive motion have been investigated.⁴⁵ The Maxwell stress tensor (MST) force is responsible for the strongly coupled particle–fluid–electric field interactions.⁴⁶ On the other hand, the calculation of the AC DEP force is considered as the most rigorous approach.⁴⁷ Similarly, Kang *et al.* analyzed two-particle system dynamics on AC fields by comparing the relative movement of particles.⁴⁸

On the other hand, the iterative dipole moment method (IDM) can also be used to calculate the electrostatic force between two adjacent beads and to investigate the relationship between the size of the beads and the distance among them.⁴⁹ Some experiments were carried out on the orientation of particles in frequency-selectable directions,⁵⁰ predicting the threshold potential or individual and chain velocities of microparticles and cells under DEP forces.^{51,52} The motion of particles along DEP forces showed subtle vibrations due to Brownian motion.³⁷

2.2. Polarizability of the PS beads

The basic principle of DEP depends on the extent of polarization. It is not only time-averaged translational DEP,¹⁷ but also travelling-wave DEP (twDEP).⁵³ In time-averaged translational DEP, the Clausius–Mossotti factor is frequency dependent, and thus, the DEP force will also be frequency dependent. The apparent crossover frequency (COF) is the frequency at which a particle–medium system will switch from being dominated by

Table 1 Force on the PS beads in the microfluids applied with an AC electric field

Force	Electric field	Scale of PS beads	Formula	No.
DEP	10 kHz to 10 MHz	Micro/nano	$F_{DEP} = 2\pi r_p^3 \epsilon_m \text{Re}[K(\omega)] \nabla E^2$ $K(\omega) = \frac{\epsilon_p^* - \epsilon_m^*}{\epsilon_p^* + 2\epsilon_m^*}$	(1)
Stokes' drag	High flow rates	Micro/nano	$F_{Drag} = 6\pi c \eta r_p (v_m - v_p)$	(2)
Gravity and buoyancy	Unlimited	Unlimited	$F = F_{Buoy} - F_{Gra} = \frac{4}{3} \pi r_p^3 g (\rho_m - \rho_p)$	(3)
ACEO	AC (0–100 kHz)	Micro	$F_{Eo} = \frac{1}{12} r_p^3 \rho_m \frac{d}{dt} (V_m - V_p)$	(4)
EP	High (DC or AC in low frequency)	Charged	$F_{EP} = qE$	(5)
Brownian	Low	Nano	$F_B = \xi \sqrt{\frac{12\pi k_B \eta T r_p}{\Delta t}}$	(6)
Interaction	10 kHz to 10 MHz	—	$F_{chain} = -C_b \pi \epsilon_m \text{Re}[K(\omega)]^2 r_p^2 E^2$	(7)



the conductivities of its system to being dominated by the permittivities; this prompts the direction of the DEP force to be reversed.¹⁸ This can be extracted from the balance between the Stokes' drag force F_{Drag} due to the background flow over a stationary particle and DEP force at the vicinity of the electrode edge. The proximity of the particle to the wall necessitates the inclusion of a correction factor to the Stokes' drag force F_{Drag} to account for the wall effect under the assumption that the gap size d between electrodes is less than the particle radius r_p . Thus, the crossover angular frequency ω_c is calculated by taking $\text{Re}(K) = 0$, which may be defined as follows:⁵⁴

$$\omega_c = \sqrt{\frac{(\sigma_p - \sigma_m)(\sigma_p + 2\sigma_m)}{(\epsilon_p - \epsilon_m)(\epsilon_p + 2\epsilon_m)}} \quad (8)$$

For low-permittivity particles in aqueous media (e.g. DI water), particles move towards the high field region at $\omega < \omega_c$, referred to as positive DEP (pDEP), and towards the low field region at $\omega > \omega_c$, referred to as negative DEP (nDEP). For particles in water, nDEP occurs when the permittivities of the system dominate due to the large $\epsilon = 79$ of water as compared to that of nearly all other substances. The COF is given by

$$f_{\text{COF}} = \frac{\omega_c}{2\pi} \quad (9)$$

The key parameter, *i.e.* the electrical conductivity of a solid homogeneous PS bead, can be expressed as

$$\sigma_p = \sigma_s + \frac{2K_s}{r_p} \quad (10)$$

where K_s represents the surface conductance and r_p is the radius of a particle.⁵⁵ K_s can be calculated as

$$K_s = \frac{r_p}{4} \left(-\sigma_m + \sqrt{9\sigma_m^2 - 4(\epsilon_p - \epsilon_m)(\epsilon_p + 2\epsilon_m)(2\pi f_{\text{COF}})^2} \right) \quad (11)$$

where f_{COF} is the crossover frequency and the DEP force is zero ($\text{Re}(K) = 0$). Therefore, the value of K_s has a linear relationship with the PS bead diameters if the permittivities of the suspending medium and PS beads are constant under f_{COF} . Surface conductance was calculated for the PS beads,⁵⁶ and the $\text{Re}(K)$ values of PS beads with different diameters were calculated, as shown in Fig. 3. If the 10 μm PS beads ($\epsilon_p = 2.6$ and $\sigma_p = 0.001 \text{ mS m}^{-1}$) were suspended in DI water ($\epsilon_m = 79$ and $\sigma_m = 0.2 \text{ mS m}^{-1}$), the frequency was on the order of 10 kHz for normal particle surface conductivities (about 1 nS). Thus, the 10 μm PS beads in DI water will distribute at the electrode gap under nDEP from 1 kHz to 100 MHz frequencies. However, other values (e.g. 1 or 0.1 μm) refer to the pDEP distributed at the electrode edge where the electric field is strongest. The conductance and DEP force are particularly sensitive to an increase in the surface charge,⁵⁷ especially in the case of nanoparticles. When the concentration of the suspending medium increases (e.g., conductivity, σ_m , increases), $\text{Re}(K)$ decreases and reverses its sign from positive to negative at low frequencies.^{58,59}




Fig. 3 $\text{Re}(K)$ with frequency on the PS beads.

The frequency is one of the important factors to ensure the polarizability of the PS beads. The voltage is applied to generate stable and sufficient DEP forces, further inducing particles within a certain electrode geometry. Several electro-physiological effects need to be considered while designing electrode architectural structures since DEP can directly affect the cell physiology.

2.3. Numerical analysis

One of major topics to be investigated in this field is numerical simulation or modelling, which can save resources, shorten the experimental period, and predict results. Simulation is used to identify an optimized geometry to enhance the capture of particles. With the rapid development of electronic computers, calculation softwares, such as COMSOL Multiphysics (COMSOL, Burlington, MA), computational fluid dynamics (CFD, ESI Group, France) and ANSYS Fluent (Fluent Inc, Lebanon, USA), based on various mathematical models have been widely used for the calculation of DEP forces. The distribution of a non-uniform electric field, the force of particles and the trajectory of motion are mainly analyzed in these software, as summarized in Table 2. To date, COMSOL is the most popularly used software in the calculation of both flow and electric fields due to the implementation of a finite element method (FEM) with adaptive meshing, error control and a variety of numerical solvers. COMSOL focuses on multiphysics modelling in electric fields, flow fields, thermal fields and particle trajectories, which are characterized by coupled calculation and tracer demonstration of the coupling F_{DEP} , F_{Drag} , F_{Gra} , F_{Eo} and thermal parameters. CFD and ANSYS FLUENT are good for calculating the F_{DEP} and F_{Drag} of particles in fluid. However, the problem associated with simulation is that most simulations only provide the distribution of the electric field and does not accurately predict the movement of particles. The difficulty lies in how to simulate the integrated force of particles in a complex fluid and track the movement of particles under combined forces at a controlled velocity.

Although the general motion of a polarizable PS bead under the DEP force may be deduced from the electrical field distribution, the trajectory solution of the PS bead under moving DEP is not trivial. Cao *et al.* simulated the electric field distribution for different pitch values while keeping other geometrical



Table 2 Commercial software used for DEP simulation (“–” for nDEP and “+” for pDEP)

Software	Scaling	Force	Parameter	2D	3D	Application	Ref.	
COMSOL	3–5 μm	$+F_{\text{DEP}}$	Electric field	Cross-section	No	Improvement of collection	63	
	1 μm	$+F_{\text{DEP}}, F_{\text{Drag}}$	Electric field	Cross-section	No	Collecting bacteria	64	
	—	$+F_{\text{DEP}}, F_{\text{Drag}}$	Electric field; electroosmotic	Top-section	No	Optimizing the conditions	65	
	—	F_{Eo}						
	—	$+F_{\text{DEP}}, F_{\text{Drag}}, F_{\text{Eo}}$	Electric field; particles trajectories; Joule heating	Top-section	Yes	Separation of healthy RBCs and RBCs infected by <i>P. falciparum</i>	66	
	9 μm	$+F_{\text{DEP}}$	Electric field	Cross-section	No	Transporting of yeast	67	
	20–30 μm	$+F_{\text{Drag}}$	Flow field	Cross-section	No	Separation of cancer cell	68	
	15 μm	$-F_{\text{DEP}}, F_{\text{Gra}}$	Electric field	Cross-section	Yes	Collection PS beads and Si beads	69	
CFD + ACE	10 μm	$+F_{\text{DEP}}, \text{nDEP}$	Electric field	Cross-section	No	Manipulation of large groups of particles	70	
	10 μm	$-F_{\text{DEP}}, F_{\text{Drag}}$	Electric field	Cross-section; top-section	Yes	Trapping of PS beads	71	
	10 μm	$-F_{\text{DEP}}$	Electric field; particles trajectories	Top-section	Yes	Trapping of PMMA NPs and cancer cells	72 and 73	
	—	$+F_{\text{DEP}}$	Electric field	Cross-section	Yes	Transporting of the oocyte and the sperm	74	
	ANSYS	2–3 μm	$+F_{\text{DEP}}$	Electric field	Cross-section	No	Trapping of erythrocyte	75
	FLUENT	9 μm	$+F_{\text{DEP}}$	Electric field	Cross-section	Yes	Manipulation of yeast	76
		4.8 μm	$+F_{\text{DEP}}$	Electric field	Cross-section	No	Manipulation of PS beads and SnO ₂ nanobelts	77
1 μm		$+F_{\text{DEP}}$	Electric field	Cross-section	Yes	Separation of PS beads and MWCNTs	55	
	0.08 μm	$+F_{\text{DEP}}, F_{\text{Drag}}$	Electric field	Cross-section	Yes	Manipulation of WO ₃ nanoparticles	78	

parameters and the applied voltage identical to the values used in their experiments.⁶⁰ Yan *et al.* tested the potential effect of voltage on the DEP-active hydrophoretic focusing and the PS beads passing through and obtained different patterns, as shown in Fig. 4A.⁶¹

2.3.1. Hydrophoretic focusing of the 10 μm PS beads. PS beads smaller than half of the channel height in diameter⁶¹

were introduced at a 20 $\mu\text{L min}^{-1}$ flow rate. They migrated back and forth inside the channel, as evidenced in (a₁) in Fig. 4A, due to unsatisfied hydrophoretic ordering. They could not form hydrophoretic focusing without an electric field. However, the PS beads tended to focus on channel sidewalls when 5 V_{pp} voltage was applied at 1 MHz frequency (a₂ in Fig. 4A). The beads were subjected to a nDEP force and were levitated to




Fig. 4 Particle trajectory in numerical analysis. (A) Particle trajectory at a voltage of 5 V_{pp}: (a₁) inlet and (a₂) outlet. Particles focused onto channel sidewalls under the effect of steric hindrance.⁶¹ (B) Trajectory of (b₁) the PS beads in DI water and (b₂) HeLa cells in a phosphate buffer saline (PBS) medium.⁶² (C) Simulation of the isolated cells in optoelectronic tweezers (OET): two randomly distributed cells were separated in 7 s under the (c₁) pDEP and (c₂) nDEP force.⁷⁹ (D) Particle position at 0 s, 30 s and 60 s, and the magnitude contours of the electric fields. The gap of two bottom electrodes was 5 times the height of the dispersion chamber, and 15 μm PS beads were used for the simulation: (d₁) $U = 100 \text{ V}$ and (d₂) $U = 150 \text{ V}$.⁶⁹



a higher position in the channel. This resulted in a greater intensive steric interaction between the particle and groove that led to hydrophoretic ordering. Regardless of their initial positions, the PS beads were focused onto each side of the channel. Das *et al.* performed numerical simulations in COMSOL to predict the distributions of a non-uniform electric field, DEP forces and particle trajectories, as shown in Fig. 4B.⁶² When a 10 V_{pp} voltage was applied at 1 MHz, the HeLa cells (pDEP) moved away from the center of the streamline (b_2 of Fig. 4B), whereas the PS beads (nDEP) followed the central streamline along the channel length (b_1 of Fig. 4B).

2.3.2. Dipolar cells. A 3D numeric model of cells was investigated with two different types of CM values in an optically induced DEP (ODEP) chip.⁷⁹ The particle tracing for the fluid model in COMSOL was used to solve the numerical model of cells under the DEP force. The DEP force acting on the dipole of 11.8 μm cells subjected to a non-uniform electric field under 60 μm Gaussian-distributed beam spot was simulated in the enrichment process. The simplified structure and numeric model of the nucleated cell provided a theoretical basis for research on biosensors and natural life. The start point of microparticles in the optoelectronic chip was scattered randomly. The direction of the DEP force, which depends on the size of the cells, permittivity, and conductivity of the suspending medium, changed upon the movement of the cells. The CM factor that referred to two types of cells was opposite if adjusting frequency on a function generator. As seen in (c_1) in Fig. 4C, the pink and white cells were shown to simulate the optoelectronic tweezer (OET) chip due to their different CM values. The result of separation demonstrated that the optical spot captured the cells with a specific frequency. On the other hand, the white cells were repelled far away from the pink cells (c_2 in Fig. 4C). The CM value of the pink and white cell was equal to 0.4 and -0.2 , respectively, and the frequency f was 0.5 MHz. The assembly and isolation of cells were confirmed by the numerical model of cells, which maintained better consistency with the literature referring to experiments.

2.3.3. PS beads focusing in time-domain. In addition, the 15 μm PS bead focusing was further simulated using (d_1) $U =$

100 V and (d_2) $U = 150$ V, as shown in Fig. 4D.⁶⁹ For both cases, the difference between the gravitational force and the vertical component of the DEP force was illustrated, where in the white region, $e < 0$, and in the grey region, $e > 0$. The three images show the position of the particles at $t = 0$ s, 30 s and 60 s.

The optimization of microfluidics numerical study offers the advantages such as requirement of less sample volume, faster analysis, precise fluid handling, and reduced biological and chemical wastes. The 3D structure design of electrode arrays will avoid the decentralization of a planar electric field and improve the sensitivity and separation efficiency of particles. The development of microfluidics also allows for the creation of better bio-devices for flow management and better understanding of the interaction between the hydrodynamic and electrokinetic forces.

3. Application

The applications of DEP covering from nanoscale to micro and macroscale materials, such as particles, wires, tubes and rods towards DNA, cells, and bacteria, are summarized in Table 3. Nano-scaled subjects include NPs, NTs, nanowires, nanorods, DNA and viruses. NPs have been extensively investigated by DEP manipulation, as shown in Fig. 2, due to their relative simplicity and repeatability. The evaluation of the performance and efficiency of NPs through the analysis and theoretical prediction of DEP can be regarded as a new research direction of DEP technology. In addition, NPs can be substantially and effectually used as nano-parts to fabricate functional and/or electrical biosensors.

DEP has been a useful tool for the study of different cells since its development. In 1966, yeast cells were first applied in the DEP separation of living and dead cells.⁸⁰ After this, DEP equipment has been mainly used in the enrichment and isolation of yeast cells. Since 1982, when human malignant melanocytes were first successfully arranged by DEP,⁸¹ DEP has been widely used in the manipulation of tumor cells. In 1984, the blood cells obtained from patients suffering from sickle cell anemia were first used to distinguish normal red blood cells

Table 3 DEP applications in different scales

Nano-scale	Micro-scale	Milli-scale
Nanoparticles (PS, ¹⁹ Au; Ag, ⁸⁵ Pd, ⁸⁶ CdTe, ⁸⁷ Silica, ⁸⁸ PMMA, ⁷² tungsten trioxide (WO ₃) ⁷⁸ and ZnO ⁸⁹)	Non-bioparticles (PS, ⁹⁰ Janus; Silica, ⁹¹ and melamine particle ⁹²)	Algae ⁹³
Nanowires (SiC; ZnO, ^{21,110} CuO, ¹¹¹ Pt, ¹¹² and Fe ₂ O ₃ ¹¹³)	Cells (yeast cell, ⁹⁴ RBC, cancer cell, fibroblast cell, ⁹⁵ osteosarcoma cell, Neural, ⁹⁶ human leukemia cell, ^{97,98} circulating tumor cells, ⁹⁸ <i>M. smegmatis</i> , ⁹⁹ and <i>B lymphoma</i> cell ¹⁰⁰)	Clay ¹⁰⁷
Carbon-based nanomaterials (carbon nanotubes ²⁰ and graphene ¹¹⁴)	Bacteria (<i>E. coli</i> , ^{58,59} <i>Staphylococcus epidermidis</i> , ¹⁰¹ <i>Salmonella</i> , <i>Listeria monocytogenes</i> , ¹⁰² <i>K. pneumonia</i> , ¹⁰³ and <i>Lactobacillus</i> ^{104,105})	Silicone ¹⁰⁸
Nanorods (ZnO ^{22,115}) Nanobelts (SnO ₂ ¹¹⁶ and ZnO ¹¹⁷) DNA ^{23,118} Virus ^{24,119} Bacteriophage ¹²⁰	Single-stranded DNA ¹⁰⁶	Minerals ¹⁰⁹



Table 4 Structures fabricated by trapping of micro-scaled PS beads

Structures (efficiency)	Theory	Scale (μm)	Medium (mS m^{-1})	Electric field	Ref.
Accumulated crystals (17–341%)	nDEP	10	DI water	2–6 V_{pp} , 1 MHz	124
	AC-DEP	2	KCl (1.42)	10 V_{pp} , 0.1–10 MHz	126
	DEP	10, 25, 45	KCl (0.7–15.9)	4–12.5 V_{pp} , 1 kHz to 1 MHz	132
	DEP	10	DI water (1.7); NaCl (8.4)	10 V_{pp} , 100 Hz to 5 MHz	125
	AC-DEP	10	—	21 V_{pp} , 50 kHz	133
	AC-DEP	1	—	12 V_{pp} , 1–4 MHz	127
	DC-iDEP	6	DI water	200–500 V	130
	ODEP	2.9	KCl	555–885 V_{RMS} , 30–100 kHz	131
	Lined chains (95–100%)	twDEP	10	DI water (1)	2 V_{pp} , 1 MHz
AC-DEP		4.3	DI water (10)	5 V_{pp} , 100 Hz to 10 MHz	34
AC-DEP		10	DI water (0.2)	5 V_{pp} , 1 MHz	63
ACEO-DEP		0.5, 2, 3	DI water	2–5 V_{pp} , 0.1–500 kHz	134
pDEP		1, 2, 5, 10	DI water (152/158)	2–10 V, 500 Hz	135
Acoustic-DEP		10	DI water (0.001)	16 V_{pp} , 5 MHz	70
DC-iDEP		10	PBS	100–850 V	136
DEP		10	DI water (20)	10 V, 500 kHz	137
iDEP		18, 24.9	DI water (0.6)	15–300 V, 20 kHz	138
CP-DEP		5, 10	PBS (1–250)	141 V_{RMS} , 0.5–500 kHz	139
Wires	cDEP	10	DI water (0.5)	95–160 V_{pp} , 100 kHz to 1 MHz	140
	IM-DEP	6	NaCl	2 V_{pp} , 184 Hz	141
	AC-DEP	6, 10	(20)	2–10 V, 1 MHz	142
Arrays (63%)	AC-DEP	10	Tris–acetate–EDTA (60)	7 V_{pp} , 1 MHz	143

and sickle cells by DEP.⁸² Subsequently, the study of mammalian cells (*i.e.* tumor cells⁸³ and fibroblasts⁸⁴) was applied in DEP research. In 1989, *Micrococcus lysodeikticus* was first investigated by optical DEP measurements.¹²¹ Subsequently, bacteria began to be introduced into the DEP application. Since 1994, with the rapid development of micro-electromechanical systems, DEP devices have exhibited greater precision and can be used in the manipulation of smaller particles such as DNA¹²² and viruses.¹²³ Cells obtained from various species possessing dissimilar particles experience different DEP forces. Furthermore, the same type of particles present irregular spheres and homogeneity due to their different growth states. Moreover, it is difficult to obtain accurate electrical conductivity and dielectric constants; this leads to some differences between the simulation and actual results. Therefore, PS beads with a standardized size, dielectric constant and conductivity can be used to more accurately simulate the force and motion of particles.

DEP forces are versatile and can be tuned simply by changing the frequencies or by fabricating electrodes with different shapes in an array to achieve DEP trapping. This approach plays a vital role in biosensors, diagnostics and medical research. Particles should be first trapped in the detection region. The location has an important impact on the accuracy of the measurement results. PS beads are often used to verify the feasibility of devices with biological particles because they are comparable in size. As has been discussed in the previous section, trapping and separation of PS beads are major issues in DEP manipulation.

3.1. Trapping of single-sized PS beads

3.1.1. Micro-scaled PS beads. DEP manipulation of PS beads is primarily carried out for trapping various scaled beads

in a suspending medium by an electric field. The DEP voltages applied are significantly lowered from the thousands to tens scale; however, the electric field density is sufficient for particle manipulation. Micro-scaled beads are typical and fundamental subjects for the investigation of trapping conditions. Recently, nano-scaled and functionalized PS beads were specially applied for affinity biosensors that exploited antibody–antigen interactions.

(1) *Accumulated crystals.* Electrode geometry is an important factor to ensure that stable and sufficient DEP forces are applied to induce micro-scaled PS beads. A number of microelectrode geometries were designed for specific research purposes. In previous studies, DEP-manipulated PS beads were used in solution to fabricate many structures including crystals, chains, wires and arrays, as summarized in Table 4.

The crystal was the earliest structure of PS beads trapped by DEP, which depended on electrode shapes including interdigitated, quadrupole, probe, and circular post, as shown in Fig. 5A. At low field-strengths, chaining structures are barely generated, whereas crystal formation is initiated at high field-strengths. Kim¹²⁴ *et al.* used nDEP to immobilize a single-crystal domain on a glass substrate without losing its crystallinity and resorted to a larger gravitation force of 10 μm PS beads to allow the formation of defect-free single-crystal domains. A quadrupole microelectrode dedicated for trapping single cells was designed and constructed on a multilayer LOC structure called the sandwiched insulator with a back contact (SIBC) biochip platform, in which 22% Ishikawa cancer cells and 17% 10 μm PS beads were successfully trapped.¹²⁵ Haapalainen and Mäkylen presented a microfluidic DEP platform with a hyperbolic quadrupole electrode geometry, which was implemented for particle characterization.¹²⁶ The electrodes of the implemented platform were made of ITO to achieve full transparency and






Fig. 5 Trapping modes of the micro-scaled PS beads. (A) Structure of crystals: (a₁) 10 μm PS beads,¹²⁴ (a₂) 2 μm PS beads,¹²⁶ and (a₃) 2 μm PS beads by iDEP.¹³⁰ (B) Structure of chains: (b₁) 10 μm PS beads in the ratchet microchannel,¹³⁶ (b₂) 10 μm PS beads in a combined acoustic DEP system,⁷⁰ and (b₃) 18 μm PS beads in the curved microchannel.¹³⁸ (C) (c₁) Structure of wires: 10 μm PS beads in X-patterned insulating structures,¹⁴⁰ (c₂) Structure of arrays: 10 μm PS beads in the cello-type mechanical trap.¹³²

consequently, better view of the particle motion using common transillumination microscopy. Freedman *et al.* tracked the 1 μm PS beads used for DNA sensing to demonstrate the distinct forces observed for both DEP and electrothermal flow (ETF).¹²⁷ Mohammadi's group studied a geometrical relationship to optimize the efficient geometry of the post array distribution in insulator-based DEP (iDEP)¹²⁸ (also known as contactless DEP (cDEP)^{129,130}). They presented microfilters, with a similar footprint and channel thickness, that maximized the trapping condition while minimizing the required voltage. The obtained results indicated that a post radius of 40 μm, larger than the transversal distance between posts, could enhance the trapping, which supported and extended a master scaling for the field strength efficiency from 56% to 341%. McMullan *et al.* reported the formation and kinetics of DEP-directed self-assembled ordered structures *via* a combination of microscopy and light scattering over an extended range of particle sizes, electrokinetic properties, and frequencies required to induce particle chaining.¹³¹ The high field applied in cDEP and ODEP increases the trapping efficiency of the PS beads. However, these techniques have some drawbacks, especially for bioparticles. Implementation of a high electric field has some disadvantages. The complicated fabrication procedures reduce the layer lifetime. Joule heating and an increase in temperature were caused by a highly conductive biological fluid and high electric field intensity (*e.g.*, 500 V). Additionally, manipulation of cells is difficult with iDEP and ODEP due to their collecting patterns and devices. The trapping of PS beads can reach high efficiency

by iDEP and ODEP; however, this is not the case for bioparticles, which are still challenging.

(2) *Lined chains.* One of the interesting structures of the PS beads trapped is chains, located in the electrode gaps by nDEP and the interaction force of beads, as shown in Fig. 5B. An interdigitated electrode is the most common geometry used. Cui and Morgan presented a generic twDEP separation system to handle a complete particle and fractionate system.⁵⁶ Chen *et al.* have demonstrated that the levitation height of a specific PS bead strongly depends on the combined contributions of parameters such as the frequency of the electric field, dielectric properties of PS beads and suspension medium.³⁴ Ravula *et al.* discussed the modeling, fabrication and characterization of a platform that combined acoustic forces and AC DEP, in which 10 μm PS beads were further focused into single file bead streams (b₂ of Fig. 5).⁷⁰ Kale *et al.* presented an experimental study of particle trapping in an asymmetric ratchet microchannel under DC-biased AC electric fields.¹³⁶ The DC/AC DEP accumulation of PS beads in the first pair of ratchets and the DC electrokinetic shifting of particles into the second and subsequent ratchets were studied, which depended on the moving direction of beads with respect to the asymmetric ratchets. The PS beads were trapped into triangular zones in all, but later they were trapped in the first pair of ratchets for both forward and backward motions. The trapping efficiency was 100% in this ratchet microchannel. Allen *et al.* demonstrated both pDEP and nDEP using silver-coated hollow glass spheres and PS beads, respectively, with isomotive DEP (isoDEP) devices.¹³⁸ Both AC/



DC electric fields could trap the PS beads by nDEP with a trapping efficiency of approximately 95–100%.

(3) *Wires and arrays.* Another similar interesting structure is wires. The PS beads are placed in a planer-electrode gap by nDEP, as shown in Fig. 5C. Jen's group examined the feasibility of using cDEP in an iDEP microdevice to effectively trap 10 μm and 13 μm PS beads (c_1 of Fig. 5C).¹⁴⁰ Su and Voldman developed an automated system and presented characterization of the method using 6 μm and 10 μm PS beads and HL-60 cells as well as its application to rapidly discriminate neutrophils with different activation states.¹⁴⁴ Carboxyl-modified PS beads were used for modelling validation and calibration. The system could accumulate large datasets for different cell types for electrical studies of cells and applications of label-free DEP separation. Guler *et al.* utilized 3D microwire electrodes to detect 6 μm PS beads and erythrocytes using an LCR meter without any additional electronics (*e.g.*, lock-in amplifiers and custom-made electronic circuitries).¹⁴¹ This yielded very high sensitivity as compared to the case of conventional coplanar electrodes, which require several repetitive cleanroom processes for the fabrication of micro-electrodes. This provided a method for integrating electrodes into a standard soft lithography process.

A set of electrode arrays accurately fabricated by MEMS technology was used for trapping the PS beads by pDEP, as shown in Fig. 5C. The reversibility of the guiding and trapping effectiveness was predicted using the 1 μm , 2 μm and 10 μm PS beads.^{143,145} The PS bead flow was up to roughly 200 $\mu\text{m s}^{-1}$ after approximately 2 min as the voltage was changed to 7 V_{pp} , and 98% of the PS beads were released when the voltage applied was turned off. An iDEP study clearly indicated the effects of particle size and shape on DEP trapping by employing 1 μm and 2 μm PS beads and *E. coli* cells.¹⁴⁶ The PS bead size and flow speed have a significant effect on the magnitude, location, and shape of the DEP trapping regions, which are defined by the DEP isovelocity lines and EK isovelocity lines (c_2 of Fig. 5C). Both AC/DC electric fields can trap the PS beads by pDEP with an approximately 63% efficiency.

DC-DEP enables highest trapping efficiency for PS beads due to its high electric field. However, bioparticles will be overcooked by its high electric field. Although DEP manipulation of particles and molecules has been extensively performed using microelectrodes or sharp tips, some hurdles still exist such as heat and bubble generation and unwanted surface reactions due to high voltages. Additionally, the DEP devices used for micro-particles require a special electrode design and some

specific instruments. Especially, plenty data needs to be obtained to achieve statistically reliable results; however, this will consume a lot of time.

3.1.2. Nano-scaled PS beads. Considering that many important biological subjects, such as protein molecules and viruses, are on the order of 10 nm in size, it is desirable to maximize the trapping efficiency and overcome the Brownian motion of particles for the DEP electrodes with equally small gap sizes. Smaller gaps can create stronger electric field gradients to reduce the voltages applied for DEP trapping, avoiding unwanted surface electro-chemical reactions, bubble formation, and heat generation, all of which are major challenges for broader DEP applications. The validation and calibration of research on nano-scaled PS beads are summarized in Table 5, and the experimental evidence of manipulation is shown in Fig. 6.

(1) *Joule heating-aided.* Lewpiriyawong *et al.* demonstrated that a DC-AC offset reduced the Joule heating by lowering the necessary voltage; however, insights into temperature were not provided.³⁵ Kale's group utilized a 3D numerical model to study Joule heating and its effects on the coupled transport of heat, where 0.59 μm fluorescent and 3 μm PS beads were applied in an iDEP device with a rectangular constriction micro-channel.^{40,147} Buffer conductivity variations should be considered to determine the magnitude of heating effects by both simulations and experiments.¹⁴⁸ The nDEP behaviour of inert 1 μm PS beads was observed in the experiments under the reported conditions, and the particle-trapping capacity of the device was observed to decrease due to heating of the buffer inside the microchannel by a DC voltage, as shown in Fig. 6A.

(2) *Nanofabrication-aided.* Tanaka *et al.* reported the use of DEP to manipulate single-carboxylated PS beads passing through the electrode-embedded pore structures. The variation in the tunnelling current was expected to diminish by controlling the molecular conformations residing in the electrode gap through the DEP mechanism.¹⁴⁹ Barik *et al.* have demonstrated an ultralow-power DEP using nanogap electrodes fabricated *via* atomic layer lithography to trap the 30 nm PS beads, which can potentially enable high-density integration on a chip and portable biosensing, as shown in Fig. 6B.³⁶

(3) *Affinity cytochemistry-aided.* Kirmani *et al.* reported a DEP spectroscopy label-free immunoassay for rare analyte quantification in biological samples with the capability to detect and quantify about 850 avidin molecules attached to biotin-functionalized PS beads, as shown in Fig. 6C.¹⁵⁰ The

Table 5 Trapping of the nano-scaled PS beads

Type	Scale (μm)	Suspending medium (mS m^{-1})	Electric field	Time	Reference
CE-DEP	0.39; 0.46; 0.59; 0.62	KOH and MOPS (11–20)	2 V_{pp} ; 100 kHz to 2 MHz	—	54
SAW-DEP	0.19	KCl (0.2–1)	7 V_{RMS} ; 10 Hz to 10 MHz	9 s	153
DC-iDEP	0.59	PBS (47)	100–850 V; 1 kHz	4 s	40
DC-iDEP	1	DI water (0.2); K_2HPO_4 (10)	100–1500 V	30 s	148
IM-DEP	0.78	Tris-EDTA (20)	0.2–1.0 V; 2–10 MHz	1 s	149
AC-DEP	0.03; 0.19	DI water (0.4); PBS (18.2)	0.2–1.5 V; 100 kHz to 10 MHz	60 s	36
DEP	0.74	PBS (10)	500–2000 kHz	5 s	150
AFM-DEP	0.2	DI water	5 V_{pp} ; 10 kHz	10 s	151
AFM-DEP	0.02	DI water, PBS and Tris-EDTA (10–490)	120–150 V; 1 MHz	60 s	152



electrodes were visible in darker regions, and a bright layer on their edges was formed by the accumulation of PS beads. Moreover, improved automated quantitative measurement of DEP was achieved through image processing, and DEP spectrum curves were obtained for two different concentrations of avidin–biotin conjugates over a range of frequencies.

(4) *AFM-aided*. Zhou's group presented a novel method of applying AFM tip-induced DEP (AFM-DEP) for the manipulation of 200 nm PS beads, which were assembled into various nanostructures, including lines, ellipsoids and arrays of dots, as shown in Fig. 6D.¹⁵¹ Both the size and shape of the assembled structures were controllable by various experimental parameters; this demonstrated the potential of AFM-DEP for applications in the fabrication of nanostructures and arrays, assembly of nanodevices, and non-destructive manipulation of biological nanoparticles. Recently, a coaxial AFM nano-probe device has been studied for DEP trapping of DNA molecules in Tris–EDTA and PBS buffers.¹⁵² The 20 nm PS beads were concentrated in the high-field region of the AFM probe end due to the pDEP force in low-conductive solution environments, and they experienced nDEP force in a high-conductive buffer solution.

The microelectrode geometries used to manipulate nanoparticles often require complex and expensive multiple layer fabrication. Moreover, for nanoparticles, some of the DEP devices require precision fabrication (*i.e.* atomic layer lithography), and the fabrication procedures are not suitable for mass production.

3.1.3. Functionalized PS beads. As abovementioned, PS beads have stable physical and chemical properties and are widely used in microfluidics devices *via* DEP. When the PS beads are coated with carboxylate groups, the surface potential of the negative charges will increase. When contacted with an amine group (a positive charge), they may create an electrostatic bond and form a larger macromolecule. The net charges not only affect the polarization strength of the suspending medium, but also influence the electric field force.¹⁵⁴ The carboxyl-functionalized PS beads were covalently coupled to ethanol-amine using a single-step reaction, especially rendering the beads uncharged and hydrophilic. The ethanol layer on the particle surface acts to reduce the hydrophobic nature of the PS beads. Spherical PS beads with surface isotropic chemical groups have been widely investigated *via* their structure–property relationship and self-assembly behaviour. Protein–protein interactions and bacterial targeting or immobilization can be extended to the detection of rare biomarkers using PS beads functionalized with appropriate antibodies. These PS beads can be further fabricated as anisotropic particles such as Janus particles (JPs), having two distinct surface chemical compositions or properties on two sides. JPs are applied in a wide variety of novel material properties as well as intriguing interparticle interactions.¹⁵⁵ The trapping of functionalized PS beads is summarized in Table 6.

Javanmard's group experimentally tested a novel microfluidic platform that demonstrated the ability of nDEP to elute




Fig. 6 Examples of the trapping of the nano-scaled PS beads. (A) Particle streak images in the microchannel constriction region of an iDEP device under various DC-biased AC (fixed at 1 kHz) voltages: (a₁) 600 V DC and (a₂) 20 V DC/580 V AC. The arrowed loops in the 20 V DC/580 V AC case indicate the directions of electrothermal flow circulations.¹⁴⁸ (B) Nanogapped DEP electrode: (b₁) schematic of an array of nanogapped electrodes, each of which were individually addressable for DEP trapping. (b₂) The beads were attracted to the gap in pDEP. (b₃) and (b₄) SEM images of trapped 190 nm PS beads along the nanogap at different magnifications. Scale bars: 20 μm, 2 μm, and 400 nm.³⁶ (C) Quantification of biotin-functionalized PS beads: (c₁) and (c₂) visible darker regions and bright layer edges were electrodes formed by the accumulation of PS beads, respectively. (c₃) nDEP spectrum of biotin-functionalized PS beads with 0% and 0.8% of biotin conjugation. (c₄) Light intensity as a function of the distance from the electrode edge. The red scale bar indicates 50 μm.¹⁵⁰ (D) Different assembled nanostructures, 2D AFM images and AFM scanned profiles: (d₁) basin-shaped, (d₂) linear, (d₃) ellipsoidal, and (d₄) nanodots.¹⁵¹



Table 6 Trapping of functionalized PS beads

DEP type	Scale (μm)	Suspending medium (mS m^{-1})	Electric field	Time	Reference
nDEP	Protein-6.7; IgG-7.4	PBS; NaOH (200)	10–20 V_{pp} ; 10 MHz	10 min	156
ICEP-DEP	3.8, JP-3.8	NaCl (0.7–116)	6 V; 40 kHz	17 min	90
ODEP	10; 15	DI water (2.14), NaCl (95/106)	5–7.5 V_{pp} ; 5 kHz to 1 MHz		157
SERS-DEP	AuNPs-1	EDC, PBS	—	2 h	158
DFS-DEP	Func-15	DI water (0.64)	0–2 V_{pp} ; 10–100 Hz	0.04 s	159
Raman-DEP	Flu-3.3	PBS (4)	350 V_{rms} ; 100 Hz	25 s	160

specifically bound beads with a switch-like behavior, as shown in Fig. 7A.¹⁵⁶ nDEP in conjunction with shear force was used at an optimal NaOH concentration to illustrate the DEP responses in a singleplex assay. With nDEP off, the flow rate was increased to $0.95 \mu\text{L min}^{-1}$, and the majority (70%) of the beads was removed. This platform offers the potential for performing a bead-based (6.7 μm protein-G-coated PS and 7.4 μm goat IgG-coated PS) multiplexed assay in various regions of a single channel. Different antibodies were immobilized to target the corresponding antigens. The DEP characteristics of non-functionalized and carboxyl-functionalized 5 μm PS beads were investigated in solutions with different conductivities by associating the measured crossover frequencies with a theoretical DEP model. In the experiments, specific (streptavidin-biotin) and nonspecific (NH_3^+ -biotin) interactions and two ligand-receptor interactions (streptavidin-biotin and avidin-

biotin) were simultaneously evaluated, and different individual bond-rupture forces associated with the measured values were clearly determined.¹⁵⁷

Peptide coupling on AuNP-decorated 1 μm PS beads by DEP enabled the surface-enhanced Raman scattering (SERS) detection of glycine. DEP-enabled immobilization of higher-order nanostructures for Raman spectroscopy of a specific analyte was demonstrated by linker bridges on fabricated AuNP-decorated PS beads, as shown in Fig. 7B.^{158,161} A similar method was also demonstrated by an automated DEP tweezer-based force spectroscopy system to examine the intermolecular weak binding interactions.¹⁵⁹ In the experiment, according to the numerical expectation, the carboxyl-terminated 15 μm PS beads were arranged through the center of the IDT electrodes covered by silicon dioxide with an nDEP force, and then, the rupture event of the arranged particles occurred since the




Fig. 7 Trapping modes of the functionalized PS beads. (A) Percentage distribution and detachment time profile of the detached beads in 0.2 M NaOH. Upon turning the nDEP on, about 80% of the beads detached.¹⁵⁶ (B) Characteristic Raman spectra of glycine immobilized on AuNP-decorated PS beads with linked bridges via peptide coupling and the blank (ultrapure water with AuNP-decorated 20 nm PS beads and no analyte).¹⁵⁸ (C) The measured unbinding (rupture) forces of 15 μm PS beads: (c₁) hydrogen bond and (c₂) van der Waals interaction.¹⁵⁹






Fig. 8 Separation of the PS beads and mixed particles. (A) Time-based mode I, with different electrodes acting as sample plugs, which could lead beads into the separation region and split into two species or more. The species were separated in time. (B) Space-based mode II, in which a sample was continuously led in a separation region and species were spatially separated. (C) Separation of multiple-sized PS beads. Mode I:¹⁸ (c₁) 1 and 5 μm PS beads and (c₂) 1, 6 and 15 μm PS beads. Mode II: (c₃) 3, 5 and 10 μm PS beads¹⁶⁹ and (c₄) 3, 10 and 25 μm PS beads.¹⁶⁸ (D) Separation of PS beads and non-bioparticles. Mode I: (d₁ and d₂) 1 μm PS beads with SWCNTs.⁵⁵ Mode II: (d₃) 5 μm PS beads and magnetic beads¹⁷³ and (d₄) 15 μm PS beads and 14 μm sliver-coated hollow glass beads.¹⁷³ (E) Separation of PS beads and bioparticles. Mode I: (e₁) 7 μm PS beads and yeast cells,¹⁷⁴ (e₂) carboxylate-modified PS beads (red) and live *E. coli* (green),¹⁴⁵ (e₃) PS beads and K562 cells.⁴¹ Mode II: (e₄) 3 μm PS particles and yeast cells¹⁷² and (e₅) 5 μm PS beads and algae.⁹³

vertical nDEP force was increased by increasing the applied AC voltage. The rupture forces of the weak binding interactions (e.g., van der Waals interaction and hydrogen bond) were measured by the developed system and compared with the previous results, as shown in Fig. 7C. Glycine was bound with the PS beads through peptide coupling induced *via* EDC. More recently, Hanson and Vargis presented an alternative method to fabricate a cDEP device, allowing higher operating voltages, improved replication, and the opportunity for analysis using Raman spectroscopy.¹⁶⁰ The cDEP devices were capable of simultaneously trapping and analyzing the 3.3 μm PS beads *via* Raman spectroscopy; however, their fabrication, replication, and reuse were difficult. In addition, their operating parameters

were limited by the dielectric breakdown of polydimethylsiloxane (PDMS).¹⁶² However, cDEP is still a powerful label-free tool to sort biological samples without tags, fluorescent markers, or specific DNA sequences for subsequent identification. It avoids common problems associated with DEP (e.g., electrode fouling and electrolysis).

Functionalized PS beads with antibodies can be applied in the detection and quantification of rare analytes for the prevention and treatment of diseases such as cancer and myocardial infarction. In addition, additional expertise and equipment are required to characterize the functionalization performance of functional particles. Surface micromachining



techniques allow electrodes to be very close to each other (*e.g.*, tens of micrometers).

3.2. Separation of multiple-sized PS beads

In addition to trapping, the DEP force is very suitable for size fractionation, which can improve the screening performance of varied sized particles when microelectrode arrays are used or when it is combined with other devices (*e.g.* microfluidic platforms). The size of the PS beads is similar to that of cells such as bacterial cells (0.8–2 μm), red blood cells (7–8 μm), liver cells (20–30 μm) and many cancer cells (10–30 μm). Separation of multiple-sized PS beads can provide prediction for PS beads and mixed particles.

3.2.1. Characterization of multiple-sized PS beads. Levitation is vital for the separation of multiple-sized PS beads. The levitation height results from the resultant force, such as nDEP and gravity, acting on particles in a vertical direction. Efficient positioning of outlets for different particles depends on the accurate prediction of the levitation height of particles. The levitation height is dependent on the CM factor, where larger PS beads are deflected more than smaller particles. Multiple-sized PS beads in suspensions are consequently separated into different transverse positions (or level heights) by size as they flow down into a microfluidic device along its length. In addition, the surface conductivity of PS beads is inversely proportional to the radius of the PS beads. The surface conductivity is significantly larger than bulk conductivity. Therefore, the CM factor of the 1 μm beads is smaller than that of the 10 μm beads. The levitation height of 1 μm beads is smaller than that of the 10 μm beads. As a result, 1 μm beads are dragged to the electrode edges due to their lower levitation height. Tunable multiplex separation of multiple-sized PS beads is classified into two modes: time-based separation mode I (Fig. 8A) and space-based separation mode II (Fig. 8B), as detailed in Table 7.

3.2.2. Multiple-sized PS beads in mode I. In the time-based separation mode I, PS beads are trapped due to frequency changes. The DEP electrode architectures act as sample trappers, inducing the PS beads to the separation regions by pDEP. Multiple-sized PS beads split into two species or more, as shown

in Table 7 and Fig. 8C. The species are thus separated in time, and the distance in the *y* direction will continue to respectively separate and accumulate various PS beads. Khoshmanesh *et al.* utilized curved microelectrodes to separate 1, 6 and 15 μm PS beads at the frequencies of 100 kHz, 200 kHz, and 20 MHz, respectively, as shown in Fig. 8(c₁ and c₂).¹⁸ Lin *et al.* utilized a circular micro-electrode array to separate the PS beads.³⁸ Herein, 83% of 10 μm beads were trapped in a circular band at a radius ranging from 100 to 120 μm , and 91% of the 6 μm beads were located between from 380 and 400 μm radii. More recently, Zhang *et al.* fabricated a DEP-enhanced microfluidic impedance biosensor for pathogen detection.¹⁶³ Pathogens were isolated from samples using PS immune-beads, followed by DEP-assisted pathogen capture at boron-doped ultrananocrystalline diamond (BD-UNCD) microelectrodes and quantification of the capture using impedance spectroscopy. The carboxylate-modified beads with poly(ethylene glycol) (PEG) provided the highest capture efficiency (65%) and selectivity (95%) in isolating live *E. coli* O157 : H7. Higher specificity was achieved upon the addition of PEG to the antibody-functionalized bead surface, the highest specificity was achieved with the epoxy-sulfate beads (85–86%), followed by carboxylate-modified beads (76–78%) and aldehyde-sulfate beads (74–76%). Romero-Creel *et al.* demonstrated how the PS beads that were of the same size and shape and made from the same substrate material could have different DEP behaviors as a result of their distinct surface charge magnitudes.¹⁶⁴ The red-carboxyl beads in the mixture were considered to have a higher surface charge than the green-amine beads. Another DEP-based PS bead sensing system was recently developed by Yahya for separating 3 and 10 μm PS beads in an array of tapered square electrodes.¹⁶⁵

3.2.3. Multiple-sized PS beads in mode II. The other separation mode is the space-based mode II, which can synchronously separate PS beads with various sizes by nDEP at the microelectrodes with shapes of sharp tips, triangular, planar and wires flowing to different outlets at the same time, as shown in Table 7 and Fig. 8C. Previously, Kralj *et al.* used DEP to separate a mixture of 4, 5 and 6 μm PS beads, with the flow rate

Table 7 The separation of multiple-sized PS beads

Mode	Type of DEP	Scale (μm)	Suspending medium (mS m^{-1})	Electric field	Ref.
I	AC-DEP	1/5; 1/6/15	DI water (0.2)	30 V, 10 kHz to 20 MHz	18
	AC-DEP	PS: 10/15; streptavidin-terminated PS: 10/12/15	DI water	5–7.5 V; 5–1000 kHz	157
	AC-DEP	6/10	KCl (1.8)	300–3000 Hz	38
	AC-DEP	PS: 4; Dynal magnetic beads; goat antimouse conjugated beads	PBS	6 V; 40 kHz	163
	DC-DEP	0.1/0.2/0.5/1	NaCl (3.5)	DC: 1300 V; AC: 2800–3600 V _{pp} ; 1 kHz	164
	AC-DEP	3/10	DI water	10 V, 0–1.5 MHz	165
II	DC-DEP	1.97/4.84	NaCl; PBS	100 V	166
	DC/AC-DEP	0.02/0.1	PBS	DC: 5–50 V; AC: 600 Hz	167
	AC-DEP	3/10/25	DI water (0.15)	120–150 V, 1 MHz	168
	DC-DEP	3/5/10	PBS	600–1320 V	169
	DEP	11/25/45	Milli-Q water	200 V, 200 kHz	170
	DEP	4/5/6	DI water	4–12.5 V, 1 kHz	39
	DC-DEP	PS: 3/5; Janus: 3/5	–50	120–290 V	171



and voltage parameters from the model used as initial estimates for experiments.³⁹ Zhu *et al.* demonstrated the continuous-flow electrokinetic separation of 1 and 5 μm PS beads in a serpentine microchannel through curvature-induced DEP.¹⁷² Sun *et al.* separated 1.97 and 4.84 μm PS beads through the integration of a metallic oxide semiconductor field-effect transistor (MOSFET)-based microfluidic Coulter counter with a DC-DEP cell sorter.¹⁶⁶ Viefhues *et al.* presented a microfluidic device capable of performing two tasks: it could be operated in either mixing or demixing mode with 20 and 100 nm PS beads.¹⁶⁷ These parameters could be adapted in real time, and the continuous separation and mixing efficiency of 85–100% was achieved. A similar method was demonstrated by separating both a binary and ternary mixture of 5 and 10 μm PS beads and a ternary separation of 3, 5 and 10 μm PS beads by intrinsic properties (*e.g.*, size and surface charge) in a T-shaped microchannel, as shown in Fig. 8(c₃).¹⁶⁹ Tao *et al.*¹⁶⁸ used acupuncture needle electrodes to facilitate the integration of working electrodes in a DEP system. Multiple-sized PS beads (3, 10 and 25 μm) were separated and moved to different outlets by DEP, and a purity efficiency of over 90% for the 10 μm PS beads was achieved, as shown in Fig. 8(c₄). Wang *et al.* developed a sheath flow-assisted DEP field-flow separator with a tailored arrangement of cylindrical IDT electrodes and observed the size-dependent trajectories of dispersed particles.¹⁷⁰ Using a voltage of 200 V at a frequency of 200 kHz, 11, 25 and 45 μm PS beads were levitated to different heights along the channel length due to the nDEP forces.

The DEP separation of multiple-sized PS beads only depends on the surface conductivity of these beads. For particles with different sizes, separation by DEP based on their electrical properties (*e.g.* conductivity) and suspending medium requires specific integrated and accurate electrodes arrays for improved separation efficiency. The limits (*e.g.*, design limitations and geometric tolerances due to the fabrication process) and stability of the separation modes play a major role in state-of-the-art applications with complex mixtures. The time-based mode I seems to be more suitable for separating particles of different sizes than model II. However, for particles of similar sizes, the separation by DEP based on their electrical properties and suspending medium requires discrete processes and becomes challenging.

3.3. Separation of the PS beads and non-bioparticles

Separation of the PS beads and other particles (organic and inorganic) has been discussed as follows. For inorganic particles in mode I, Zhang and Khoshmanesh presented a platform for separating 1 μm PS beads and MWCNTs (20–50 nm in diameter and 0.5–2 μm in length) according to their dielectric response to alternating electric fields at specific frequencies (about 150 kHz), as shown in (d₁–d₂) of Fig. 8D.⁵⁵ More recently, Zhao *et al.* presented a nano-orifice-based microfluidic device using a DC-DEP method to continuously separate different types of micro and nanoparticles, which included 5.2 μm magnetic-coated PS beads, 7 μm fluorescent PS beads, 14 μm silver-coated hollow glass beads and 15 μm plain PS beads, with

similar sizes by their different electric conductivities in a pressure-driven flow.¹⁷³ They also numerically demonstrated the separation of 5 μm Janus particles and homogeneous PS beads and the separation of 3 and 5 μm Janus particles using DC-DEP in the mode II, as shown in Fig. 8(d₃).¹⁷¹

3.4. Separation of the PS beads and bioparticles

The separation of bioparticles from a complex mixture is important for numerous applications in cell sorting, infection diagnosis, food safety, and enrichment of particle populations for drug development. Separation is accomplished by the intrinsic physical properties of a cell type, which defines a specific finger print that can be expressed as an induced force that drives the separation of cells in complex mixtures.

3.4.1. Characterization of bioparticles. Commercialized cell sorter systems, such as fluorescence-activated cell sorting (FACS) and cell labeling, require tedious cell preparation, whereas DEP is a simple and cost-effective technique that can be implemented in this area. DEP-based cell separation has significant promise for the separation of cells from heterogeneous mixtures based on their electrical properties and is used in cell purification, such as in the purification of human promyelocytic leukaemia cells (HL-60)³⁸ and Chinese hamster ovary (CHO) cells,¹⁷⁵ or target cell isolation. The DEP technique offers advantageous identification and differentiation of target cells as compared to the conventional FACS approach. In the DEP rapid detection, various target biological particles (such as antigens, antibodies, cells, and DNA) can be trapped, concentrated, separated, and purified from multiplex media, including inorganic impurities (*e.g.*, silica). The PS beads act as a known inorganic particle, which is suitable for the preliminary investigation of cell purification due to their various sizes, surface area and morphology.

The separation of PS beads and organic particles (*e.g.*, cells) from a fluid is a vital step in many biochemical tests. PS beads and numerous cells have distinctive differences in their dielectric properties, rendering their COF obviously different. For the PS beads, pDEP occurs from 10 to 500 kHz, whereas *E. coli* under pDEP between 20 kHz and 20 MHz moves towards the direction of a high electric field. Beyond 20 MHz, both the PS beads and bacteria suffer nDEP and are pushed towards the electrode gap, where the minimum field exists. Between around 500 kHz and 20 MHz, the PS beads exhibit nDEP, whereas *E. coli* exhibits pDEP, and thus, separation is possible in this frequency region.¹⁸⁰

3.4.2. PS beads and bioparticles in mode I. In the time-based separation mode I illustrated in Fig. 8A, the PS beads usually flow away from, whereas the bioparticles will be trapped at the electrodes, as summarized in Table 8 and shown in Fig. 8(e₁–e₃). Tang demonstrated the capability of a developed system by studying the interaction of viable yeast cells with micro/nano materials including 850 nm PS beads and MWCNTs.¹⁷⁶ A microfluidic AC-DEP chip for yeast cell separation was fabricated using screen printing technology.¹⁷² This is especially suitable for high-throughput mass production due to its low cost, a simple operating procedure and facile fabrication



Table 8 The separation of PS beads and bioparticles

Mode	DEP type	Scale (μm)	Suspending medium (mS m^{-1})	Electric field	Time (min)	Efficiency	Ref.
I	AC-DEP	PS (0.85) and yeast	PBS (30)	—	5	—	176
	DC-iDEP	PS(0.5/1/2) and yeast	DI water and KOH (2)	440–3000 V	4	99%	177
	AC-DEP	PS (7/20) and yeast	Aqueous (0.5–10)	10 V, 100–200 kHz	5	94–96%	174
	F-DEP	PS (10) and WBC (K562 cell)	PBS (80)	7.5 V, 800 MHz	4	93–100%	41
II	DC-DEP	PS (3) and yeast	PBS (20)	0–200 V	15	90%	172
	AC-DEP	PS (2.9) and yeast; PS (2.9) and <i>E. coli</i>	NaCl (38; 60)	31.2 V, 300 kHz	6.2	97%	178
	DC-DEP	PS (5) and algae	Sodium borate buffer (10)	0–192 V	—	—	93
	AC-DEP	Gold-coated PS (25) and yeast	KCl (1.4)	12.5–25 V, 1 MHz	5	90–100%	179

conditions as compared to the conventional micro-fabrication process, as evidenced in Fig. 8(e₁). The results showed that a high capture rate (95%) and separation efficiency (94–96%) could be achieved under the optimized conditions. Lapizco-Encinas's group focused on the separation of different particles (*e.g.*, PS beads and yeast cells¹⁷⁷ and PS beads and *E. coli*¹⁴⁵ by iDEP, as shown in Fig. 8(e₂)). The results demonstrated the successful and stable capture and enrichment of the rare PS beads and cells (efficiency >99%). Sadeghian *et al.* presented a design and optimized flowchart for the separation of PS beads and white blood cells (WBCs), and the recovery and purity efficiency of 93% and 100% were achieved, as shown in Fig. 8(e₃), respectively.⁴¹

3.4.3. PS beads and bioparticles in mode II. In the space-based separation mode II illustrated in Fig. 8B, the PS beads and bioparticles are separated based on their different DEP behaviours, as summarized in Table 8 and shown in Fig. 8(e₄ and e₅). Previously, Lewpiriyawong *et al.* have presented a PDMS microfluidic device that utilizes AC-DEP *via* 3D Ag PDMS composites as sidewall electrodes for cell characterization and continuous separation of particles and cells of similar sizes.¹⁷⁸ The 2.9 μm PS beads and *E. coli*, 5 μm PS beads and yeast cells, and especially live and dead yeast cells were separated based on their different polarizabilities and DEP behaviors, and a high separation efficiency (>97%) was achieved. However, these devices suffered from particle adhesion on the microelectrode surface, fabrication complexity, and chemical reactions on the electrode surface. Zhu demonstrated the continuous-flow electrokinetic separation of 3 μm PS beads and yeast cells, with a separation efficiency of both over 90%, as shown in Fig. 8(e₄).¹⁷² Similarly, Song *et al.* reported the DC-DEP separation of two types of algae by their sizes and separation of algae and 5 μm PS beads by their different dielectric properties, as shown in Fig. 8(e₅).⁹³ Jia *et al.* further presented a microfluidic device for the separation of 25 μm gold-coated PS beads (100% separation efficiency) and yeast cells (90% separation efficiency).¹⁷⁹ Utilizing a range of frequencies, distinct DEP responses were observed. As a result, the particles moved away from the electrodes; this reduced the bioparticle viability *via* the nDEP effects of the electric field.

For a typical DEP system, high separation efficiency is one of the most significant performance features that can be achieved by quantitative evaluation of the PS beads and bioparticles. Although the efficiency of the two abovementioned modes is

over 90%, that of the time-based mode I is higher (>93% on average). Not only separation efficiency, but also time is one of the important factors in separation. Cells exposed to a buffer solution for long time will suffer from biological deactivation due to the lack of adequate nutrition. By contrast, the space-based mode II takes less time to carry all particles out of the microfluidic channel than the time-based separation mode. Moreover, the microstructure of bioparticles is protected by an appropriate shape of electrodes and AC electric field, whereas a DC electric field is harmful to the cell viability. Therefore, the space-based mode II is more suitable for the separation of bioparticles by DEP.

4. Conclusions and future perspectives: DEP-on-a-chip

This review highlighted the recent studies carried out on the DEP manipulation of various scaled particles. In addition, modelling and simulation aspects helped in the better design and optimization of process parameters for the effective manipulation of particles. The systematic studies on the PS beads, especially nanoparticles and functional particles, were highlighted. The detailed descriptions on the trapping and separation of PS beads and other particles (non-bioparticles and bioparticles) were addressed clearly, which provide great potential for the *in situ* monitoring of bioparticles by DEP. There is also a vast variety of opportunities to design advanced protocols for the handling of cells and particles. However, compared to the case of conventional techniques, the throughput of DEP systems is still low, which needs to be increased for the commercialization of these systems. Although the microfluidic devices to handle DEP manipulation are simple, these instruments and the preparation of samples to run the system may require technical skills. Moreover, the instruments needed to run the system may be complex and bulky. The seamless integration of the different components will determine the usability, portability, and simplicity of manufacturing and costs.

Real-time detection of pathogenic microorganisms in water, food, and airborne sources can prevent food poisoning and respiratory infections. The DEP technology precisely discriminates bio-targets, such as DNAs, proteins, viruses and bacteria, from a sample containing non-target species. The major issue is that individual biomolecules have very similar and very small






Fig. 9 A DEP-on-a-chip platform advanced by IM, QCM and SAW in the network of microfluidic channels.

induced dipole moments, and a high field gradient is required to generate a substantial DEP force.¹⁸¹ Thus, further research on the design of DEP systems is required for the effective fluidic manipulation of multi-particles. Microfabrication techniques can help in achieving better designs, cost-effectiveness and efficiency. In addition, some quantitative technologies and methodologies, such as impedance spectroscopy (IM), quartz crystal microbalance (QCM) and surface acoustic wave (SAW), can be integrated into microfluidic channels to form DEP-on-a-chip, which will improve the efficiency and accuracy of separation and detection, as shown in Fig. 9.

IM allows continuous monitoring of the electrical properties, such as impedance amplitude, phase, conductance, and capacitance, of bio-samples in a microfluidic network for the characterization and differentiation of various types of particles in a frequency domain. QCM biosensors exploit the changes in mechanical properties of the interface between particles and an oscillating quartz crystal to detect micro-scaled particles (e.g., cells, bacteria and minerals). Moreover, QCM-based immunosensors¹⁸² are capable of detecting specificity and non-specificity by immobilization of selective antibodies.¹⁸³ Furthermore, SAW-based immunosensors¹⁸⁴ have a higher mass sensitivity to accurately detect nano-scaled particles.¹⁸⁵

The combination of high-throughput capabilities of SAW with the exquisite discriminatory capabilities of DEP and mass sensitivity of QCM in a DEP-on-a-chip system can be a key to the development of rapid, accurate, portable, simple and cost-effective microfluidic devices with a promising impact in multi-particle manipulation, microfluidic, immune analysis, micro-total analysis systems (μ TAS) and bond-rupture analysis systems.

Conflicts of interest

There are no conflicts to declare.

Acknowledgements

This study was supported by the National Key R&D Program of China (2016YFB1200401-102E) to JYJ.

References

- H. Towbin, T. Staehelin and J. Gordon, *Proc. Natl. Acad. Sci. U. S. A.*, 1979, **76**, 4350–4354.
- H. A. Pohl and J. S. Crane, *Biophys. J.*, 1971, **11**, 711–727.
- Z. Gagnon, J. Gordon, S. Sengupta and H. C. Chang, *Electrophoresis*, 2008, **29**, 2272–2279.
- K. C. Neuman and A. Nagy, *Nat. Methods*, 2008, **5**, 491–505.
- Y. Kim, S. Hong, S. Jung, M. S. Strano, J. Choi and S. Baik, *J. Phys. Chem. B*, 2006, **110**, 1541–1545.
- R. Krupke, F. Hennrich, H. von Lohneysen and M. M. Kappes, *Science*, 2003, **301**, 344–347.
- J. Gimsa, T. Muller, T. Schnelle and G. Fuhr, *Biophys. J.*, 1996, **71**, 495–506.
- J. Suehiro, T. Hatano, M. Shutou and M. Hara, *Sens. Actuators, B*, 2005, **109**, 209–215.
- D. Di Carlo, D. Irimia, R. G. Tompkins and M. Toner, *Proc. Natl. Acad. Sci. U. S. A.*, 2007, **104**, 18892–18897.
- E. Lauga and T. R. Powers, *Rep. Prog. Phys.*, 2009, **72**, 096601.
- X. Ding, S.-C. S. Lin, B. Kiraly, H. Yue, S. Li, I. K. Chiang, J. Shi, S. J. Benkovic and T. J. Huang, *Proc. Natl. Acad. Sci. U. S. A.*, 2012, **109**, 11105–11109.
- N. G. Green, in *Electrokinetics And Electrohydrodynamics In Microsystems*, ed. A. Ramos, 2011, vol. 530, pp. 61–84.
- N. G. Green and H. Morgan, *J. Phys. D: Appl. Phys.*, 1997, **30**, 2626–2633.
- R. Pethig, *Biomicrofluidics*, 2010, **4**, 022811.
- T. B. Jones and G. W. Bliss, *J. Appl. Phys.*, 1977, **48**, 1412–1417.
- W. M. Arnold, *IEEE Trans. Ind. Appl.*, 2001, **37**, 1468–1475.
- Y. J. Yuan, M. K. Andrews and B. K. Marlow, *Appl. Phys. Lett.*, 2004, **85**, 130–132.
- K. Khoshmanesh, C. Zhang, S. Nahavandi, F. J. Tovar-Lopez, S. Baratchi, A. Mitchell and K. Kalantar-Zadeh, *J. Appl. Phys.*, 2010, **108**, 034904.
- M. P. Hughes and H. Morgan, *Anal. Chem.*, 1999, **71**, 3441–3445.
- M. Duchamp, K. Lee, B. Dwir, J. W. Seo, E. Kapon, L. Forro and A. Magrez, *ACS Nano*, 2010, **4**, 279–284.
- D.-J. Kim, J.-H. Hyung, D.-W. Seo, D.-I. Suh and S.-K. Lee, *J. Electron. Mater.*, 2010, **39**, 563–567.
- C. S. Lao, J. Liu, P. X. Gao, L. Y. Zhang, D. Davidovic, R. Tummala and Z. L. Wang, *Nano Lett.*, 2006, **6**, 263–266.
- J. Regtmeier, T. T. Duong, R. Eichhorn, D. Anselmetti and A. Ros, *Anal. Chem.*, 2007, **79**, 3925–3932.
- M. Nakano, Z. Ding and J. Suehiro, *Jpn. J. Appl. Phys.*, 2016, **55**, 017001.
- M. P. Hughes, M. F. Flynn and H. Morgan, in *Electrostatics*, ed. D. M. Taylor, 1999, vol. 163, pp. 81–84.
- M. A. Mansor and M. R. Ahmad, *Int. J. Mol. Sci.*, 2015, **16**, 12686–12712.
- S. Kasetsirikul, J. Buranapong, W. Srituravanich, M. Kaewthamasorn and A. Pimpin, *Malar. J.*, 2016, **15**, 358.
- E. O. Adekanmbi and S. K. Srivastava, *Lab Chip*, 2016, **16**, 2148–2167.



- 29 C. Páez-Avilés, E. Juanola-Feliu, J. Punter-Villagrasa, B. del Moral Zamora, A. Homs-Corbera, J. Colomer-Farrarons, P. Lluís Miribel-Catala and J. Samitier, *Sensors*, 2016, **16**, 1514.
- 30 M. Viefhues and R. Eichhorn, *Electrophoresis*, 2017, **38**, 1483–1506.
- 31 C. Modchang, W. Triampo, P. Kanthang, D. Triampo and Y. Lenbury, *Anal. Chem.*, 2008, **53**, 851–862.
- 32 B. Cetin and D. Li, *Electrophoresis*, 2011, **32**, 2410–2427.
- 33 S. Dash and S. Mohanty, *Electrophoresis*, 2014, **35**, 2656–2672.
- 34 D. F. Chen, H. Du and W. H. Li, *Sens. Actuators, A*, 2007, **133**, 329–334.
- 35 N. Lewpiriyawong, C. Yang and Y. C. Lam, *Microfluid. Nanofluid.*, 2012, **12**, 723–733.
- 36 A. Barik, X. Chen and S.-H. Oh, *Nano Lett.*, 2016, **16**, 6317–6324.
- 37 A. T. J. Kadaksham, P. Singh and N. Aubry, *Electrophoresis*, 2004, **25**, 3625–3632.
- 38 S.-C. Lin, Y.-C. Tung and C.-T. Lin, *Appl. Phys. Lett.*, 2016, **109**, 053701.
- 39 J. G. Kralj, M. T. W. Lis, M. A. Schmidt and K. F. Jensen, *Anal. Chem.*, 2006, **78**, 5019–5025.
- 40 A. Kale, S. Patel, G. Hu and X. Xuan, *Electrophoresis*, 2013, **34**, 674–683.
- 41 H. Sadeghian, Y. Hojjat and M. Soleimani, *J. Electrostat.*, 2017, **86**, 41–49.
- 42 T. B. Jones and M. Washizu, *J. Electrostat.*, 1996, **37**, 121–134.
- 43 V. Giner, M. Sancho, R. S. Lee, G. Martinez and R. Pethig, *J. Phys. D: Appl. Phys.*, 1999, **32**, 1182–1186.
- 44 M. Sancho, G. Martinez and M. Llamas, *J. Electrostat.*, 1988, **21**, 135–144.
- 45 C. Xie, B. Chen, C.-O. Ng, X. Zhou and J. Wu, *European Journal of Mechanics, B: Fluids*, 2015, **49**, 208–216.
- 46 N. Couniot, L. A. Francis and D. Flandre, *Lab Chip*, 2015, **15**, 3183–3191.
- 47 Y. Ai, Z. Zeng and S. Qian, *J. Colloid Interface Sci.*, 2014, **417**, 72–79.
- 48 S. Kang, *Comput. Fluids*, 2014, **105**, 231–243.
- 49 L. Liu, C. Xie, B. Chen, N. Chiu-On and J. Wu, *Microsyst. Technol.*, 2016, **22**, 2223–2232.
- 50 W. M. Arnold, *Microsyst. Technol.*, 2008, **15**, 144–151.
- 51 J. L. Griffin and C. D. Ferris, *Nature*, 1970, **226**, 152–154.
- 52 H. Moncada-Hernandez, E. Nagler and A. R. Minerick, *Electrophoresis*, 2014, **35**, 1803–1813.
- 53 X. B. Wang, M. P. Hughes, Y. Huang, F. F. Becker and P. R. C. Gascoyne, *Biochim. Biophys. Acta, Gen. Subj.*, 1995, **1243**, 185–194.
- 54 C. M. White, L. A. Holland and P. Famouri, *Electrophoresis*, 2010, **31**, 2664–2671.
- 55 C. Zhang, K. Khoshmanesh, F. J. Tovar-Lopez, A. Mitchell, W. Wlodarski and K. Klantar-zadeh, *Microfluid. Nanofluid.*, 2009, **7**, 633–645.
- 56 L. Cui and H. Morgan, *J. Micromech. Microeng.*, 2000, **10**, 72–79.
- 57 F. Hennrich, R. Krupke, M. M. Kappes and H. V. Lohneysen, *J. Nanosci. Nanotechnol.*, 2005, **5**, 1166–1171.
- 58 M. Elitas, N. Dhar, K. Schneider, A. Valero, T. Braschler, J. D. McKinney and P. Renaud, *Biomedical Physics & Engineering Express*, 2017, **3**, 015005.
- 59 B. H. Lapizco-Encinas, B. A. Simmons, E. B. Cummings and Y. Fintschenko, *Anal. Chem.*, 2004, **76**, 1571–1579.
- 60 Q. Cao, S.-j. Han and G. S. Tulevski, *Nat. Commun.*, 2014, **5**, 5071.
- 61 S. Yan, J. Zhang, M. Li, G. Alici, H. Du, R. Sluyter and W. Li, *Sci. Rep.*, 2014, **4**, 5060.
- 62 D. Das, K. Biswas and S. Das, *Med. Eng. Phys.*, 2014, **36**, 726–731.
- 63 M. R. Buyong, F. Larki, M. S. Faiz, A. A. Hamzah, J. Yunas and B. Y. Majlis, *Sensors*, 2015, **15**, 10973–10990.
- 64 R. Hamada, H. Takayama, Y. Shonishi, L. Mao, M. Nakano and J. Suehiro, *Sens. Actuators, B*, 2013, **181**, 439–445.
- 65 J. I. Martinez-Lopez, H. Moncada-Hernandez, J. L. Baylon-Cardiel, S. O. Martinez-Chapa, M. Rito-Palomares and B. H. Lapizco-Encinas, *Anal. Bioanal. Chem.*, 2009, **394**, 293–302.
- 66 M. Nahavandi, *Ind. Eng. Chem. Res.*, 2016, **55**, 5484–5499.
- 67 J. Fathy, A. Pourmand and H. Badri Ghavifekr, *Iranian Conference on Electrical Engineering*, 2017.
- 68 T.-K. Chiu, W.-P. Chou, S.-B. Huang, H.-M. Wang, Y.-C. Lin, C.-H. Hsieh and M.-H. Wu, *Sci. Rep.*, 2016, **6**, 32851.
- 69 E. Samiei, H. R. Nejad and M. Hoorfar, *Appl. Phys. Lett.*, 2015, **106**, 204101.
- 70 S. K. Ravula, D. W. Branch, C. D. James, R. J. Townsend, M. Hill, G. Kaduchak, M. Ward and I. Brener, *Sens. Actuators, B*, 2008, **130**, 645–652.
- 71 C.-P. Jen, C.-H. Weng and C.-T. Huang, *Electrophoresis*, 2011, **32**, 2428–2435.
- 72 C.-H. Chuang and Y.-W. Huang, *Electrophoresis*, 2013, **34**, 3111–3118.
- 73 C.-H. Chuang, Y.-W. Huang and Y.-T. Wu, *Biomed. Microdevices*, 2012, **14**, 271–278.
- 74 H. Y. Huang, Y. H. Huang, W. L. Kao and D. J. Yao, *Biomicrofluidics*, 2015, **9**, 022404.
- 75 S. Golan, D. Elata and U. Dinnar, *Sens. Actuators, A*, 2008, **142**, 138–146.
- 76 M. Nasabi, K. Khoshmanesh, F. J. Tovar-Lopez, K. Kalantar-zadeh and A. Mitchell, *Electrophoresis*, 2013, **34**, 3150–3154.
- 77 S. Rajaraman, H. Noh, P. J. Hesketh and D. S. Gottfried, *Sens. Actuators, B*, 2006, **114**, 392–401.
- 78 A. A. Kayani, A. F. Chrimes, K. Khoshmanesh, V. Sivan, E. Zeller, K. Kalantar-zadeh and A. Mitchell, *Microfluid. Nanofluid.*, 2011, **11**, 93–104.
- 79 Y. Zhao, S. Hu and Q. Wang, *Microfluid. Nanofluid.*, 2014, **17**, 287–294.
- 80 H. A. Pohl and I. Hawk, *Science*, 1966, **152**, 647–649.
- 81 M. Mischel, F. Rougé, I. Lamprecht, C. Aubert and G. Prota, *Arch. Dermatol. Res.*, 1983, **275**, 141.
- 82 J. Vienken, U. Zimmermann, A. Alonso and D. Chapman, *Naturwissenschaften*, 1984, **71**, 158–160.
- 83 F. F. Becker, X. B. Wang, Y. Huang, R. Pethig, J. Vykoukal and P. R. C. Gascoyne, *Proc. Natl. Acad. Sci. U. S. A.*, 1995, **92**, 860–864.
- 84 K. Kaler and T. B. Jones, *Biophys. J.*, 1990, **57**, 173–182.



- 85 A. F. Chrimes, K. Khoshmanesh, P. R. Stoddart, A. A. Kayani, A. Mitchell, H. Daima, V. Bansal and K. Kalantar-zadeh, *Anal. Chem.*, 2012, **84**, 4029–4035.
- 86 A. Bezryadin, C. Dekker and G. Schmid, *Appl. Phys. Lett.*, 1997, **71**, 1273–1275.
- 87 S.-H. Jung, C. Chen, S.-H. Cha, B. Yeom, J. H. Bahng, S. Srivastava, J. Zhu, M. Yang, S. Liu and N. A. Kotov, *J. Am. Chem. Soc.*, 2011, **133**, 10688–10691.
- 88 S. Kuehn, P. Measor, E. J. Lunt, B. S. Phillips, D. W. Deamer, A. R. Hawkins and H. Schmidt, *Lab Chip*, 2009, **9**, 2212–2216.
- 89 R. Riahifar, E. Marzbanrad, B. Raissi, C. Zamani, M. Kazemzad and A. Aghaei, *Mater. Lett.*, 2011, **65**, 632–635.
- 90 L. Zhang and Y. Zhu, *Langmuir*, 2012, **28**, 13201–13207.
- 91 S. Park, M. Koklu and A. Beskok, *Anal. Chem.*, 2009, **81**, 2303–2310.
- 92 M. Antfolk, S. H. Kim, S. Koizumi, T. Fujii and T. Laurell, *Sci. Rep.*, 2017, **7**, 46507.
- 93 Y. Song, J. Yang, X. Shi, H. Jiang, Y. Wu, R. Peng, Q. Wang, N. Gong, X. Pan, Y. Sun and D. Li, *Sci. China: Chem.*, 2012, **55**, 524–530.
- 94 N. Demierre, T. Braschler, R. Muller and P. Renaud, *Sens. Actuators, B*, 2008, **132**, 388–396.
- 95 S. Gupta, R. G. Alargova, P. K. Kilpatrick and O. D. Velev, *Langmuir*, 2010, **26**, 3441–3452.
- 96 A. Ismail, M. P. Hughes, H. J. Mulhall, R. O. C. Oreffo and F. H. Labeed, *J. Tissue Eng. Regen. Med.*, 2015, **9**, 162–168.
- 97 S. Pui-ock, M. Ruchirawat and P. Gascoyne, *Anal. Chem.*, 2008, **80**, 7727–7734.
- 98 S. Kang, *European Journal of Mechanics, B: Fluids*, 2015, **54**, 53–68.
- 99 B. G. Hawkins, C. Huang, S. Arasanipalai and B. J. Kirby, *Anal. Chem.*, 2011, **83**, 3507–3515.
- 100 M. Safavieh, S. Khetani, F. Juillard, V. Kaul, M. K. Kanakasabapathy, K. M. Kaye and H. Shafiee, *Biosens. Bioelectron.*, 2016, **80**, 230–236.
- 101 M. D. Pysher and M. A. Hayes, *Anal. Chem.*, 2007, **79**, 4552–4557.
- 102 L. Yang, *Anal. Lett.*, 2012, **45**, 187–201.
- 103 C. C. Chung, I. F. Cheng, H. M. Chen, H. C. Kan, W. H. Yang and H. C. Chang, *Anal. Chem.*, 2012, **84**, 3347–3354.
- 104 M. Llamas, V. Giner and M. Sancho, *J. Phys. D: Appl. Phys.*, 1998, **31**, 3160–3167.
- 105 K. Khoshmanesh, S. Baratchi, F. J. Tovar-Lopez, S. Nahavandi, D. Wlodkowic, A. Mitchell and K. Kalantar-zadeh, *Microfluid. Nanofluid.*, 2012, **12**, 597–606.
- 106 J. Ramon-Azcon, T. Yasukawa and F. Mizutani, *Anal. Chem.*, 2011, **83**, 1053–1060.
- 107 F. Du, A. Hawari, M. Baune and J. Thoeming, *J. Membr. Sci.*, 2009, **336**, 71–78.
- 108 P. Dommersnes, Z. Rozynek, A. Mikkelsen, R. Castberg, K. Kjerstad, K. Hersvik and J. O. Fossum, *Nat. Commun.*, 2013, **4**, 2066.
- 109 G. R. Ballantyne and P. N. Holtham, *Miner. Eng.*, 2014, **55**, 75–79.
- 110 Q. Tao, M. Jiang and G. Li, *IEEE Trans. Nanotechnol.*, 2014, **13**, 517–526.
- 111 J. Wu, B. Yin, F. Wu, Y. Myung and P. Banerjee, *Appl. Phys. Lett.*, 2014, **105**, 183506.
- 112 A. Nerowski, M. Poetschke, M. Bobeth, J. Opitz and G. Cuniberti, *Langmuir*, 2012, **28**, 7498–7504.
- 113 A. W. Maijenburg, M. G. Maas, E. J. B. Rodijk, W. Ahmed, E. S. Kooij, E. T. Carlen, D. H. A. Blank and J. E. ten Elshof, *J. Colloid Interface Sci.*, 2011, **355**, 486–493.
- 114 S. Ahadian, U. Naito, V. J. Surya, S. Darvishi, M. Estili, X. Liang, K. Nakajima, H. Shiku, Y. Kawazoe and T. Matsue, *Carbon*, 2017, **123**, 460–470.
- 115 M. K. Gupta, N. Sinha, B. K. Singh and B. Kumar, *Mater. Lett.*, 2010, **64**, 1825–1828.
- 116 S. Kumar, Z. Peng, H. Shin, Z. L. Wang and P. J. Hesketh, *Anal. Chem.*, 2010, **82**, 2204–2212.
- 117 W. Liu, J. Zhu, Z. Wang and X. Tang, *Phys. E*, 2010, **42**, 1653–1658.
- 118 J. Regtmeier, R. Eichhorn, L. Bogunovic, A. Ros and D. Anselmetti, *Anal. Chem.*, 2010, **82**, 7141–7149.
- 119 T. Masuda, H. Maruyama, A. Honda and F. Arai, *PLoS One*, 2014, **9**(6), e94083.
- 120 A. Sonnenberg, J. Y. Marciniak, J. McCanna, R. Krishnan, L. Rassenti, T. J. Kipps and M. J. Heller, *Electrophoresis*, 2013, **34**, 1076–1084.
- 121 J. P. H. Burt, T. A. K. Al-Ameen and R. Pethig, *J. Phys. E: Sci. Instrum.*, 1989, **22**, 952–957.
- 122 G. Fuhr, T. Muller, V. Baukloh and K. Lucas, *Hum. Reprod.*, 1998, **13**, 136–141.
- 123 S. Archer, H. Morgan and F. J. Rixon, *Biophys. J.*, 1999, **76**, 2833–2842.
- 124 S. Kim, R. Asmatulu, H. L. Marcus and F. Papadimitrakopoulos, *J. Colloid Interface Sci.*, 2011, **354**, 448–454.
- 125 S. N. Ibrahim, L. Murray, V. Nock, J. J. Evans and M. M. Alkaisy, *Microelectron. Eng.*, 2012, **97**, 369–374.
- 126 M. Haapalainen and A. Mäkynen, in *Optical Trapping And Optical Micromanipulation VII*, ed. K. Dholakia and G. C. Spalding, 2010, p. 7762.
- 127 K. J. Freedman, L. M. Otto, A. P. Ivanov, A. Barik, S. H. Oh and J. B. Edell, *Nat. Commun.*, 2016, **7**, 10217.
- 128 P. V. Jones, G. L. Salmon and A. Ros, *Anal. Chem.*, 2017, **89**, 1531–1539.
- 129 H. Shafiee, J. L. Caldwell and R. V. Davalos, *JALA*, 2010, **15**, 224–232.
- 130 M. Mohammadi, M. J. Zare, H. Madadi, J. Sellares and J. Casals-Terre, *Anal. Bioanal. Chem.*, 2016, **408**, 5285–5294.
- 131 J. M. McMullan and N. J. Wagner, *Langmuir*, 2012, **28**, 4123–4130.
- 132 P. Garcia-Sanchez, Y. Ren, J. J. Arcenegui, H. Morgan and A. Ramos, *Langmuir*, 2012, **28**, 13861–13870.
- 133 S. Li, M. Li, K. Bougot-Robin, W. Cao, I. Y. Y. Chau, W. Li and W. Wen, *Biomicrofluidics*, 2013, **7**(2), 024106.
- 134 D. Ben-Bassat, A. Boymelgreen and G. Yossifon, *J. Colloid Interface Sci.*, 2015, **442**, 154–161.
- 135 V. H. Perez-Gonzalez, H. Vinh, M. Vazquez-Pinon, S. O. Martinez-Chapa and L. Kulinsky, *J. Micromech. Microeng.*, 2015, **25**(11), 115007.



- 136 A. Kale, X. Lu, S. Patel and X. Xuan, *J. Micromech. Microeng.*, 2014, **24**, 075007.
- 137 V. Nerguizian, A. Alazzam, J. Gauthier, D. Roman, I. Stiharu and M. Burnier, *Proc. Inst. Mech. Eng., Part B*, 2016, **230**, 1309–1316.
- 138 D. J. Allen, R. P. Accolla and S. J. Williams, *Electrophoresis*, 2017, **38**, 1441–1449.
- 139 Y. Luo and L. Yobas, *Appl. Phys. Lett.*, 2014, **105**(9), 094105.
- 140 C.-P. Jen, N. A. Maslov, H.-Y. Shih, Y.-C. Lee and F.-B. Hsiao, *Microsyst. Technol.*, 2012, **18**, 1879–1886.
- 141 M. T. Guler, I. Bilican, S. Agan and C. Elbuken, *J. Micromech. Microeng.*, 2015, **25**(9), 095019.
- 142 H.-W. Su, J. L. Prieto and J. Voldman, *Lab Chip*, 2013, **13**, 4109–4117.
- 143 J. Tirapu-Azpiroz, Y. Temiz and E. Delamarche, *Appl. Phys. Lett.*, 2015, **107**, 204102.
- 144 J. Voldman, R. A. Braff, M. Toner, M. L. Gray and M. A. Schmidt, *Micro Total Analysis Systems*, 2000, pp. 431–434.
- 145 M. A. Saucedo-Espinosa, A. LaLonde, A. Gencoglu, M. F. Romero-Creel, J. R. Dolas and B. H. Lapizco-Encinas, *Electrophoresis*, 2016, **37**, 282–290.
- 146 M. A. Saucedo-Espinosa and B. H. Lapizco-Encinas, *Electrophoresis*, 2015, **36**, 1086–1097.
- 147 A. Kale, S. Patel, S. Qian, G. Hu and X. Xuan, *Electrophoresis*, 2014, **35**, 721–727.
- 148 R. C. Gallo-Villanueva, M. B. Sano, B. H. Lapizco-Encinas and R. V. Davalos, *Electrophoresis*, 2014, **35**, 352–361.
- 149 S. Tanaka, M. Tsutsui, H. Theodore, Y. He, A. Arima, T. Tsuji, K. Doi, S. Kawano, M. Taniguchi and T. Kawai, *Sci. Rep.*, 2016, **6**, 31670.
- 150 S. A. M. Kirmani, F. D. Gudagunti, L. Velmanickam, D. Nawarathna and I. T. Lima Jr, *J. Biomed. Opt.*, 2017, **22**, 037006.
- 151 P. Zhou, H. Yu, W. Yang, Y. Wen, Z. Wang, W. J. Li and L. Liu, *ACS Appl. Mater. Interfaces*, 2017, **9**, 16715–16724.
- 152 Y. Tao and H. K. Wickramasinghe, *Appl. Phys. Lett.*, 2017, **110**, 073701.
- 153 S. Stanke, F. F. Bier and R. Hoesel, *Electrophoresis*, 2011, **32**, 2448–2455.
- 154 Y. Yan, J. Luo, D. Guo and S. Wen, *PLoS One*, 2015, **10**, 0117456.
- 155 L. Zhang and Y. Zhu, *Appl. Phys. Lett.*, 2010, **96**(14), 141902.
- 156 M. Javanmard, S. Emaminejad, R. W. Dutton and R. W. Davis, *Anal. Chem.*, 2012, **84**, 1432–1438.
- 157 I. S. Park, K. Eom, J. Son, W.-J. Chang, K. Park, T. Kwon, D. S. Yoon, R. Bashir and S. W. Lee, *ACS Nano*, 2012, **6**, 8665–8673.
- 158 A. Yamaguchi, T. Fukuoka, K. Kuroda, R. Hara and Y. Utsumi, *Colloids Surf., A*, 2016, **507**, 118–123.
- 159 M. H. Kim, J. Lee, K. Nam, I. S. Park, M. Son, H. Ko, S. Lee, D. S. Yoon, W.-J. Chang, S. Y. Lee, Y. R. Yoon and S. W. Lee, *Sensors*, 2017, **17**(10), 2272.
- 160 C. Hanson and E. Vargis, *Sensors*, 2017, **17**, 327.
- 161 A. Yamaguchi, T. Fukuoka, R. Takahashi, R. Hara and Y. Utsumi, *Sens. Actuators, B*, 2016, **230**, 94–100.
- 162 T. T. Kanagasabapathi, C. Dalton, and K. V. I. S. Kaler, *Asme 3rd International Conference on Microchannels and Minichannels, Pt B*, 2005.
- 163 W. Zhang and A. D. Radadia, *Anal. Chem.*, 2016, **88**, 2605–2613.
- 164 M. F. Romero-Creel, E. Goodrich, D. V. Polniak and B. H. Lapizco-Encinas, *Micromachines*, 2017, **8**, 239.
- 165 N. N. Yahya, N. Abd Aziz, M. R. Buyong and B. Y. Majlis, *IEEE. RSM*, 2017.
- 166 J. Sun, Y. Gao, R. J. Isaacs, K. C. Boelte, P. C. Lin, E. M. Boczek and D. Li, *Anal. Chem.*, 2012, **84**, 2017–2024.
- 167 M. Viefhues, R. Eichhorn, E. Fredrich, J. Regtmeier and D. Anselmetti, *Lab Chip*, 2012, **12**, 485–494.
- 168 Y. Tao, Y. Ren, H. Yan and H. Jiang, *Chin. J. Mech. Eng.*, 2016, **29**, 325–331.
- 169 X. Lu, J.-P. Hsu and X. Xuan, *Langmuir*, 2015, **31**, 620–627.
- 170 Y. Wang, F. Du, G. R. Pesch, J. Koeser, M. Baune and J. Thoeming, *Chem. Eng. Sci.*, 2016, **153**, 34–44.
- 171 K. Zhao and D. Li, *J. Micromech. Microeng.*, 2017, **27**, 274–284.
- 172 J. Zhu, R. C. Canter, G. Ketten, P. Vedantam, T.-R. J. Tzeng and X. Xuan, *Microfluid. Nanofluid.*, 2011, **11**, 743–752.
- 173 K. Zhao and D. Li, *Sens. Actuators, B*, 2017, **250**, 274–284.
- 174 H. Zhu, X. Lin, Y. Su, H. Dong and J. Wu, *Biosens. Bioelectron.*, 2015, **63**, 371–378.
- 175 S. Afshar, E. Salimi, K. Braasch, M. Butler, D. J. Thomson and G. E. Bridges, *IRE Trans. Microwave Theory Tech.*, 2016, **64**, 991–998.
- 176 S.-Y. Tang, W. Zhang, R. Soffe, S. Nahavandi, R. Shukla and K. Khoshmanesh, *PLoS One*, 2014, **9**, e104109.
- 177 A. LaLonde, M. F. Romero-Creel, M. A. Saucedo-Espinosa and B. H. Lapizco-Encinas, *Biomicrofluidics*, 2015, **9**(6), 064113.
- 178 N. Lewpiriyawong, K. Kandaswamy, C. Yang, V. Ivanov and R. Stocker, *Anal. Chem.*, 2011, **83**, 9579–9585.
- 179 Y. Jia, Y. Ren and H. Jiang, *Electrophoresis*, 2015, **36**, 1744–1753.
- 180 K.-S. Chow and H. Du, *Sens. Actuators, A*, 2011, **170**, 24–31.
- 181 K. Tian, K. Decker, A. Aksimentiev and L.-Q. Gu, *ACS Nano*, 2017, **11**, 1204–1213.
- 182 Y. J. Yuan and R. Jia, *Nanomedicine*, 2012, **7**, 381.
- 183 Y. J. Yuan, A. T. Pyatenko and M. Suzuki, *Nanoscale Res. Lett.*, 2007, **2**, 119–122.
- 184 Z. Xu and Y. J. Yuan, *Biosens. Bioelectron.*, 2018, **99**, 500–512.
- 185 M. Seidel and R. Niessner, *Anal. Bioanal. Chem.*, 2014, **406**, 5589–5612.

