



Cite this: *RSC Adv.*, 2019, 9, 9500

# Weighted spectral reconstruction method for discrimination of bacterial species with low signal-to-noise ratio Raman measurements

Shanshan Zhu,<sup>a</sup> Xiaoyu Cui,<sup>ad</sup> Wenbin Xu,<sup>b</sup> Shuo Chen<sup>\*ad</sup> and Wei Qian<sup>c</sup>

Raman spectroscopy is a label-free and non-destructive spectroscopic technique that has been explored for bacterial identification. However, noise often interferes with the interesting Raman peaks because the Raman signal is inherently weak, especially for bacterial samples. Although this problem can be solved by increasing the exposure time or the power of the excitation laser, a longer acquisition time is required or the risk of sample damage is increased. In contrast, short exposure time and low laser power often lead to inadequate acquisition of Raman scattering, in which the Raman spectra with low signal-to-noise ratio (SNR) is difficult to be further analyzed. In order to quickly and accurately characterize biological samples by using low SNR Raman measurements, a weighted spectral reconstruction based method was developed and tested on Raman spectra with low SNR from 20 bacterial samples of two species. Principal component analysis followed by support vector machine was applied on the reference Raman spectra and the spectra recovered from the low SNR Raman measurements by the proposed method, the traditional spectral reconstruction method, and four other commonly used de-noising methods for the discrimination of bacterial species. The results showed that a classification accuracy of 90% was achieved based on our method, which was comparable to that of the reference Raman spectra and showed significant advantages over other spectral recovery methods. Therefore, the weighted spectral reconstruction method can preserve the most biochemical information for the bacterial species' identification while removing the noise from the low SNR Raman spectra, in which the advantages of lesser sample damage and shorter acquisition time would promote wider biomedical applications of Raman spectroscopy.

Received 14th January 2019

Accepted 8th March 2019

DOI: 10.1039/c9ra00327d

[rsc.li/rsc-advances](http://rsc.li/rsc-advances)

## Introduction

The quick identification and discrimination of bacterial species have been a hot point of concern especially for clinicians, which is critical for the treatment of infections.<sup>1,2</sup> These concerns are highly relevant in the area of medical applications and aim at fast and precise diagnosis and treatment. Different approaches for bacterial identification include morphological and biochemical tests, such as Gram staining, enzyme activity tests, antibiotic susceptibility profiles as well as 16S ribosomal deoxyribonucleic acid or 16S ribosomal ribonucleic acid analysis.<sup>3,4</sup> The conventional examinations can be applied only on pure, isolated bacteria derived from culture plates.<sup>5</sup> However, bacterial culture and purification are often time-consuming and complicated, and usually require several days for the clinicians

to get the final laboratory report after the sample collection from patients. Although genotypic methods are faster and more accurate than the traditional biochemical tests, DNA extraction from pure cultures and consumption of expensive reagents are still needed.<sup>4</sup> Due to the low efficiency of the current bacterial identification methods, empirical therapy usually has to be performed while awaiting laboratory results, which leads to a low treatment effect on the infection and loss of the optimal time for appropriate treatment, even resulting in deterioration of the infection or mortality. Tumbarello *et al.* indicated that 20.1% patients with bloodstream infections caused by *Escherichia coli* do not receive proper antimicrobial therapy and have a higher proportion of extended-spectrum  $\beta$ -lactamase bloodstream infections (74.0% versus 15.8%) and higher 21 day mortality rates (40.7% versus 5.6%) compared with patients who receive the appropriate therapy from the beginning.<sup>6</sup> Werarak *et al.* demonstrated that although carbapenem is the most frequently used treatment for hospital-acquired pneumonia and ventilator-associated pneumonia, the patients who are treated by carbapenem in the beginning of infection have more severe complications and high mortality rate.<sup>7</sup> Hence, a rapid

<sup>a</sup>Sino-Dutch Biomedical and Information Engineering School, Northeastern University, Shenyang, China, 110169. E-mail: [chenshuo@bmie.neu.edu.cn](mailto:chenshuo@bmie.neu.edu.cn); Tel: +86 24 83680230

<sup>b</sup>Science and Technology on Optical Radiation Laboratory, Beijing, China, 110854

<sup>c</sup>College of Engineering, University of Texas at El Paso, El Paso, USA, 79968

<sup>d</sup>Key Laboratory of Data Analytics and Optimization for Smart Industry (Northeastern University), Ministry of Education, China



and accurate identification of the bacterial pathogens is critical for the precise treatment of patients.

In recent years, Raman spectroscopy has been explored in the detection and identification of bacteria as a fast, non-destructive, and label-free spectroscopic technique based on measuring the inelastic scattering of photons by vibrating molecules or crystal lattices.<sup>8,9</sup> Based on the specific vibrational modes of the molecules, a large amount of qualitative and quantitative biochemical information enables the identification and differentiation of bacterial biochemical components.<sup>10,11</sup> Unfortunately, noise often interferes with the interesting Raman peaks because the Raman signal is inherently weak, especially for bacterial samples.<sup>12,13</sup> Although this problem can be solved by increasing the exposure time or the power of the excitation laser, a longer acquisition time is required or the risk of sample damage is increased. If a short exposure time and low laser power is applied during the Raman measurement, inadequate acquisition of Raman scattering would lead to low signal-to-noise ratio (SNR) measurements, which makes the further spectral data processing and analysis difficult. Therefore, a method to quickly and accurately discriminate the bacterial species by using low SNR Raman measurements with low laser power and short exposure time can potentially solve the above problems.

In this study, a weighted spectral reconstruction based method was newly developed and tested on Raman spectra with low SNR equal to 0.98 from 20 bacterial samples of two species, *i.e.*, *Pseudomonas aeruginosa* and *Staphylococcus aureus*. For identifying the different bacterial species, principal component analysis (PCA) followed by support vector machine (SVM) was applied on the reference Raman spectra (high SNR), low SNR Raman spectra, and low SNR Raman spectra processed by the proposed method, traditional spectral reconstruction method, and four other commonly used de-noising methods, *i.e.*, Savitzky–Golay (SG) algorithm, wavelet transform, finite impulse response (FIR) filtration, and factor analysis. Compared with other methods, the proposed method demonstrated significant improvement in the spectral recovery and higher accuracy in the discrimination of bacterial species, in which the classification accuracy of the Raman spectra recovered by the proposed method was even comparable to that of the reference Raman spectra with high SNR. Therefore, our method demonstrated significant potential in the rapid and accurate bacterial species' identification based on low SNR Raman spectra, wherein the sample damage was lesser and a shorter acquisition time was required.

## Experimental

### Sample preparation

The bacterial samples of the two species, *i.e.*, *Staphylococcus aureus* (ATCC 29213) and *Pseudomonas aeruginosa* (ATCC 9027), were cultured overnight at a fixed temperature of 35 °C in Tryptic Soy Agar Plates. Thereafter, few bacterial colonies were selected and placed in the phosphate buffered saline to reach a final concentration of  $1 \times 10^8$  CFU mL<sup>-1</sup>, followed by centrifugation at 10 000 rpm for five minutes to concentrate and

purify the bacterial samples. The bacterial samples were rinsed and immersed twice in distilled water to wash away the culture medium, and finally resuspended in 100  $\mu$ L distilled water. To prepare the samples for Raman measurements, 2  $\mu$ L of the suspension was repeatedly dropped and air dried at the same location on an aluminum foil, in which the bacterial samples were concentrated and a relatively higher Raman signal could be achieved.<sup>14</sup>

### Raman spectra acquisition

Twenty pairs of Raman spectra, *i.e.*, 10 pairs of Raman measurements from *Pseudomonas aeruginosa* and 10 pairs of Raman measurements from *Staphylococcus aureus*, were obtained in the wavenumber range from 600 cm<sup>-1</sup> to 1600 cm<sup>-1</sup> by a confocal Raman microscope (inVia, Renishaw, UK). A 633 nm diode laser was used for excitation and the spectral resolution was 2 cm<sup>-1</sup>. In each pair of Raman measurements, both the low SNR and high SNR Raman spectra were collected from the exact same position. The exposure time for low SNR Raman measurement was 1 s and accumulated once, whereas that for the corresponding high SNR Raman measurement was 10 s and accumulated 30 times, so that the actual acquisition time for the high SNR Raman measurements was 300 times of that for the low SNR Raman measurements. In this study, the high SNR Raman spectra were treated as the reference to evaluate the proposed method. The noise level was quantified by a published method,<sup>15</sup> *i.e.*, by dividing the maximum intensity of the normalized reference Raman spectrum by the root mean square error (RMSE) between the normalized low SNR Raman spectrum and the normalized reference Raman spectrum, in which the normalization for each Raman spectrum was performed by dividing the Raman intensity at each wavenumber by the summation of the Raman intensities at all wavenumbers. Thus, the SNR of the low SNR Raman spectra used in this study was 0.98.

### Spectral recovery methods

Twenty Raman spectra with low SNR were de-noised by the traditional spectral reconstruction method, weighted spectral reconstruction method, and four other commonly used de-noising methods, *i.e.*, SG algorithm, wavelet transform, FIR filtration, and factor analysis. The corresponding high SNR Raman spectra were treated as the reference Raman spectra to evaluate the performance of each method.

Traditional spectral reconstruction was performed to retrieve the Raman spectra from the low SNR Raman measurements,<sup>15–17</sup> in which a calibration data set was required, as shown in Fig. 1. In the calibration data set, both the reference Raman spectra and the narrow-band measurements derived from the low SNR Raman spectra were included, whereas the test data set contained only the low SNR Raman spectra. The narrow-band measurements were numerically calculated by the inner production of low SNR Raman spectra and the transmittance of the non-negative PC based filters,<sup>18</sup> in which the first six non-negative principal component (PC) based filters were used. In the calibration stage, the Wiener matrix  $W$  was






Fig. 1 The flowchart of the traditional spectral reconstruction method.

calculated based on the calibration dataset to extract the relation between the narrow-band measurements  $C_{\text{cal}}$  and the reference Raman spectra  $R_{\text{ref}}$ , as shown in eqn (1).

$$W = E(R_{\text{ref}}C_{\text{cal}}^T)[E(C_{\text{cal}}C_{\text{cal}}^T)]^{-1} \quad (1)$$

where  $E(\ )$  represents the ensemble average,  $T$  represents the matrix transpose, and  $-1$  represents the matrix inverse. In the test stage, the reconstructed high SNR Raman spectra can be retrieved by the multiplication of the Wiener matrix and the corresponding narrow-band measurements in the test data set.

The weighted spectral reconstruction was a newly developed method and different from the traditional spectral reconstruction method, in which the ensemble average was replaced by the weighted average when constructing the weighted Wiener matrix  $\hat{W}$ , as shown in eqn (2).

$$\hat{W} = \sum_{i=1}^n \rho_i R_{\text{ref}(i)} C_{\text{cal}(i)}^T \sum_{i=1}^n (\rho_i C_{\text{cal}(i)} C_{\text{cal}(i)}^T)^{-1} \quad (2)$$

where  $\rho_i$  refers to the weight for the  $i$ -th set of the calibration data. The weight can be calculated based on the similarity of the narrow-band measurements according to eqn (3), in which the calibration data with higher similarity to the test data would contribute more to the weighted Wiener matrix.

$$\rho_i = \frac{S_i}{\sum_{i=1}^n S_i} \quad (3)$$

where  $S_i$  is the similarity between the narrow-band measurements of the low SNR Raman spectra from the test data set and the  $i$ -th set of the narrow-band measurements in the calibration data set, and can be calculated according to eqn (4).

$$S_i = \frac{D_i^m}{\sum_{i=1}^n D_i^m} \quad (4)$$

where  $D_i$  is the difference between the narrow-band measurements from the test data and the  $i$ -th set of the narrow-band measurements in the calibration data, and  $m$  is the power to adjust the contribution of  $D_i$ . The reason for using the difference of the narrow-band measurement instead of the difference of the low SNR Raman spectra is that the difference of the

narrow-band measurement can more precisely represent the similarity between the test data and calibration data, in which the narrow-band measurement is the integration of the Raman intensities along the wavelength dimension and its SNR should be much higher compared to its corresponding low SNR Raman spectrum. In contrast to the traditional spectral reconstruction method, the weighted reconstructed high SNR Raman spectra were retrieved by the multiplication of the weighted Wiener matrix and the narrow-band measurements in the test stage, as shown in Fig. 2. Since both the weighted spectral reconstruction and the traditional spectral reconstruction methods were supervised learning methods in nature, the leave-one-out cross validation method was used by selecting one sample for testing and the rest for training until all the samples were tested.<sup>19</sup>

Besides the traditional spectral reconstruction and weighted spectral reconstruction methods, SG algorithm, wavelet transform, FIR filtration, and factor analysis were applied on the same set of low SNR Raman spectra for comparison. For the SG algorithm, each part of the original spectrum with a selected window size was fitted to a polynomial function for smoothing purpose.<sup>20</sup> In contrast, the wavelet transform, FIR filtration, and factor analysis commonly remove noise by filtering techniques. For the wavelet transform,<sup>21,22</sup> the spectral data were decomposed into the wavelet domain by various wavelet basis and reconstructed after noise removal by certain thresholds. The FIR filtration is a linear filtration technique, in which a window-based FIR filter is designed based on the frame size and cut-off frequency and was subsequently used for noise removal in this study.<sup>23</sup> For factor analysis,<sup>24</sup> the original spectral information is projected into the linear combination of a certain number of subspectra, and those subspectra related to the noise can be subsequently removed. The parameters of each de-noising method and the range of the parameters are shown in Table 1.

After the low SNR Raman spectra were de-noised, the broad and slowly varying fluorescence background was estimated by using the fifth order polynomial fitting and subtracted from the original spectra.<sup>25</sup> Normalization was subsequently performed on each Raman spectrum by dividing the Raman intensity at each wavenumber by the summation of the Raman intensities at all the wavenumbers. In order to evaluate the accuracy of the recovered Raman spectra, the above fluorescence background






Fig. 2 The flowchart of the weighted spectral reconstruction method.

removal algorithm and normalization were applied on the corresponding reference Raman spectra as well and the mean relative RMSE<sup>26</sup> was used as the metric.

### PCA-SVM based classification methods

After obtaining the recovered Raman spectra from the low SNR Raman spectra by each of the above methods, principal component analysis (PCA) followed by support vector machine (SVM) was applied on the reference Raman spectra, low SNR Raman spectra, and the recovered Raman spectra for bacterial species' discrimination. As a multivariate statistical method, PCA can reduce the large original spectral data into small number of independent variables named as PC scores, which can effectively carry the most important information of the

corresponding spectra. Thus, PCA can be used to group the spectra by choosing different combinations of the PC scores to build a new coordinate system.<sup>27</sup> In order to achieve a fair comparison of the different spectral recovery methods in this study, the first four PC scores were chosen for bacterial species' discrimination because the highest classification accuracy was achieved based on the reference Raman spectra with high SNR when using the first four PC scores. Support vector machine is a state-of-the-art supervised machine learning method especially suited to classify and identify different bacterial species.<sup>28,29</sup> SVM belongs to the group of maximum margin classifiers and can efficiently find the optimal solution for the given parameters. In this paper, a classification model was built by inputting the first four PC scores for SVM to identify the bacterial species. The leave-one-out method was used for cross validation and the performance of the different spectral recovery methods was compared in terms of accuracy, sensitivity, and specificity in identifying the different bacterial species. In addition, the integrated area under the receiver operating characteristic (ROC) curve was used to quantify the performance of the classification model on different groups of Raman spectra.

## Results and discussion

Fig. 3 shows the comparison among the average spectra after the fluorescence background removal and normalization based on the different spectral recovery methods between *Pseudomonas aeruginosa* and *Staphylococcus aureus* samples. It was found that both the traditional spectral reconstruction method (see Fig. 3(c)) and weighted spectral reconstruction method (see Fig. 3(d)) showed excellent agreement in the Raman peaks and the spectral shape compared to the reference Raman spectra (see Fig. 3(a)). The excellent performance of the spectral reconstruction based methods can be attributed to two factors. One is that the Raman measurements are integrated along the wavenumber dimension in the procedure of synthesizing the

Table 1 Parameters of each de-noising method and the range of the parameters

Method	Parameter	Parameter range
Traditional spectral reconstruction	Number of non-negative PC scores based filters	6
	Weighted spectral reconstruction	Number of non-negative PC scores based filters
SG algorithm	Power	−0.1 to −10
	Window size	3 to 729
Wavelet transform	Polynomial degree	1 to 9
	Wavelet basis	Common wavelet filters built in Matlab
	Decomposition level	1 to 10
FIR filtration	Threshold	Soft threshold or hard threshold: threshold value were selected according to the Birge–Massart strategy
	Frame size	2 to 243
Factor analysis	Cut-off frequency	$1 \times 10^{-10}$ to 1
	Number of subspectra	1 to 20








Fig. 3 Comparison of (a) the mean reference Raman spectrum, (b) the mean low SNR Raman spectrum, and the mean recovered Raman spectrum based on (c) the traditional spectral reconstruction method, (d) the weighted spectral reconstruction method, (e) SG algorithm, (f) wavelet transform, (g) FIR filtration, and (h) factor analysis for *Pseudomonas aeruginosa* and *Staphylococcus aureus*.

narrow-band measurements from the low SNR Raman spectra, thus, the SNR can be significantly improved by eliminating the shot noise in Raman measurements.<sup>30,31</sup> The other factor is that the spectral reconstruction based methods contain prior information about the samples, in which the high SNR Raman spectra are used to associate with the narrow-band measurements in the calibration stage. Furthermore, the weighted spectral reconstruction method shows the best performance in spectral recovery among all the methods, and improves 4.3% in the mean relative RMSE compared to the traditional spectral reconstruction method, as shown in Table 2. The reason is that the weighted spectral reconstruction method takes the advantage of collecting more effective information from the samples with higher similarity when constructing the weighted Wiener matrix. However, the time efficiency of the weighted spectral reconstruction is lower because different weighted Wiener matrices have to be created for each test sample, whereas the traditional spectral reconstruction method only requires a single Wiener matrix when the calibration dataset is fixed. Although the SG algorithm is a commonly used smoothing method for de-noising Raman spectra, the spectral shape was distorted and some of the peak locations were shifted, as shown in Fig. 3(e), because those weak features in the Raman spectra are comparable to the noise level, and thus, can be easily smoothed out during noise removal.<sup>20</sup> The wavelet transform (see Fig. 3(f)) and FIR filtration (see Fig. 3(g)) methods showed good performance in noise removal, at the cost of losing some

important spectral shape information, e.g., the central wavelengths and bandwidths of the peaks.<sup>21</sup> Unfortunately, the factor analysis did not work well when the SNR was extremely low, which could not even smoothen out the noise and lost plenty of useful information during bacterial discrimination, as shown in Fig. 3(h). The reason might be that the factor analysis loses the ability to decompose the noise and signal when the SNR is extremely low because the noise contributes similarly or even more to the Raman spectra compared with the signal.<sup>24</sup>

Table 3 shows the comparison of the classification accuracy, sensitivity, and specificity of *Pseudomonas aeruginosa* and *Staphylococcus aureus* from the Raman spectra after fluorescence background removal and normalization of the reference Raman spectra, low SNR Raman spectra, and Raman spectra recovered by using the traditional spectral reconstruction method, weighted spectral reconstruction method, SG algorithm, wavelet transform, FIR filtration, and factor analysis. For identifying *Pseudomonas aeruginosa* and *Staphylococcus aureus*, the Raman spectra recovered by both traditional and weighted spectral reconstruction methods can achieve a classification accuracy of 90%, which was exactly the same as that of the reference Raman spectra and showed significant advantages over the other commonly used de-noising methods as well as the results of the low SNR Raman spectra. Furthermore, the weighted spectral reconstruction method also successfully demonstrated exactly the same sensitivity and specificity compared to the reference Raman spectra, whereas the

Table 2 Comparison of the mean relative RMSE between the reference Raman spectra with low SNR Raman spectra and the Raman spectra recovered from low SNR Raman measurements using the traditional spectral reconstruction, weighted spectral reconstruction, SG algorithm, wavelet transform, FIR filtration, and factor analysis

	Low SNR Raman spectra	Traditional spectral reconstruction	Weighted spectral reconstruction	SG algorithm	Wavelet transform	FIR filtration	Factor analysis
Mean relative RMSE	$1.98 \times 10^{-1}$	$8.21 \times 10^{-2}$	$7.86 \times 10^{-2}$	$1.47 \times 10^{-1}$	$1.45 \times 10^{-1}$	$1.54 \times 10^{-1}$	$1.48 \times 10^{-1}$



**Table 3** Comparison of the classification accuracy, sensitivity, and specificity of bacterial species identification based on the reference Raman spectra, low SNR Raman spectra, and Raman spectra recovered from the low SNR Raman measurements using the traditional spectral reconstruction, weighted spectral reconstruction, SG algorithm, wavelet transform, FIR filtration, and factor analysis

	Reference Raman spectra	Low SNR Raman spectra	Traditional spectral reconstruction	Weighted spectral reconstruction	SG algorithm	Wavelet transform	FIR filtration	Factor analysis
Classification accuracy	90%	75%	90%	90%	70%	85%	70%	35%
Sensitivity	90%	80%	80%	90%	70%	90%	50%	40%
Specificity	90%	70%	100%	90%	70%	80%	90%	30%

traditional spectral reconstruction failed. Although the specificity of the traditional spectral reconstruction is the highest, it sacrifices the sensitivity, as shown in Table 3, and we believe some improper prior information is used during the spectral recovery process of the traditional spectral reconstruction. Thus, the higher spectral recovery accuracy of the weighted spectral reconstruction method is indeed critical for better performance in the following spectral data analysis. In practical applications, the choice of the traditional spectral reconstruction and weighted spectral reconstruction should be mainly dependent on its specific applications, in which the compromise between time efficiency and spectral recovery accuracy should be considered. Interestingly, the classification accuracy did not fully comply with the mean relative RMSE, *i.e.*, the agreement between the recovered Raman spectra and the reference Raman spectra. The reason might be that most of the information is preserved whereas some critical information for bacterial identification is lost during the noise removal, especially for the SG algorithm, FIR filtration, and factor analysis methods. The classification results of these three methods are even lower than that of the low SNR Raman spectra, indicating that more information is lost compared to the information gained during the spectral recovery by these three methods. This can be attributed to the fact that the importance of the information cannot be distinguished by these commonly used de-noising methods. For the SG algorithm, the weak features in the Raman spectra comparable to the noise level can be easily smoothed out during noise removal, resulting in some shifted Raman peaks and the distorted spectral shape.<sup>20</sup> By the FIR filtration method, some important spectral shape information is lost simultaneously, while the noise is well removed. Although the Raman spectra de-noised by the SG algorithm and FIR filtration methods retain some important information about the peak locations and the spectral shape, the information regarding the discrimination of the two bacterial samples might be removed during noise removal, resulting in relatively low classification accuracy of only 70%. The classification accuracy of the factor analysis method was the lowest among all the methods, whereas the mean relative RMSE was not the worst. The reason is that factor analysis loses the ability to decompose the noise and signal when the SNR is extremely low, thus, plenty of important Raman peaks for bacterial discrimination as well as the noise are smoothed out simultaneously.<sup>24</sup>

To further evaluate the performance of the PCA-SVM-based classification model for bacterial species, the ROC curves were generated at different threshold levels for different groups of Raman spectra, respectively. The integrated area under the ROC curve (AUC) is a quantitative indicator used to represent the classifier performance, in which the larger AUC value usually means that the classifier has higher prediction accuracy.<sup>32</sup> According to the results in Fig. 4, the integrated areas under the ROC curves (AUC) are 0.96, 0.79, 0.98, 0.99, 0.66, 0.86, 0.78, and 0.32 for the reference Raman spectra, low SNR Raman spectra, and Raman spectra recovered from the low SNR Raman measurements using the traditional spectral reconstruction, weighted spectral reconstruction, SG algorithm, wavelet transform, FIR filtration, and factor analysis, respectively. Thus, the Raman spectra recovered from the low SNR Raman measurements using the weighted spectral reconstruction method demonstrates the strongest ability of bacterial species' identification with high sensitivity and specificity, and even outperforms the reference Raman spectra with high SNR. The reason might be that the spectral reconstruction procedure can remove some of the useless information within the reference Raman spectra, which may provide negative impacts on the bacterial species' identification.

Fig. 5 shows the average Raman spectra based on the reference Raman spectra and Raman spectra recovered by the spectral reconstruction method of *Pseudomonas aeruginosa* and *Staphylococcus aureus*, respectively. By the visual inspection of the reference Raman spectra (red curve) and the Raman spectra after spectral reconstruction (blue curve) in Fig. 5(a) and (b), it can be noted that the major Raman features were at 853  $\text{cm}^{-1}$  (tyrosine ring breathing vibration of protein), 1003  $\text{cm}^{-1}$  (phenylalanine ring vibration of protein), 1126  $\text{cm}^{-1}$  (C–N, C–C stretching of protein and C–C lipid stretch), 1447  $\text{cm}^{-1}$  ( $\text{CH}_2$ ,  $\text{CH}_3$  lipid, and protein), and 1556  $\text{cm}^{-1}$  (C=C vibration of protein).<sup>33,34</sup> Moreover, the Raman peaks at 725  $\text{cm}^{-1}$  and 751  $\text{cm}^{-1}$  can be identified and assigned to the adenine and thymine ring breathing vibrations of DNA. From Fig. 5(c), it can be seen that the intensity of these bands were consistently lower in *Pseudomonas aeruginosa* compared to *Staphylococcus aureus*, indicating a significant reduction in both the DNA and RNA concentrations in *Pseudomonas aeruginosa*. In addition, the Raman peaks at 1298  $\text{cm}^{-1}$  and 1447  $\text{cm}^{-1}$ , which are assigned to the  $\text{CH}_2$  and  $\text{CH}_3$  bending modes were found primarily in proteins and lipids. The intensity of these two peaks notably






Fig. 4 The ROC curves obtained by using the PCA-SVM-based spectral classification based on (a) the reference Raman spectra, (b) the low SNR Raman spectra, and Raman spectra recovered from low SNR Raman measurements using (c) the traditional spectral reconstruction, (d) the weighted spectral reconstruction, (e) the SG algorithm, (f) the wavelet transform, (g) the FIR filtration, and (h) the factor analysis.

decreased in *Pseudomonas aeruginosa* compared to those in *Staphylococcus aureus*, as shown in Fig. 5(c), which may indicate the decrease of the membranous lipids in *Pseudomonas aeruginosa*.<sup>14</sup> The proteins have a prominent peak at  $1556\text{ cm}^{-1}$ , which was assigned to the C=C vibration.<sup>35</sup> As in Fig. 5(c), the intensity at this wavenumber increased in *Pseudomonas aeruginosa* compared to that in *Staphylococcus aureus*, which indicated the difference in the protein level between these two bacterial samples.<sup>35</sup> Although the recovered Raman spectra retained the information about most of the Raman peaks, some Raman information were still lost or overlapped, especially for Raman

peaks at around  $751\text{ cm}^{-1}$  (thymine ring breathing vibration of DNA) and  $1251\text{ cm}^{-1}$  (amide III of protein and adenine ring breathing vibration of DNA), as shown in Fig. 5(c). It was found that the difference in the spectra of the two bacterial species at these two peaks after spectral reconstruction were much closer to zero than that of the reference spectra, which demonstrated that some Raman information of DNA and proteins was lost or overlapped by the surrounding peaks. Even though the information in these bands and modes of bacterial components were lost or overlapped, it has very minor impact on the final classification results between *Pseudomonas aeruginosa* and




Fig. 5 Average Raman spectra based on the reference Raman spectra (red curve) and the recovered Raman spectra (blue curve) by using the weighted spectral reconstruction method: (a) *Pseudomonas aeruginosa*, (b) *Staphylococcus aureus*, and (c) the difference between them.



*Staphylococcus aureus* (see Table 3). In other words, these bands and modes may not be the key biomarkers responsible for the discrimination in these two bacterial species.

Although only twenty pairs of Raman spectra of *Pseudomonas aeruginosa* and *Staphylococcus aureus* were tested in this study, many studies based on a large amount of bacterial samples have verified the feasibility of using Raman spectroscopy as a powerful tool for bacterial species' identification with acceptable Raman signals.<sup>34,36,37</sup> In this study, we mainly focused on the information preservation by the proposed weighted spectral reconstruction method during the noise removal from the low SNR Raman spectra. The recovered Raman spectra by the proposed weighted spectral reconstruction method show closest relative RMSE, accuracy, sensitivity, specificity, and AUC value compared to those of the reference Raman spectra, which demonstrates the proposed method's excellent preservation of the most useful spectral information such as Raman peaks and the spectral shape for bacterial species' identification. Thus, we believe that the proposed method can still work well or even better for bacterial species' identification when the amount of bacterial samples increases. In the future study, a large size of sample data set and even other bacterial species will be investigated to validate and confirm these conclusions.

## Conclusions

In this study, a weighted spectral reconstruction based method was developed and tested on low SNR Raman spectra from bacterial samples for the discrimination of two bacterial species. PCA combined with SVM classifier was used to identify different bacterial species. Based on the results, it was found that the proposed method could recover the Raman spectra in excellent agreement with the reference Raman spectra and reach a comparable classification accuracy of 90% for bacterial species' identification, which shows significant advantages over other spectral recovery methods. Therefore, the weighted spectral reconstruction based method can excellently recover the Raman signal from the low SNR Raman spectra and preserve the most important information, in which the lesser sample damage and shorter acquisition time would promote the wider application of Raman spectroscopy in biomedical applications.

## Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

The authors would like to acknowledge the financial support from the National Natural Science Foundation of China (61605025, 61501101), Science and Technology Foundation of National Defense Key Laboratory (61424080209), Program for Innovation Talents in Universities of Liaoning Province (LR2016031), Programs Supported by Ningbo Natural Science Foundation (2018A610365), the Fundamental Research Funds

for the Central Universities (N171902001, N171904006), and the 111 Project (B16009).

## References

- 1 A. M. Caliendo, D. N. Gilbert, C. C. Ginocchio, K. E. Hanson, L. May, T. C. Quinn, *et al.*, *Clin. Infect. Dis.*, 2013, **57**, S139–S170.
- 2 C. A. Cornett, S. A. Vincent, J. Crow and A. Hewlett, *Journal of the American Academy of Orthopaedic Surgeons*, 2016, **24**, 11–18.
- 3 T. Karre, E. A. Vetter, J. N. Mandrekar and R. Patel, *J. Clin. Microbiol.*, 2010, **48**, 1504–1505.
- 4 A. T. Hassanain, L. Baha and M. Z. Zaini, *J. Clin. Microbiol.*, 2014, **52**, 3244–3249.
- 5 F. Bittar, H. Richet, J. C. Dubus, M. Rwynaud-Gaubert, N. Stremmer, J. Sarles, D. Raoult and J. M. Rolain, *PLoS One*, 2008, **7**, S41.
- 6 M. Tumbarello, T. Spanu, R. Di Bidino, M. Marchetti, M. Ruggeri, E. M. Treccarichi, G. De Pascale, E. M. Proli, R. Cauda, A. Cicchetti and G. Fadda, *Antimicrob. Agents Chemother.*, 2010, **54**, 4085–4091.
- 7 P. Werarak, P. Kiratisin and V. Thamlikitkul, *J. Med. Assoc. Thailand*, 2010, **93**, s126–s138.
- 8 I. F. Cheng, H. C. Chang, T. Y. Chen, C. Hu and F. L. Yang, *Sci. Rep.*, 2013, **3**, 2365.
- 9 K. Rebrošová, M. Šiler, O. Samek, F. Růžička, S. Bernatová, V. Holá, J. Ježek, P. Zemánek, J. Sokolová and P. Petráš, *Sci. Rep.*, 2017, **7**, 14846.
- 10 S. Y. Feng, W. B. Wang, T. I. Tai, G. N. Chen, R. Chen and H. S. Zeng, *Biomed. Opt. Express*, 2015, **6**, 3494–3502.
- 11 K. Lin, W. Zheng, C. M. Lim and Z. W. Huang, *Biomed. Opt. Express*, 2016, **7**, 3705–3715.
- 12 Z. M. Zhang, S. Chen, Y. Z. Liang, Z. X. Liu, Q. M. Zhang, L. X. Ding, F. Yei and H. Zhou, *J. Raman Spectrosc.*, 2010, **41**, 659–669.
- 13 P. Wang and W. R. Yao, *Sci. Technol. Food Ind.*, 2012, **33**, 427–430.
- 14 Y. H. Ong, M. Lim and Q. Liu, *Opt. Express*, 2012, **20**, 22158–22171.
- 15 S. Chen, X. Lin, C. Yuen, S. Padmanabhan, R. W. Beuerman and Q. Liu, *Opt. Express*, 2014, **22**, 12102–12114.
- 16 H. Haneishi, T. Hasegawa, A. Hosoi, Y. Yokoyama, N. Tsumura and Y. Miyake, *Appl. Opt.*, 2000, **39**, 6621–6632.
- 17 S. Chen and Q. Liu, *J. Biomed. Opt.*, 2012, **17**, 030501.
- 18 R. Piché, *J. Opt. Soc. Am. A*, 2002, **19**, 1946–1950.
- 19 S. Arlot and A. Celisse, *Stat. Surv.*, 2009, **4**, 40–79.
- 20 K. Chen, H. Zhang, H. Wei and Y. Li, *Appl. Opt.*, 2014, **53**, 5559–5569.
- 21 A. E. Villanueva-Luna, J. Castro-Ramos, S. Y. Montiel, A. Flores-Gil, J. A. D. Atencio and E. E. O. Guillén, *Optical Memory and Neural Networks*, 2010, **19**, 310–317.
- 22 S. Sidhik, *Optik*, 2015, **126**, 5952–5955.
- 23 Č. Martin, M. Pavel and V. Karel, *J. Raman Spectrosc.*, 2007, **38**, 1174–1179.





- 24 K. M. Omberg, J. C. Osborn, S. L. Zhang, J. P. Freyer, J. R. Mourant and J. R. Schoonover, *Appl. Spectrosc.*, 2002, **56**, 813–819.
- 25 B. D. Beier and A. J. Berger, *Analyst*, 2009, **134**, 1198–1202.
- 26 S. Chen, G. Wang, X. Y. Cui and Q. Liu, *Opt. Express*, 2017, **25**, 1005–1018.
- 27 S. Wold, K. Esbensen and P. Geladi, *Chemom. Intell. Lab. Syst.*, 1987, **2**, 37–52.
- 28 S. Meisel, S. Stöckel, M. Elschner, F. Melzer, P. Rösch and J. Popp, *Appl. Environ. Microbiol.*, 2012, **78**, 5575–5583.
- 29 S. Meisel, S. Stöckel, M. Elschner, P. Rösch and J. Popp, *Analyst*, 2011, **136**, 4997–5005.
- 30 M. J. B. Moester, F. Ariese and J. F. D. Boer, *J. Eur. Opt. Soc.*, 2015, **10**, 15022.
- 31 D. Sompel, E. Garai, C. Zavaleta and S. S. Gambhir, *J. Raman Spectrosc.*, 2013, **44**, 841–856.
- 32 S. Y. Feng, R. Chen, J. Lin, J. Pan, G. Chen, Y. Li, M. Cheng, Z. Huang, J. Chen and H. Zeng, *Biosens. Bioelectron.*, 2010, **25**, 2414–2419.
- 33 I. Notingher, S. Verrier, H. Romanska, A. E. Bishop, J. Polak and L. L. Hench, *Spectroscopy*, 2014, **16**, 43–51.
- 34 G. Rusciano, P. Capriglione, G. Pesce, P. Abete, V. Carnovale and A. Sasso, *Laser Phys. Lett.*, 2013, **10**, 075603.
- 35 A. Rygula, K. Majzner, K. M. Marzec, A. Kaczor, M. Pilarczyk and M. Baranska, *J. Raman Spectrosc.*, 2013, **44**, 1061–1076.
- 36 E. E. Rossi, A. L. Pinheiro, O. C. Baltatu, M. T. Pacheco and L. J. Silveria, *J. Photochem. Photobiol., B*, 2012, **107**, 73–78.
- 37 B. Lorenz, C. Wichmann, S. Stöckel, P. Rösch and J. Popp, *Trends Microbiol.*, 2017, **25**, 413–424.

