RSC Advances



View Article Online

View Journal | View Issue

CORRECTION

Check for updates

Cite this: RSC Adv., 2021, 11, 28106

DOI: 10.1039/d1ra90138a

bilayer MoS₂ nanopores with high efficiency Payel Sen^a and Manisha Gupta^{*b}

Correction: Single nucleotide detection using

rsc.li/rsc-advances

Correction for 'Single nucleotide detection using bilayer MoS₂ nanopores with high efficiency' by Payel Sen et al., RSC Adv., 2021, 11, 6114-6123, DOI: 10.1039/D0RA10222A

The authors would like to include additional sentences in their RSC Advances article to reference the updated ESI. The paragraph beginning on line four in the right hand column on page 6118 should read as follows:

Fig. 3a-h presents truncated single nucleotide peaks obtained for ML and BL MoS₂ nanopores for a direct comparison of dwell times. The raw, 100-fold upscaled and the data filtered at 20 kHz for all the different nucleotide translocations for ML and BL MoS₂ nanopores are shown in Fig. S12. The protocol used for the analysis is also described in the ESI⁺. It is observed that the dwell times are higher for BL as compared to those of the ML MoS₂ nanopores for all the different nucleotides. Blockade current is plotted as a function of dwell time for 3000 single nucleotide transport events in Fig. 3i. We observe four distinct blockade current regions for the different nucleotides. Thus, we can conclude that both ML and BL MoS₂ nanopores are capable of detecting single nucleotides. The blockade current for the nucleotide translocation is plotted as histograms to observe their distribution (ESI Fig. S4[†]). We observe normal distribution for all the nucleotides for both ML and BL nanopores. Thus, the mean blockade current values along with their standard deviations can be obtained.

The Royal Society of Chemistry apologises for these errors and any consequent inconvenience to authors and readers.

^aDepartment of Chemical and Materials Engineering, University of Alberta, Edmonton, Canada. E-mail: payel@ualberta.ca ^bDepartment of Electrical and Computer Engineering, University of Alberta, Edmonton, Canada. E-mail: mgupta1@ualberta.ca