


# Lab on a Chip

Devices and applications at the micro- and nanoscale

[rsc.li/loc](https://rsc.li/loc)



ISSN 1473-0197



Cite this: *Lab Chip*, 2023, 23, 1547

## 3D-printed capillarc ELISA-on-a-chip with aliquoting†

Azim Parandakh,<sup>iD</sup><sup>ab</sup> Oriol Ymbern,<sup>iD</sup><sup>‡ab</sup> William Jogia,<sup>‡ab</sup> Johan Renault,<sup>‡ab</sup> Andy Ng<sup>iD</sup><sup>ab</sup> and David Juncker<sup>iD</sup><sup>\*ab</sup>

Sandwich immunoassays such as the enzyme-linked immunosorbent assay (ELISA) have been miniaturized and performed in a lab-on-a-chip format, but the execution of the multiple assay steps typically requires a computer or complex peripherals. Recently, an ELISA for detecting antibodies was encoded structurally in a chip thanks to the microfluidic chain reaction (Yafia *et al. Nature*, 2022, **605**, 464–469), but the need for precise pipetting and intolerance to commonly used surfactant concentrations limit the potential for broader adoption. Here, we introduce the ELISA-on-a-chip with aliquoting functionality that simplifies chip loading and pipetting, accommodates higher surfactant concentrations, includes barrier channels that delay the contact between solutions and prevent undesired mixing, and that executed a quantitative, high-sensitivity assay for the SARS-CoV-2 nucleocapsid protein in 4×-diluted saliva. Upon loading the chip using disposable pipettes, capillary flow draws each reagent and the sample into a separate volumetric measuring reservoir for detection antibody (70  $\mu\text{L}$ ), enzyme conjugate (50  $\mu\text{L}$ ), substrate (80  $\mu\text{L}$ ), and sample (210  $\mu\text{L}$ ), and splits washing buffer into 4 different reservoirs of 40, 40, 60, and 20  $\mu\text{L}$ . The excess volume is autonomously drained via a structurally encoded capillarc aliquoting circuit, creating aliquots with an accuracy of >93%. Next, the user click-connects the assay module, comprising a nitrocellulose membrane with immobilized capture antibodies and a capillary pump, to the chip which triggers the step-by-step, timed flow of all aliquoted solutions to complete the assay in 1.5 h. A colored precipitate forming a line on a nitrocellulose strip serves as an assay readout, and upon digitization, yielded a binding curve with a limit of detection of 54 and 91 pg mL<sup>-1</sup> for buffer and diluted saliva respectively, vastly outperforming rapid tests. The ELISA chip is 3D-printed, modular, adaptable to other targets and assays, and could be used to automate ELISA in the lab; or as a diagnostic test at the point of care with the convenience and form factor of rapid tests while preserving the protocol and performance of central laboratory ELISA.

Received 22nd September 2022,  
Accepted 13th November 2022

DOI: 10.1039/d2lc00878e

rsc.li/loc

## Introduction

The enzyme-linked immunosorbent assay (ELISA) is utilized for the detection and quantification of proteins, antibodies, or antigens. The sandwich format with a capture antibody immobilized on the surface and a detection antibody applied in solution is used for assays requiring high sensitivity and specificity. Laboratory microplate ELISA still serves as a gold standard for assays and benefits from high sensitivity thanks to enzymatic signal amplification (down to sub-picomolar concentration for the best antibody pairs), quantitative results, standardized operations, off-the-shelf consumables,

and a comparably high throughput thanks to the use of 96-well plates. Long incubation times and copious washing between different steps to reduce non-specific binding and assay background are critical to achieving high assay sensitivity. However, the ELISA suffers from several downsides, such as being laborious, lengthy (~2–12 h depending on the protocol), requiring precise timing for each step, dependence on technical skills notably for adding and removing reagents (and thus susceptible to inter-operator variation) and necessitating a plate reader for signal readout.<sup>1,2</sup>

The miniaturization of ELISA has proceeded thanks to microfluidic lab-on-a-chip systems that can also automate the protocol.<sup>3,4</sup> Microfluidics successfully reduced the consumption of reagents and the total assay time while preserving assay performance. However, whereas the chips are small, they rely on bulky peripherals such as syringe pumps<sup>3</sup> or control motors,<sup>4</sup> and a computer or an instrument for operation.<sup>5,6</sup> Capillary phenomena and gravity have been

<sup>a</sup> Biomedical Engineering Department, McGill University, 740 Dr. Penfield Ave, Montreal, QC, H3A 0G1, Canada. E-mail: david.juncker@mcgill.ca

<sup>b</sup> McGill Genome Centre, McGill University, Montreal, QC, Canada

<sup>†</sup> Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d2lc00878e>

<sup>\*</sup> These authors contributed equally to this work.



harnessed to automate simple liquid manipulation, reducing or obviating the need for an external/active power supply.<sup>7–9</sup> For instance, a disk-like microfluidic platform (powered by a combination of centrifugal and capillary forces)<sup>8</sup> and a microfluidic siphon platform (powered by gravitational forces)<sup>9</sup> have been developed to carry out the common steps of a conventional ELISA with reduced reagents consumption and assay time while preserving assay sensitivity. Yet, both examples require multiple precise pipetting steps and timed user interventions for operation.

Sandwich assays can also be performed at point-of-care using so-called lateral flow assays (LFAs), also called rapid diagnostics, and are used globally for pregnancy tests and COVID-19 (coronavirus disease 2019) diagnosis. LFAs replace the enzyme amplification with conjugated colorimetric particles (either gold nanoparticles or polystyrene beads) that become visible to the naked eye upon accumulation. LFAs are simple to use as they only require the application of the sample, which flows thanks to capillarity without the need for peripherals, and produce a test result within a few minutes. However, LFAs offer only qualitative yes-no results, their sensitivity is typically lower compared to that of laboratory ELISA and are not suitable for archival as the readout must be completed within a few minutes of the test, because otherwise, the result can be compromised.<sup>10–12</sup> Enzymatic amplification has been implemented in the LFAs,<sup>13,14</sup> for instance by using a microfluidic interface,<sup>14</sup> to improve sensitivity. Yet, they cannot implement various fluidic handling tasks of common ELISA such as timed incubation of reagents and multiple rinsing steps between each incubation interval.

Paper-based microfluidics has been developed to introduce more advanced fluidic functions such as sequential delivery, additional rinsing step, and enzymatic amplification that collectively help improve assay sensitivity compared to LFAs.<sup>15–17</sup> Sponge actuators that upon swelling connect or disconnect different parts of a paper-based microfluidic circuit, along with flow paths with different lengths and resistance, have been used to time the delivery of multiple reagents for completing a *bona fide* ELISA.<sup>17</sup> However, these systems lacked the intermediate washing steps characteristic of classical ELISA, and undesired mixing of consecutive reagents occur at their mutual interface. Enzyme–substrate mixing may limit the potential for higher sensitivity as it could contribute to non-specific signal amplification.

Capillary circuits (CCs) are capillary microfluidics in microchannels designed and built using capillary elements which can automate liquid handling operations by pre-programming them structurally using capillary phenomena and powering them by capillary flow, without the need for peripheral equipment.<sup>18,19</sup> Multiple CCs have been designed to perform and automate ELISA with new functionality including flow reversal,<sup>18</sup> timing, reagent lyophilization,<sup>20</sup> and portable readers,<sup>21,22</sup> but they skip intermediate washing steps. Aliquoting a single solution into multiple reservoirs

has been shown for a nucleic acid test.<sup>23</sup> For an ELISA in a CC, multiple solutions including sample, buffer, and reagents must be serially delivered, and fluidically connected to effect fluid flow by hydraulic transmission of pressure differentials, which would subject them to unwanted mixing due to diffusion and/or convection, negatively affecting assay sensitivity and reliability.

The microfluidic chain reaction (MCR) introduces conditional initiation of capillary flow events, whereby event  $n$  is triggered after the preceding event  $n - 1$  has been completed, and completion of  $n$ , in turn, initiates event  $n + 1$ ; the condition is encoded using the so-called capillary domino valves.<sup>24</sup> MCR can drive hundreds of sequential flow operations robustly, thus opening new opportunities for CCs. CCs and MCRs are susceptible to failure in the presence of surfactants that reduce surface tension and contact angles. Yet surfactants are essential ingredients to assays, and often 0.05% Tween 20 is used to prevent non-specific binding.<sup>25</sup> In the initial MCR demonstration,<sup>24</sup> only 0.0125% was used because higher concentrations led to corner flow and trapping of air bubbles<sup>19</sup> and failure of stop valves. The initial MCR was used for a SARS-CoV-2 (severe acute respiratory syndrome coronavirus 2) antibody assay that could help determine the immune or vaccination status or detect prior infection, but which does not typically require high sensitivity needed for antigen assays where insufficient assay sensitivity often limits clinical usefulness.

User-friendliness is a critical feature for a device to be used at the point of need,<sup>26–28</sup> and accurate aliquoting and volumetric consistency are essential to reliable immunoassays.<sup>29,30</sup> Whereas in the lab they can be met using precision pipettes operated by technicians, they are difficult to achieve in a point-of-care setting. Both ELISA and the MCR CC introduced previously are dependent on precision pipetting. In particular, for the MCR chips,<sup>24</sup> reagent loading occurred *via* miniature inlets that require precise positioning of the pipette, precise volumetric control and hence precision laboratory pipettes, and controlled pipetting rates to match the capillary flow rate of the reservoir being filled. LFAs for COVID-19 often include a dropper and instructions for delivering a precise number of droplets.<sup>31</sup> Droppers are prone to occasional miscounting,<sup>31</sup> while for more complex assays multiple solutions with different volumes are required that could not be serviced using droppers to be supplied and aliquoted with different volumes.

Here, we introduce the ELISA chip that automates ELISA protocol on a chip using an MCR CC while preserving the washing steps used in classical ELISA. The ELISA chip eases the pipetting procedure thanks to the automated aliquoting of solutions and can accommodate higher surfactant concentrations commonly used in immunoassays. Akin to measuring spoons that are used to size ingredients in cooking, measuring reservoirs with different volumetric capacities are used to aliquot reagents, buffer, and sample.




Upon loading, solutions spontaneously fill their respective measuring reservoir, while an integrated capillary aliquoting circuit (CAC) autonomously drains excess liquid from all reservoirs simultaneously, forming precise aliquots. We describe the capillary circuitry and its components, and the step-by-step automated capillary flow operations of both aliquoting of all solutions and the sequential MCR-controlled ELISA protocol. We characterize the volumetric accuracy of aliquoting, the timing precision of delivery of the reagents for the ELISA, and the performance and limit of detection (LOD) of an assay for the detection of the SARS-CoV-2

nucleocapsid (N) protein spiked in buffer and 4×-diluted human saliva.

## Results and discussion

### ELISA chip

We designed a capillary ELISA chip without a moving part that automates aliquoting of sample, reagents, and buffer, and autonomously executes an ELISA protocol by flowing eight solutions in sequence according to predetermined flow rates using an MCR (Fig. 1A and S1†). The solutions are



**Fig. 1** 3D-printed ELISA chip with aliquoting for SARS-CoV-2 N protein assay in saliva. (A) Picture of the chip with superposed process flow graphics. (i<sub>1</sub>) Reagents (R<sub>1-3</sub>) and (i<sub>2</sub>) washing buffer (W) are added to the inlets using disposable, low-precision pipettes shown in (B) and fill the measuring reservoirs by capillary flow; the washing buffer is split into four reservoirs automatically (W<sub>1-4</sub>). (i<sub>3</sub>) Addition of the sample brings the drainage capillary pump to a fluidic connection, initiating drainage of excess solutions from all inlets via the capillary aliquoting circuit (CAC, see C), and producing (ii) final aliquots schematized by graduated tubes. (iii) Connection of the nitrocellulose strip triggers the ELISA via the sequential flow of sample, reagents, and washing buffer through the chip according to the propagation of the MCR. (iv) Following the enzymatic conversion of the substrate into a permanent line of a brown precipitate, the result can be read out by eye, and quantified following digitization. (B) Picture of a disposable pipette used to load the chip. (C) 3D design view of the ELISA chip with the structurally encoded capillary elements without moving parts. The CAC is highlighted in yellow, and insets showcase (i) inlet barrier channel and supply/drainage conduit and air vent, (ii) reservoir outlet, outlet barrier channel, main assay conduit, and air vent (iii) sample inlet and reservoir, main drainage conduit and connection to supply/drainage conduit, and drainage capillary pump.



delivered for the ELISA in the order of (i) sample, (ii) buffer, (iii) biotinylated detection antibody, (iv) buffer, (v) streptavidin poly-HRP (horseradish peroxidase), (vi) buffer, (vii) colorimetric substrate DAB (3,3'-diaminobenzidine) and (viii) buffer. The ELISA was developed for measuring the SARS-CoV-2 N protein in natural saliva and includes a strip of nitrocellulose membrane with a positive control line and a test line that will produce a permanent colorimetric signal proportional to the concentration of the N protein.

The need for the use of laboratory micropipettes is circumvented by integrating CAC into the chip which accepts the delivery of larger-than-needed volumes into the adequately sized inlets, followed by the spontaneous flow of the solutions *via* capillarity into serpentine measuring reservoirs for each solution. To complete the aliquoting process, the CAC removes excess volumes into a drainage capillary pump. The volume of each reservoir was set following assay optimization.

Simple squeeze pipettes, shown in Fig. 1B, can be used to load the chip while visually monitoring filling progression, and the measuring reservoirs form the aliquots (symbolized by graduated tubes) after draining excess liquid. The ELISA chip has five inlets servicing the eight reservoirs, three for the reagents, one for the buffer, and one for the sample, which need to be added in this order. The buffer is automatically split into four separate measuring reservoirs on the chip. The addition of the sample simultaneously fills the sample reservoir and initiates the drainage of excess volumes *via* the CAC. Next, upon click-connection of the main capillary pump, the step-by-step execution of the ELISA protocol is triggered, and the eight reservoirs chained by capillary domino valves are drained one by one as the MCR progresses. Finally, the signal is read out *via* the naked eye or using a scanner for quantification. A fluidic flow chart to explain the step-by-step manual user interventions and the ELISA chip's autonomous functions is provided in Fig. S2.†


The ELISA chip is made by additive manufacturing using an off-the-shelf 3D printer and light-based photopolymerization, followed by a series of post-processing steps for a total time of  $\sim 1$  h (Fig. 2; see Experimental section for additional details).

### Step-by-step loading and aliquoting

The ELISA chip loading and aliquoting are shown in Video S1.† Using a low-precision pipette (Fig. 1B), the user first deposits each reagent in the respective inlet in an arbitrary order with an excess of solution. The three serpentine reservoirs for the detection antibody, streptavidin poly-HRP, and DAB solutions have nominal measuring volumes of 70, 50, and 80  $\mu\text{L}$ , respectively. Fig. 3A and B show the loading of the detection antibody and are representative of the other reagents. Capillary flow fills the reservoir up to the trigger valve at the empty outlet barrier conduit, while excess solution remains in the inlet and fills the trigger valve connecting to the conduit at the inlet (called inlet barrier channel) that is part of the CAC (see Fig. 1C). Next, the washing buffer is delivered to the common inlet to the right of the ELISA chip and flows by capillarity through the supply/drainage conduit into the four washing buffer measuring reservoirs with a nominal volume of 40, 40, 60, and 20  $\mu\text{L}$  (Fig. 3C). The buffer simultaneously fills the main assay conduit, which is supplied by the vertical channel adjacent to the buffer inlet. The buffer distribution and splitting simplify chip loading as only one buffer solution needs to be added. However, the chip design imposes contact between buffer and reagents both at the inlet and outlet of the reagent reservoirs which could lead to unwanted mixing by diffusion and convection between various solutions.


### Suppressing unwanted mixing and bubble trapping

We isolated each of the reagents measuring reservoirs to prevent unwanted mixing between them and/or the sample prior to assay completion. We equipped them with two



**Fig. 2** Additive manufacturing of ELISA chip and post-processing steps. The manufacturing includes (A) ELISA chip design in computer-aided design software, (B) digital light processing-based 3D printing, (C) rinsing with isopropanol (IPA) to remove uncured resin, (D) drying under a stream of pressurized nitrogen gas ( $\text{N}_2$ ), (E) 1 min UV curing, (F) 10 s air plasma treatment and (G) sealing with a microfluidic tape with cut-out openings for servicing sample, buffer, and reagent inlets.





**Fig. 3** ELISA chip loading, filling into measuring reservoirs, and drainage of excess liquid by the CAC. Most figures show the part of the chip outlined by the dashed rectangle in Fig. 1C. (A and B) Loading of the detection antibody reagent into the inlet followed by capillary flow into the measuring reservoir and stop of the liquid at the capillary stop valves located next to the inlet and at the outlet. Loading of the other two reagents, *i.e.*, streptavidin poly-HRP and DAB, follows the same procedure. (C and D) Loading of buffer in buffer inlet and splitting into the buffer reservoirs supplied by the supply/drainage conduit. (E–H) Sample loading with the filling of the sample reservoir and simultaneous triggering of the drainage of all excess solutions on the chip *via* the CAC. See Video S1† and text for further details.

diffusion barrier channels located upstream (at the inlet) and downstream (at the outlet) of each reagent reservoir. By isolation, we suppress undesired preliminary mixing of reagents that could affect assay performance and reproducibility during both aliquoting (*i.e.*, filling and excess volume drainage), and subsequent ELISA sequential delivery (*e.g.* premature reaction between enzyme and colorimetric

substrate DAB could lead to unwanted background signal). Because diffusion will occur for a long time at the outlet of each reservoir, and for almost the entire assay duration particularly for the DAB solution, the last reagent to flow in the ELISA protocol, a long diffusion barrier is needed. By implementing diffusion barrier channels, we increase the distance between ELISA reagents to prevent premixing that



could occur by both diffusion and/or convection (e.g. small convective flows due to evaporation). The outlet barrier channel was designed with a U-turn and a separation length between the outlet of the reservoir and the main assay conduit with a length of 13–17 mm. The characteristic one-dimensional diffusion time is given as  $t = L^2/D$  with  $L$  = length,  $D$  = diffusion constant, and  $t$  = time.<sup>32</sup> For an IgG antibody with  $D \cong 36 \mu\text{m}^2 \text{s}^{-1}$  and DAB with  $D \cong 570 \mu\text{m}^2 \text{s}^{-1}$  (calculated by the Stokes–Einstein law for diffusion),<sup>32</sup> the time for diffusion across 10 mm is  $\sim 16$  days and  $\sim 1$  day, respectively. In our case, the functional connection connecting the reservoir and the barrier forms a constriction,<sup>24</sup> hence the diffusion time would be further increased relative to the calculation. The diffusion barriers thus ensure that no unwanted mixing occurs among reagents during the ELISA.

The inlet and outlet barrier channels form a separation between each reagent reservoir and the supply/drainage conduit (upstream), and main assay conduit (downstream) respectively. Both barrier channels remain empty during the autonomous filling of the reagents due to the action of trigger valves (see Fig. 3A), and autonomously fill by buffer after buffer injection (Fig. 3D). The inlet of the inlet (upstream) barrier conduit is located to the left of the reagent reservoir and branches out from the serpentine buffer reservoir, close to the outlet of the buffer reservoir. Upon reaching the branching point, capillary flow filling the buffer reservoir immediately branches into the barrier channel and fills it up to the extremity, which forms a dead-end. To prevent bubble trapping, an air vent is included that is connected to the channel *via* four stop valves, including three positioned upstream of the dead-end. At the reservoir outlet, the outlet barrier channel, which forms an extension to the buffer reservoir, is similarly filled by the buffer. However, the buffer reservoir to the right of the reagent channel supplies the corresponding outlet barrier channel and fills only after the buffer reservoir is filled completely (Fig. 3D bottom close-up, Video S1†). Akin to the inlet barrier channel, the extremity of the outlet barrier channels also forms a dead-end and is connected by three stop valves to a venting opening. Note that the design ensures that only pristine buffer fills each of the buffer measuring reservoirs. The inlet and outlet barrier channels are empty as the buffer starts flowing, and hence there is no contact or mixing between the reagents and the buffer flowing into the reservoirs.

The provision of multiple venting connections/stop valves is needed to accommodate liquids with low surface tension such as buffer containing a surfactant that is used in ELISA (e.g., Tween 20) and other assays. Low surface tension liquids have a lower contact angle than water (which has a high surface tension), which induces corner flow preceding the main filling front. The corner flow reaches the extremity of the conduit before the main filling front and can thus clog small air vents located at the extremity of dead-end channels before all the air can escape (see Fig. S3†). Upstream connections to the vent allow air to escape even after corner

flow reaches the dead-end and here prevent air bubbles from being trapped. For the ELISA chip, we evaluated the chip fluidic functionality under Tween 20 concentrations of 0.0125%, 0.025%, 0.05%, and 0.1%, and because of occasional failure at 0.1%, limited it to 0.05% for subsequent experiments.

### Excess volume draining


Drainage of the excess volumes is initiated by supplying the sample to the sample inlet (Fig. 3F). The sample inlet is connected to both the sample measuring reservoir and the main drainage conduit, both filling simultaneously by capillary flow. The outlet of the sample reservoir connects to a now pre-filled outlet barrier channel supplemented with an air vent to permit complete filling. The sample flowing into the main drainage conduit also initiates drainage of the excess volumes by triggering a trigger valve connecting the filled supply/drainage conduit to the main drainage conduit, and subsequently flowing to the drainage capillary pump, all part of the CAC. Initially, most of the drainage flow is supplied by the sample owing to excess solution in the sample inlet, the absence of a retention capillary pressure, and lower flow resistance. Once the sample inlet is emptied to the point where only corners remain filled, the drainage flow rates from the inlets containing excess buffer and reagents increase. The inlet barrier channel, in addition to minimizing the diffusion between the main drainage conduit and each reservoir, also forms the fluidic connection through which the excess reagent is drained from each reagent inlet.

### Flow path routing during drainage

Whereas conceptually simple, in CCs, drainage must both be triggered and stopped without external intervention. In addition, multiple possible drainage paths exist when the circuit is operating in the absence of active valves to close off sections of the circuit, yet to avoid unwanted mixing and drainage of solutions in the reservoirs, the liquid flow must flow through a pre-designed flow path, Fig. 4A. In the desired path, the excess reagents flow from the common inlet to the drainage capillary pump as follows: reagent inlet  $\rightarrow$  inlet barrier channel  $\rightarrow$  supply/drainage conduit  $\rightarrow$  main drainage conduit  $\rightarrow$  drainage capillary pump. To facilitate flow and favor this fluidic path while preserving other capillary functionality, we included 4 parallel connection conduits ( $0.1 \times 0.2 \times 0.75 \text{ mm}^3$ ) between the inlet barrier channel and the supply/drainage conduit (Fig. 1C close-up i), which yielded an overall flow resistance of  $135 \text{ Pa s mm}^{-3}$  for path i.

For each reagent inlet, three unwanted, leaky flow paths exist, as illustrated for the detection antibody in Fig. 4B. Flow through any of the parasitic paths should be disabled as they would result in unwanted reagents mixing and/or deviation from the pre-programmed volumes. Path ii proceeds from reagent inlet  $\rightarrow$  inlet barrier channel  $\rightarrow$  washing buffer





**Fig. 4** Drainage of excess volumes *via* the CAC. (A) The flow path for draining excess volumes in sample, reagent, and buffer inlets is shown with the black arrows. (B) In addition to the desired flow path i, three alternative, parasitic paths ii, iii, and iv are schematized using colored arrows. Drainage of reagent *via* a parasitic flow path could lead to mixing with other solutions and jeopardize the function of the ELISA chip. The design minimizes parasitic flow by ensuring that their flow resistance is much higher because they are longer and if necessary, sections with a small cross-section (and therefore a high flow resistance) are included. Note that in panel B, all flow paths are shown for only one reagent, and the same exist for other reagents as well.

reservoir → supply/drainage conduit, and path iii from reagent inlet → reagent reservoir → reagent outlet barrier channel → washing buffer reservoir → supply/drainage conduit to the main drainage conduit and drainage capillary pump. Both paths would result in reagents flowing into, and mixing with, the washing buffer in the adjacent reservoirs. To impede drainage through path ii, the resistance at the inlet of the inlet barrier channel as well as that of the corresponding washing buffer reservoir was increased, resulting in overall flow resistance of  $250 \text{ Pa s mm}^{-3}$  for path ii. Likewise, to impede drainage through path iii, the resistance at the inlet of the outlet barrier channel as well as that of the corresponding washing buffer reservoir was increased, thereby the overall resistance of path iii was  $550 \text{ Pa s mm}^{-3}$ .

Path iv follows the reagent inlet → reagent reservoir → reagent outlet barrier channel → main assay conduit → sample outlet barrier channel → sample reservoir → sample inlet → main drainage conduit → capillary drainage pump (Fig. 4B, path iv), with an overall resistance of  $700 \text{ Pa s mm}^{-3}$ . This drainage path would lead to mixing with the sample, yet

the last part of it being in charge of the excess drainage of the sample. This path only exists until the point where the excess sample is drained to the drainage capillary pump through the main drainage conduit (with an overall resistance of  $50 \text{ Pa s mm}^{-3}$ ) and is disconnected from the sample reservoir as the inlet is emptied.

Here we discussed the overall flow resistance of each of the four paths, but the parasitic flow is governed by the difference between the point where a given parasitic path splits from the drainage path i to the point where it merges again, which is a shorter distance. The ratio of flow resistance for the split section between path i, and each of paths ii, iii, and iv was calculated to be 7.8, 8.4, and 6.9 times higher respectively (Table S1†). Path iv is only active for a short time, and thus only contributes marginally. This design hence ensures that most of the excess volume will flow through path i.


### The automated ELISA protocol

The ELISA is initiated by the user click-connecting the main capillary pump to the chip (Fig. 5A), which wicks liquid from the CC and the reservoirs, and initiates the MCR that propagates from one reservoir to the next *via* the capillary domino valves that chain the reservoirs<sup>24</sup> (Video S2†). Briefly, the first reservoir is the sample reservoir that is capped by a retention burst valve (RBV) with a nominal capillary/burst pressure of  $-200 \text{ Pa}$  (*i.e.*, it will burst when the negative sucking pressure in absolute value is  $>200 \text{ Pa}$ ), which quickly bursts and results in draining of the reservoir. Air drawn in from the outside replaces liquid, and once all liquid is displaced, the capillary domino valve that forms an air link to the next reservoir is now connected to the outside *via* the just-emptied reservoir. At the top of the buffer reservoir, there is another RBV (here with the nominal capillary pressure of  $-215 \text{ Pa}$ ), which in turn bursts, leading to the draining of the first buffer reservoir, and in turn, opening the air link to the next (reagent) reservoir, and so on. The sequence of flow is thus sample, buffer, detection antibody, buffer, streptavidin poly-HRP, buffer, DAB, and buffer (Fig. 5B–F). Drainage of the main assay conduit is not controlled *via* the MCR because the chip lacked space for an additional capillary domino valve. Instead, we implemented an additional RBV with a nominal burst pressure of  $-550 \text{ Pa}$ . That is the RBV with the highest threshold within the CC, and hence the last one to burst. The ELISA chip illustrates the possibility of combining an MCR with an RBV to control the sequence of liquids flowing through an assay conduit. Upon draining the main assay conduit at the last step, the main capillary pump and therefore the nitrocellulose membrane, is fluidically disconnected from the microfluidic chip to complete the ELISA.

### ELISA chip over-loading capacity and overflow

We evaluated the volumetric operational window of the ELISA chip that preserves aliquoting and overall ELISA chip functionality. We analyzed the video of four ELISA chips loaded using precision pipettes with known volumes of





**Fig. 5** MCR progression of the ELISA chip to complete ELISA steps. (A) Click-connecting the main capillary pump to the chip (the region indicated by the black circle and arrow) triggers (B–G) sequential delivery of the sample, reagents, and corresponding washes to the sensor. Black arrows indicate the direction of flow. The black rectangle in panel A shows the small serpentine implemented in the ELISA chip to avoid mixing the sample with other reagents due to evaporation-induced backflow during sample delivery.

sample, reagents, and buffer, determined the minimum volume required to fill the measuring reservoirs, tested a range of excess volumes, and verified flow functionality while also monitoring unwanted mixing.

The nominal aliquot volume corresponds to the capacity of the measuring reservoir. However, in practice, some additional volume is needed to account for dead volumes in the chip. Conversely, if the inlets are overfilled, leakage into other parts of the chip could occur. We verified the nominal aliquoting accuracy by mapping the levels of the measuring reservoirs filled with food dyes after aliquoting was completed, and calculating the volumetric error by image analysis (see Experimental and Fig. S4† for detailed explanation). For the sample, the nominal aliquot volume is 210  $\mu\text{L}$  but at least 300  $\mu\text{L}$  is required because the sample triggers the CAC (as explained above). As this ELISA chip is designed for saliva, which was collected in an amount of 1 mL per individual, the additional volume requirement could be accommodated easily. We verified that the chips preserved their functionality and nominal aliquoting accuracy for a volume of 400  $\mu\text{L}$  (maximum tested). Within the 300–400  $\mu\text{L}$  range, the nominal accuracy of sample aliquoting was 99.5% with the coefficient of variation (CV) of 1.1%.

In the case of detection antibody, streptavidin poly-HRP, and DAB, the nominal aliquot volume is 70, 50, and 80  $\mu\text{L}$  respectively, while loading the chip with a micropipette showed that at least an extra of  $\sim 1$   $\mu\text{L}$  (for a total of 71, 51,

and 81  $\mu\text{L}$ , respectively) is needed to ensure complete filling of each reservoir with desired flow rate and without bubble entrapment. The maximum volumes that reagent inlets could accommodate while avoiding pre-mixing of reagents with buffer were 110, 90, and 120  $\mu\text{L}$  for the detection antibody, streptavidin poly-HRP, and DAB, respectively. Under these conditions, the nominal accuracy of aliquoting was found to be 99.7 (CV = 2.1%), 93.4 (CV = 1.5%), and 99.9% (CV = 3.5%) respectively. In a case that overloading exceeds the maximum volumes, the chip will continue to function, but some part of the excess may spill into other reservoirs, which was observed when loading 140  $\mu\text{L}$  of detection antibody led to spilling into the adjacent buffer reservoir (Fig. S5†).

For the washing buffer,  $\sim 300$   $\mu\text{L}$  is needed to fill all washing buffer reservoirs as well as the supply/drainage conduit, inlet/outlet barrier channels, and main assay conduit. We tested its operation for a volume of up to 400  $\mu\text{L}$ , successfully. Within this operating range, the buffer was reliably aliquoted into the four reservoirs of 40  $\mu\text{L}$ , 40  $\mu\text{L}$ , 60  $\mu\text{L}$ , and 20  $\mu\text{L}$  with a nominal accuracy of 98.7, 97.4, 98.3, and 94.7%, respectively, all with the CV of  $< 1\%$ . The results are summarized in Table 1. Considering these values, the ELISA chip outperforms the bulky pipetting robots utilized in laboratory ELISA, and indicates superior performance over the previously developed microfluidic devices while requiring significantly less user intervention.<sup>29,30</sup> The computed tomography data provided



**Table 1** ELISA chip over-loading, aliquoting, and flow timing performance

	Aliquoting <sup>a</sup>				Sequential delivery <sup>a</sup>
	Nominal aliquot volume (μL)	Loading range <sup>b</sup> (μL)	Visual aliquoted volume <sup>c,d</sup> (μL)	Nominal aliquoting accuracy <sup>e</sup> (%)	Duration <sup>d</sup> (min'sec'')
Sample	210	300–400 <sup>f</sup>	209.0 ± 2.4 (1.1%)	99.5	24'39" ± 40" (2.7%)
Detection antibody	70	71–110	70.2 ± 1.5 (2.1%)	99.7	8'52" ± 18" (3.4%)
Enzyme conjugate	50	51–90	53.3 ± 0.8 (1.5%)	93.4	6'29" ± 10" (2.5%)
Substrate	80	81–120	80.1 ± 2.8 (3.5%)	99.9	10'57" ± 22" (3.4%)
Wash #1	40	300–400 <sup>f</sup>	39.5 ± 0.1 (0.3%)	98.7	4'58" ± 7" (2.4%)
Wash #2	40		39.0 ± 0.3 (0.8%)	97.4	5'2" ± 6" (2.0%)
Wash #3	60		59.0 ± 0.2 (0.3%)	98.3	7'57" ± 12" (2.5%)
Wash #4	20		19.0 ± 0.3 (0.2%)	94.7	3'11" ± 6" (3.1%)

<sup>a</sup> Number of chips tested: 4. <sup>b</sup> Loading range that can be accommodated without impairing measuring function. <sup>c</sup> Volumes are measured by visual analysis of liquid within the chip and calculating volume based on 3D design (see Experimental for detailed explanation). <sup>d</sup> Reported in mean ± standard deviation and CV in parenthesis. <sup>e</sup> Nominal aliquoting accuracy (%) =  $\left(1 - \frac{|\text{Visual aliquoted volume} - \text{Nominal aliquot volume}|}{\text{Nominal aliquot volume}}\right) \times 100$ .

<sup>f</sup> Maximum volume tested.

in our previous study also confirm the high dimensional/volumetric precision of 3D-printed CCs with CV values of ≤3.9% in width and ≤1.2% in depth for the range of 400–1000 μm.<sup>24</sup> It should be noted that due to the dimensional precision, the error in accuracy can be compensated by fine-tuning the length of a corresponding reservoir to adjust the volume.

### Flow timing

We characterized the reliability of the timing of the ELISA protocol by measuring the duration of each step pre-programmed into the ELISA chip. The timing of the assay steps is an important parameter when considering the accuracy and reproducibility of ELISA. We analyzed the video of four ELISA chips loaded with assay reagents spiked with food dye colors for visualization purposes. Table 1 summarizes the draining/delivery duration for the sample, reagents, and buffer. Remarkably, we found the timing to be reproducible within a CV of 3.5% which confirms print-to-print precision on the smallest conduits which have a remarkably high flow resistance relative to large conduits, and where imprecision could thus disproportionately affect flow timing. We consider these results, both for the volumes and incubation times, excellent for a capillary-driven microfluidic system and suitable for conducting ELISA.

Sequential delivery of all solutions takes ~1.15 h. To reduce the effect of evaporation-induced backflow and the resultant mixing of the sample with other reagents, particularly during the ~25 min of sample draining, the size and length of the main assay conduit were expanded (see serpentine in the main assay conduit in Fig. 5A).

### SARS-CoV-2 N protein ELISA


To evaluate the performance of the ELISA chip, we used it to run a saliva-based SARS-CoV-2 N protein assay. We performed

an in-depth optimization process including testing various concentrations of capturing antibody, detection antibody, streptavidin poly-HRP, and Tween 20 as well as the extent of saliva dilution on the minimum detectable signal and assay background. In particular, we observed that increasing Tween 20 concentration improved the assay background, with 0.05% and 0.1% concentrations having comparable performance (Fig. S6†). Saliva dilutions of 1×, 2×, 4×, and 10×, were tested, and 4× was selected as it eased flow through the nitrocellulose membrane, preserved a low assay background, and yielded good assay-to-assay reproducibility without significantly compromising sensitivity (see Table S2† for a summary of the optimization process).

Akin to classical microplate ELISA, the ELISA chip allowed for the implementation of enzymatic amplification that necessitates a two-step process (*i.e.*, the addition of enzyme, followed by the substrate) which is not possible for the common LFAs that are typically carried out in one-step. DAB is the substrate used here which is chromogenic and oxidized in the presence of HRP, forming a brown precipitate at HRP locations. Whereas in laboratory ELISA the assay produces a colored solution, here it forms a precipitate on the paper strip which can be read out by the naked eye or digitized with a scanner.<sup>24</sup> While the read-out time for both ELISA and LFAs is limited to a window of a few minutes, the precipitate in the ELISA chip is stable and can be read out later, thus potentially also serving as an archival record.

We generated binding curves by spiking-in N protein of SARS-CoV-2 across 6 orders of magnitude of dilution from 1 to 10<sup>6</sup> pg mL<sup>-1</sup> in 4×-diluted pooled human saliva and ELISA buffer. Fitting the experimental data using a 4-parameter logistic regression, we obtained an LOD of 54 pg mL<sup>-1</sup> and 91 pg mL<sup>-1</sup> for the N protein in buffer and 4×-diluted pooled saliva, respectively (Fig. 6). The difference in LOD between buffer and saliva is due to the use of saliva as the diluent in the saliva binding curve, which





**Fig. 6** Calibration curves of the SARS-CoV-2 N protein assay enabled by the ELISA chip. Different concentrations of SARS-CoV-2 N protein were spiked in ELISA buffer or 4×-diluted pooled saliva. The resultant time-insensitive colorimetric signals were then captured by a regular scanner and yielded a binding curve that was fitted with a 4-parameter logistic regression. The number of replicates for each concentration is 3 for each calibration curve.

leads to a higher background signal and greater variation than buffer alone. The small dilution of human saliva together with the high sensitivity of the assay is a practical advantage of the developed ELISA-on-chip given the higher performance needed for reliable COVID-19 diagnosis based on saliva testing.<sup>33</sup>

We benchmarked the ELISA chip against two commercially available microplate-based ELISA for SARS-CoV-2 N protein detection (SinoBiological, Inc. and RayBiotech Life, Inc.). These ELISA kits use blood serum as the sample and have a time-to-result of ~5 h with an LOD of 35 pg mL<sup>-1</sup> and 70 pg mL<sup>-1</sup>, respectively, as reported by the manufacturer's protocol (Table 2). Note that the LODs of these serum ELISA were established based on a binding curve using buffer, and are thus comparable to the ELISA chip LOD for buffer; besides, RayBiotech calculated the LOD based on 2 × SD (standard deviation) above the blank instead of the

widely used 3 × SD. The LOD of the ELISA chip rivals one of the classical ELISA kits while being four times faster with less hands-on time and no time-sensitive manipulations of liquids, and in a format that is compatible with the point-of-care setting.

A study compared the performance of seven LFA rapid antigen tests for N protein spiked in phosphate buffer saline.<sup>34</sup> The tests notably included the widely used Abbott Panbio COVID-19 Ag Rapid Test and the Roche-SD Biosensor SARS-CoV Rapid Antigen Test. The most sensitive one was reported to be the R-Biopharm RIDA QUICK SARS-CoV-2 Antigen Test which yielded a line visible to the naked eye for a concentration as low as 2.5 ng mL<sup>-1</sup>.<sup>34</sup> The LOD of the ELISA chip is ~50 and ~25 times higher than this LFA test in buffer and 4×-diluted pooled saliva, respectively.

## Conclusions and future work

We presented an integrated ELISA chip that miniaturized and automated an ELISA-on-a-chip using capillaries and an MCR with an encoded aliquoting function, enabling the ELISA chip to be serviced with disposable pipettes. The ELISA chip generated aliquots with various volumes and timed the assay steps both within the CV of ≤3.5%, rivaling the common pipetting robots utilized in laboratory ELISA and other automated microfluidic ELISA.<sup>29,30</sup> The ELISA chip could operate with 0.05% Tween 20 commonly used in assays. The LOD of the ELISA chip for the SARS-CoV-2 N protein in 4×-diluted saliva was 91 pg mL<sup>-1</sup>, in line with classical microplate ELISA and outperforming conventional lateral flow assays by ~25×.

The ELISA chip could be 3D-printed and assembled in less than 1 h, and ~1200 chips were manufactured as part of this work. ELISA chips were designed with superficial channels only and may thus be adaptable to mass production by injection molding with much lower mass manufacturing costs than 3D printing.

In the future, the chip may be validated with patient samples in retrospective and possibly prospective studies.

**Table 2** Figure of merit of the ELISA chip in comparison to a commercial microplate ELISA kit (SARS-CoV-2 N protein detection ELISA kit from SinoBiological, Inc.) and an LFA rapid antigen test (R-Biopharm RIDA QUICK SARS-CoV-2 antigen)

	Sino biological SARS-CoV-2 N protein detection ELISA kit	R-Biopharm RIDA QUICK SARS-CoV-2 antigen test	SARS-CoV-2 N protein assay-on-ELISA Chip
Enzymatic amplification	Yes	No	Yes
Results type	Quantitative	Qualitative (yes/no)	Qualitative and quantitative
Readout timeframe	Within 20 min after assay (= temporary)	Within 15 min after assay (= temporary)	No limitation (= permanent)
Assay duration	~5 h	~20 min	~1.5 h
LOD	35 pg mL <sup>-1</sup>	2500 pg mL <sup>-1b</sup>	91 pg mL <sup>-1</sup>
Sample type	Serum	PBS <sup>a</sup>	Human saliva
Results interpretation	Plate reader	Direct reading	Direct reading or scanner
Precise pipetting	Required	Required	Not required
Hands-on time	~1 h	~1–5 min	~2 min

<sup>a</sup> PBS = phosphate buffer saline. <sup>b</sup> This value is the minimum concentration that could be detected by the naked eye, which is not equivalent to the LOD, a formulaic value.



The ELISA-on-chip introduced here will be particularly attractive for point-of-care applications where higher sensitivity than an LFA is needed while a longer assay time can be tolerated. Shortening assay time and further simplifying operations by pre-drying reagents and rehydrating them with a buffer solution<sup>20–22</sup> would further increase the usefulness of the ELISA chip. A more systematic optimization to accommodate higher surfactant concentrations might expand the range of assays that could be automated on the chip. Finally, the effect of temperature, which is known to modulate the enzymatic turnover, will need to be studied and accounted for prior to clinical use. Following these improvements, ELISA Chips could be deployed at the point-of-need and used by non-experts, and using a cell phone for imaging and quantifying the assay results,<sup>21,22,24</sup> quantitative, point-of-care tests with the performance of a central laboratory ELISA become available for everyone.

## Experimental

### ELISA chip fabrication and preparation

The chips were designed in AutoCAD (Autodesk), exported as “STL” files, and printed *via* an automated stereolithography 3D printer with the LED wavelength of 405 nm (Pr 110, Creative CADworks, Concord, Canada) using a monocure 3D rapid clear resin (Monocure 3D, NSW, Australia) with the following printing parameters: exposure time per layer: 2.5 s (10 s for the base layer); transition buffer layers: 2; layer thickness: 20 µm; printing delay: 1 min; and gap adjustment: 0.1 mm. Once printed, the chips were rinsed with isopropanol (Fisher Scientific, Saint-Laurent, Quebec, Canada) to wash away uncured resin, dried under a stream of pressurized nitrogen gas, cured for 1 min in a UV lamp (CureZone; Creative CADWorks; Concord; Canada), plasma treated for 10 s at 100% power (PE50 plasma chamber, Plasma Etch, Carson City, USA), and sealed with a microfluidic diagnostic tape (catalog number: 9795R; 3M Science. Applied to Life.™, Ontario, Canada).

A strip of Whatman CF4 paper (Cytiva, Marlborough, Massachusetts, United States) was clamped between 2 absorbent pads (Electrophoresis and Blotting Paper, Grade 320, Ahlstrom-Munksjo Chromatography, Helsinki, Finland) from the back end to collectively serve as the capillary pump. For the main capillary pump, a strip of Vivid™ 120 lateral flow nitrocellulose membrane (catalog number: VIV1202503R; Pall Corporation, Port Washington, USA) was clamped between the same absorbent pads from the back end and to a G041 glass fiber conjugate pad (Millipore Sigma, Oakville, Ontario, Canada) from the front end to facilitate connection to the chip.

### Nitrocellulose membrane

The strips of Vivid™ 120 lateral flow nitrocellulose membranes were designed in AutoCAD with the dimensions of 5.2 mm wide and 12 mm long and cut using the Silhouette

Portrait paper cutter (Silhouette, Lindon, USA). Membranes were stripped with a 5 mm-wide test line of SARS-CoV-2 N protein mouse monoclonal antibody (catalog number: 40143-MM08; Sino Biological, Inc., Beijing, China) at the concentration of 1 mg mL<sup>−1</sup> and a 5 mm-wide control line of bovine serum albumin (BSA)–biotin solution at the concentration of 50 µg mL<sup>−1</sup>, both delivered using a programmable inkjet spotter (sciFLEXARRAYER SX, Scienion, Berlin, Germany). The membranes were dried for 1 h at 37 °C and blocked by dipping into the blocking buffer solution (1% BSA and 0.1% Tween 20 in PBS) until completely wet, followed by shaking on a rocker for 60 min at 75 rpm. The membranes were then retrieved, dried in an oven for 1 h at 37 °C, and stored with a desiccant at 4 °C until use on the next day.

### SARS-CoV-2 N protein ELISA

Fresh saliva specimens were collected from three individuals using oral cotton swabs (Salivette, Sarstedt, Numbrecht, Germany) with their informed consent. The collected saliva samples were then pooled at a 1:1:1 ratio, filtered through a 0.22-micron filter, and diluted by a factor of 4 in the ELISA buffer solution (0.1% BSA and 0.05% Tween 20 in 1× PBS). The sample solutions were prepared by spiking SARS-CoV-2 N protein (catalog number: 40588-V08B; Sino Biological, Inc., Beijing, China) at the concentrations of 0, 1, 5, 10, 50, 10<sup>2</sup>, 10<sup>3</sup>, 10<sup>4</sup>, 10<sup>5</sup>, and 10<sup>6</sup> pg mL<sup>−1</sup> in either the ELISA buffer solution or 4×-diluted pooled saliva solution. The biotinylated SARS-CoV-2 N protein rabbit monoclonal antibody (catalog number: 40143-R004-B; Sino Biological, Inc.; Beijing, China) and streptavidin poly-HRP (Pierce; catalog number: 21140; ThermoFisher; Ottawa, Canada) solutions were prepared in the ELISA buffer solution both with the concentration of 7.5 µg mL<sup>−1</sup>. The substrate solution was prepared by dissolving SIGMAFAST™ DAB tablets (catalog number: D4293-50SET; Sigma-Aldrich; Oakville, Canada) in Milli-Q water. The washing buffer solution was the same as the ELISA buffer solution.

For benchmarking, the developed SARS-CoV-2 N protein assay in buffer was compared with the SARS-CoV-2 (2019-nCoV) Nucleocapsid Detection ELISA Kit (catalog number: KIT40588; Sino Biological, Inc.; Beijing, China) and the RayBio® COVID-19/SARS-COV-2 Nucleocapsid Protein ELISA Kit (catalog number: ELV-COVID19N; RayBiotech Life, Inc.; Peachtree Corners, United States).

### Nitrocellulose membranes image analysis and LOD calculation

After completion of the ELISA, the nitrocellulose strips were removed from the ELISA chip, left to dry at room temperature, and scanned at 1200 dpi in TIFF format (Epson Perfection V600) (see Fig. S7†). The images were imported in Photoshop (version: CS5 ME) and superposed with guide structures to locate the region of interest (2.5 × 0.4 mm) for the test line as well as the bottom background and top



background, each located 1.5 mm below and above the test line respectively. The superposed images were then imported into Fiji to measure the gray value of the three regions of interest for each nitrocellulose membrane (see Fig. S8†). For each concentration and the negative control, the local signal intensity of the test line was calculated by subtracting the gray value of the test line from the average gray value of the top and bottom local backgrounds. The relative signal intensity was then calculated by subtracting the local signal intensity of the test line from the average of the local signal intensity of the negative controls.

The experimental data were fitted using a 4-parameter logistic regression with the following equation:<sup>35</sup>

$$y = d + \frac{a - d}{1 + \left(\frac{x}{c}\right)^b} \quad (1)$$

where  $a$  and  $d$  are theoretical responses at zero and infinity, respectively,  $b$  denotes the slope factor (*i.e.*, Hill slope), and  $c$  represents the mid-point concentration (inflection point).<sup>35</sup>

The LOD was then determined by adding 3 standard deviations to the mean relative signal intensity of the blank samples (*i.e.*, zero antigen concentration) and calculating the corresponding concentration from the established calibration curve.<sup>36</sup>

### ELISA chip video recording and still image capture

For video recording (Panasonic Lumix DMC-GH3K), the ELISA chips were loaded with the ELISA buffer solution in the reagents and washing buffer inlets and with the 4×-diluted pooled saliva solution in the sample inlet. Both solutions were colored with food dye for visualization purposes unless stated otherwise.

Videos were edited in Adobe Premiere Pro (version: 22.1.2) to adjust brightness, contrast, sharpness, and speed. Still images were captured using the Sony α7R III camera and edited in Adobe Photoshop for brightness, contrast, and sharpness.

### Characterization of ELISA chip aliquoting and maximum over-loading capacity

To characterize aliquoting, screenshots of the chips following the completion of aliquoting were analyzed in Fiji. The regions corresponding to the extra or lost solution shown in Fig. S4† were mapped, and the equivalent volume was calculated based on the 3D design file. The visual aliquoted volume and the nominal aliquoting accuracy were then calculated as follows:

$$\text{Visual aliquoted volume} = \text{Nominal aliquot volume} - \text{lost volume} + \text{extra volume} \quad (2)$$

where the nominal aliquot volume is equal to the capacity of the reservoir in the 3D design.

Nominal aliquoting accuracy

$$= 1 - \frac{|\text{Visual aliquoted volume} - \text{Nominal aliquot volume}|}{\text{Nominal aliquot volume}} \times 100 \quad (3)$$

To characterize the maximum overloading capacity, the chips were loaded with exact volumes of sample, reagents and washing buffer using laboratory micropipettes with an increment of 5 μL for each solution. The recorded videos were analyzed visually to investigate any unwanted mixing or spilling of reagents to the adjacent reservoirs.

### Characterization of ELISA chip timing

The videos of the ELISA chips were recorded, and the timing of each step in the MCR read from the video data was tabulated and the duration of different steps was calculated.

### Capillary pressure and resistance calculation

Capillary pressure was calculated using the Young-Laplace equation as follows:<sup>19</sup>

$$P = -\gamma \left( \frac{\cos \theta_t + \cos \theta_b}{h} + \frac{\cos \theta_l + \cos \theta_r}{w} \right) \quad (4)$$

where  $P$  denotes the capillary pressure,  $\gamma$  represents the liquid surface tension,  $h$  and  $w$  are the microchannel height and widths, and  $\theta_t$ ,  $\theta_b$ ,  $\theta_l$ , and  $\theta_r$  respectively are the contact angle of the liquid with the top, bottom, left, and right wall of the microchannel.<sup>19</sup> For the ELISA buffer and 4×-diluted saliva, the liquid surface tension  $\gamma$  was measured to be 37.8 and 38.0 mN m<sup>-1</sup> respectively using a dynamic tensiometer (DCAT11; Filderstadt; Germany). A custom-built apparatus was utilized for the contact angle measurement. Each  $\theta_b$ ,  $\theta_l$ , and  $\theta_r$  was replaced by the contact angle of ELISA buffer (10° ± 1.5°) (mean ± SD) or 4×-diluted saliva (12.5° ± 2°) on the plasma-treated, flat 3D printed samples.  $\theta_t$  was replaced by the contact angle of the ELISA buffer (86.2° ± 3.4°) or 4×-diluted saliva (88.7° ± 1.9°) on the microfluidic diagnostic tape. As reported in our previous study, flat 3D printed monocyte has a native contact angle of 77° ± 3° before, and 33° ± 5° after plasma treatment for MilliQ water.<sup>24</sup>

To calculate the resistance of a fluidic path, a lumped-element model was created where each section of the circuit was assigned a resistance calculated using the following equation:<sup>19</sup>

$$R = 12\eta \frac{L}{h^3 w} \left( 1 - 0.63 \frac{h}{w} \right)^{-1} \quad (5)$$

where  $R$  denotes the resistance,  $\eta$  represents the liquid viscosity, and  $h$  is the height of the microchannel. Then, for circuit elements in series or parallel, the equivalent resistance ( $R_{eq}$ ) was calculated as follows:

$$R_{eq, \text{Series}} = R_1 + R_2 + \dots + R_N \quad (6)$$



$$\frac{1}{R_{\text{eq.Parallel}}} = \frac{1}{R_1} + \frac{1}{R_2} + \dots + \frac{1}{R_N} \quad (7)$$

where  $N$  denotes the number of capillary elements included in the corresponding fluidic path. The liquid viscosity,  $\eta$ , was measured to be 1.11 mPa s (temperature: 16.9 °C) for the ELISA buffer and 1.07 mPa s (temperature: 17.7 °C) for the 4×-diluted saliva using a viscometer (SV-10; A&D Company Ltd; Tokyo; Japan).

## Author contributions

A. P. and O. Y. designed the ELISA chip. A. P. characterized the ELISA chip. A. P., J. R. W. J., and O. Y. optimized the SARS-CoV-2 N protein assay. A. P. and D. J. wrote the initial draft of the manuscript. A. P. and O. Y. prepared the figs. A. P., O. Y., A. N., and D. J. reviewed and edited the manuscript. A. P. and D. J. analyzed the data. D. J. conceptualized, administered, and supervised the work.

## Conflicts of interest

Authors declare that they have no conflicts of interest.

## Acknowledgements

We thank Geunyoung Kim and Molly Shen for their assistance in assay optimization. We thank Mohamed Yafia for his constructive feedback regarding chip design. We thank Galyna Shul from NanoQAM, UQAM for her assistance in operating the viscometer and tensiometer. This work was supported by the NSERC Alliance COVID-19 Grant (grant number: ALLRP 551058-20), NSERC Discovery Grant (grant number: RGPIN-2016-06723), and the McGill MI4 Emergency COVID-19 Research Funding Grant. D. J. acknowledges support from a Canada Research Chair in Bioengineering (grant number: CRC-232159).

## References

- 1 S. D. Gan and K. R. Patel, Enzyme immunoassay and enzyme-linked immunosorbent assay, *J. Invest. Dermatol.*, 2013, **133**, e12.
- 2 S. Hosseini, P. Vázquez-Villegas, M. Rito-Palomares and S. O. Martinez-Chapa, Advantages, Disadvantages and modifications of conventional ELISA. *SpringerBr. Appl. Sci. Technol.*, 2018, 67–115.
- 3 F. Costantini, C. Sberna, G. Petrucci, C. Manetti, G. de Cesare, A. Nascetti and D. Caputo, Lab-on-chip system combining a microfluidic-ELISA with an array of amorphous silicon photosensors for the detection of celiac disease epitopes, *Sens. Bio-Sens. Res.*, 2015, **6**, 51–58.
- 4 M. J. Uddin, N. H. Bhuiyan and J. S. Shim, Fully integrated rapid microfluidic device translated from conventional 96-well ELISA kit, *Sci. Rep.*, 2021, **11**, 1–9.
- 5 L. Zhu, Y. Feng, X. Ye, J. Feng, Y. Wu and Z. Zhou, An ELISA chip based on an EWOD microfluidic platform, *J. Adhes. Sci. Technol.*, 2012, **26**, 2113–2124.
- 6 A. H. C. Ng, *et al.*, A digital microfluidic system for serological immunoassays in remote settings, *Sci. Transl. Med.*, 2018, **10**, eaar6076.
- 7 J. T. Heggstad, *et al.*, Multiplexed, quantitative serological profiling of COVID-19 from blood by a point-of-care test, *Sci. Adv.*, 2021, **7**, 4901–4926.
- 8 S. Laschi, R. Miranda-Castro, E. González-Fernández, I. Palchetti, F. Reymond, J. S. Rossier, G. Marrazza and S. Laschi, A new gravity-driven microfluidic-based electrochemical assay coupled to magnetic beads for nucleic acid detection, *Electrophoresis*, 2010, **31**, 3727–3736.
- 9 N. M. Reis, S. H. Needs, S. M. Jegouic, K. K. Gill, S. Sirivisoot, S. Howard, J. Kempe, S. Bola, K. Al-Hakeem, I. M. Jones and T. Prommool, Gravity-driven microfluidic siphons: fluidic characterization and application to quantitative Immunoassays, *ACS Sens.*, 2021, **6**, 4338–4348.
- 10 K. M. Koczula and A. Gallotta, Lateral flow assays, *Essays Biochem.*, 2016, **60**, 111–120.
- 11 G. A. Posthuma-Trumpie, J. Korf and A. Van Amerongen, Lateral flow (immuno)assay: Its strengths, weaknesses, opportunities and threats. A literature survey, *Anal. Bioanal. Chem.*, 2009, **393**, 569–582.
- 12 Y. Liu, L. Zhan, Z. Qin, J. Sackrison and J. C. Bischof, Ultrasensitive and highly specific lateral flow assays for point-of-care diagnosis, *ACS Nano*, 2021, **15**, 3593–3611.
- 13 C. Parolo, A. de la Escosura-Muñiz and A. Merkoçi, Enhanced lateral flow immunoassay using gold nanoparticles loaded with enzymes, *Biosens. Bioelectron.*, 2013, **40**, 412–416.
- 14 Y. Panraksa, I. Jang, C. S. Carrell, A. G. Amin, O. Chailapakul, D. Chatterjee and C. S. Henry, Simple manipulation of enzyme-linked immunosorbent assay (ELISA) using an automated microfluidic interface, *Anal. Methods*, 2022, **14**, 1774–1781.
- 15 M. S. Verma, M. N. Tsaloglou, T. Sisley, D. Christodouleas, A. Chen, J. Milette and G. M. Whitesides, Sliding-strip microfluidic device enables ELISA on paper, *Biosens. Bioelectron.*, 2018, **99**, 77–84.
- 16 M. M. Gong and D. Sinton, Turning the Page: Advancing Paper-Based Microfluidics for Broad Diagnostic Application, *Chem. Rev.*, 2017, **117**, 8447–8480.
- 17 B. J. Toley, J. A. Wang, M. Gupta, J. R. Buser, L. K. Lafleur, B. R. Lutz, E. Fu and P. Yager, A versatile valving toolkit for automating fluidic operations in paper microfluidic devices, *Lab Chip*, 2015, **15**, 1432–1444.
- 18 R. Safavieh and D. Juncker, Capillaries: pre-programmed, self-powered microfluidic circuits built from capillary elements, *Lab Chip*, 2013, **13**, 4180–4189.
- 19 A. Olanrewaju, M. Beaugrand, M. Yafia and D. Juncker, Capillary microfluidics in microchannels: from microfluidic networks to capillary circuits, *Lab Chip*, 2018, **18**, 2323–2347.
- 20 O. Gökçe, S. Castonguay, Y. Temiz, T. Gervais and E. Delamarche, Self-coalescing flows in microfluidics for pulse-shaped delivery of reagents, *Nature*, 2019, **574**, 228–232.
- 21 T. U. Vinitha, S. Ghosh, A. Milleman, T. Nguyen and C. H. Ahn, A new polymer lab-on-a-chip (LOC) based on a microfluidic capillary flow assay (MCFA) for detecting unbound cortisol in saliva, *Lab Chip*, 2020, **20**, 1961–1974.



- 22 E. E. Ahi, H. Torul, A. Zengin, F. Sucularlı, E. Yıldırım, Y. Selbes, Z. Suludere and U. Tamer, A capillary driven microfluidic chip for SERS based hCG detection, *Biosens. Bioelectron.*, 2022, **195**, 113660.
- 23 N. Ramalingam, H. B. Liu, C. C. Dai, Y. Jiang, H. Wang, Q. Wang, K. M. Hui and H. Q. Gong, Real-time PCR array chip with capillary-driven sample loading and reactor sealing for point-of-care applications, *Biomed. Microdevices*, 2009, **11**, 1007–1020.
- 24 M. Yafia, O. Ymbern, A. O. Olanrewaju, A. Parandakh, A. S. Kashani, J. Renault, Z. Jin, G. Kim, A. Ng and D. Juncker, Microfluidic Chain Reaction of Structurally Programmed Capillary Flow Events, *Nature*, 2022, **605**, 464–469.
- 25 M. Steinitz, Quantitation of the Blocking Effect of Tween 20 and Bovine Serum Albumin in ELISA Microwells, *Anal. Biochem.*, 2000, **282**, 232–238.
- 26 P. B. Lippa, C. Müller, A. Schlichtiger and H. Schlebusch, Point-of-care testing (POCT): Current techniques and future perspectives, *TrAC, Trends Anal. Chem.*, 2011, **30**, 887–898.
- 27 V. Gubala, L. F. Harris, A. J. Ricco, M. X. Tan and D. E. Williams, Point of care diagnostics: Status and future, *Anal. Chem.*, 2012, **84**, 487–515.
- 28 S. Sachdeva, R. W. Davis and A. K. Saha, Microfluidic point-of-care testing: Commercial landscape and future directions, *Front. Bioeng. Biotechnol.*, 2021, **8**, 602659.
- 29 C. D. Chin, T. Laksanasopin, Y. K. Cheung, D. Steinmiller, V. Linder, H. Parsa, J. Wang, H. Moore, R. Rouse, G. Umvilighozo and E. Karita, Microfluidics-based diagnostics of infectious diseases in the developing world, *Nat. Med.*, 2011, **17**, 1015–1019.
- 30 Z. He, J. Huffman, K. Curtin, K. L. Garner, E. C. Bowdridge, X. Li, T. R. Nurkiewicz and P. Li, Composable microfluidic plates (cPlate): A simple and scalable fluid manipulation system for multiplexed enzyme-linked immunosorbent assay (ELISA), *Anal. Chem.*, 2021, **93**, 1489–1497.
- 31 M. Jing, R. Bond, L. J. Robertson, J. Moore and A. Kowalczyk, User experience of home-based AbC-19 SARS-CoV-2 antibody rapid lateral flow immunoassay test, *Sci. Rep.*, 2022, **12**, 1–18.
- 32 C. C. Miller, The Stokes-Einstein law for diffusion in solution, *Proc. R. Soc. London, Ser. A*, 1924, **106**, 724–749.
- 33 E. S. Savelaz, *et al.*, Quantitative SARS-CoV-2 viral-load curves in paired saliva and nasal swabs inform appropriate respiratory sampling site and analytical test sensitivity required for earliest viral detection, *J. Clin. Microbiol.*, 2022, **60**, e01785-21.
- 34 V. M. Corman, *et al.*, Comparison of seven commercial SARS-CoV-2 rapid point-of-care antigen tests: a single-centre laboratory evaluation study, *Lancet Microbe*, 2021, **2**, 311–319.
- 35 A. De Lean, P. J. Munson and D. Rodbard, Simultaneous analysis of families of sigmoidal curves: application to bioassay, radioligand assay, and physiological dose-response curves, *Am. J. Physiol.*, 1987, **4**, E:97.
- 36 C. K. Dixit, S. K. Vashist, B. D. MacCraith and R. O'Kennedy, Multisubstrate-compatible ELISA procedures for rapid and high-sensitivity immunoassays, *Nat. Protoc.*, 2011, **6**, 439–445.

