


Cite this: *RSC Adv.*, 2025, 15, 9129

An indolium inspired portable colorimetric sensor for cyanide recognition in environmental samples with smartphone integration†

Kiran,^a Anju Ranolia,^a Priyanka,^a Anil Duhan,^a Rahul Kumar Dhaka,^a Snigdha Singh,^b Gaurav Joshi,^c Parvin Kumar,^d Devender Singh,^e Muhammad Wahajuddin^f and Jayant Sindhu^{*a}

Cyanide is a highly hazardous and fast-acting blood agent widely used in various industries, making its monitoring crucial due to its severe impact on living organisms. Considering this, we had developed a novel indolium based low-cost and portable sensor for the colorimetric and fluorogenic detection of cyanide ions. The probe displayed a distinct visual color change and selectively exhibits a fluorogenic “turn-off” response to cyanide ions in pure aqueous medium. The chemodosimetric-approach-based detection of cyanide ions has been confirmed using ¹H NMR titration and DFT calculations. The presence of competing anions has minimal or no impact on the cyanide recognition process. The probe displayed an excellent detection limit as low as 3.78 nM with a response time of 5 s. The aqueous phase and optimum pH workability are some of the promising features of the developed probe. Additionally, a visual colorimetric strip based test was also developed using probe impregnated filter paper strips, which displayed a visual color change from red to colourless upon cyanide exposure. In addition to this, these test strips also showed excellent selectivity for cyanide ions, with no interference from fluoride or acetate ions. Furthermore, a smartphone-based protocol was developed to record the color change, demonstrating significant potential for cyanide detection without using sophisticated instrumentation. The formulation of the probe into an INHIBIT logic gate for facilitating cyanide recognition using electronic devices is one of the important applications.

Received 24th January 2025
Accepted 12th March 2025

DOI: 10.1039/d5ra00576k

rsc.li/rsc-advances

Introduction

The harmful effects of cyanide, often referred to as “hydrocyanic acid” or “prussic acid” have been recognized for centuries.^{1–4} Its high volatility and low molecular weight make it very dispersive. Cyanides are utilized across the globe in numerous industrial applications including textile production, paint manufacturing, metal extraction, fertilizer production, coating processes, disinfectant production, organic synthesis, mining, photography and the pharmaceutical sector.^{5–8} Cyanide dependent industries produce 2–3 million tons of annual cyanide waste as reported by Towill *et al.*⁹ Additionally, cyanide is naturally

produced in small amounts as cyanogenic glycosides by plants such as bitter almonds, coffee, chickpeas, apples, peaches, apricots, plums, cherries and other members of the Rosaceae family.^{10–14} The seeds of these foods contain a compound called amygdalin which upon hydrolysis produces HCN.¹⁵ Moreover, HCN gas is generated during the combustion of materials such as wool, silk and polyurethane. An average of 17.56–1553.98 µg of HCN was produced by cigarette smoke.¹⁶ Sodium nitroprusside, used to manage acute hypertensive crises, leads to the formation of cyanide. Rapid administration and prolonged use at high doses can result in toxic effects or kidney failure. Irreversible binding of a cyanide ion to the ferric ion (Fe³⁺) of cytochrome oxidase results into its deactivation, thus preventing oxygen exchange in the tissues, which ultimately leads to death.^{17,18} This interaction process is different from CO poisoning, as this leads to Hypoxia. Due to its harmful effects on human health and the environment, the real time monitoring of cyanide is crucial in both industrial and research settings, prompting ongoing efforts to establish new analytical strategies.^{19–24}

The various analytical techniques such as voltammetry, ion chromatography, potentiometry, titrimetry and spectrophotometry are used for the determination of cyanide ion.^{25–29}

^aDepartment of Chemistry, COBS&H, CCSHAU, Hisar-125004, India

^bDepartment of Chemistry, Delhi University, Delhi-11007, India

^cDepartment of Pharmaceutical Science, Hemvati Nandan Bahuguna Garhwal University, Uttarakhand-244713, India

^dDepartment of Chemistry, Kurukshetra University, Kurukshetra, Haryana-136119, India

^eDepartment of Chemistry, Maharshi Dayanand University, Rohtak-124001, India

^fInstitute of Cancer Therapeutics School of Pharmacy and Medical Sciences, University of Bradford, UK. E-mail: jayantchem@gmail.com; m.wahajuddin@bradford.ac.uk

† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d5ra00576k>



While these methods fulfil legal requirements, many of these requires sophisticated equipment and are time-consuming, making them inappropriate for on-site detection. Thus, there is a crucial demand for development of rapid, inexpensive, selective and sensitive methodology for cyanide detection. Recently, colorimetric and fluorescence-based methodology has gained significant attention of scientific community in various fields such as environmental chemistry, biology and biochemistry,^{30–34} owing to their advantages, including on-site monitoring, low cost, ease of use, visual and rapid detection, high sensitivity and selectivity.^{35–39} Additionally, the field of fluorescence sensing has been expanded to a great extent with the development of compact and portable devices to meet the increasing demand for field applications.^{40–46}

Indolium-based organic compounds, particularly indolium dyes, play a crucial role in cyanide detection due to their exceptional sensitivity and specificity.⁴⁷ When exposed to cyanide ions, these compounds exhibit noticeable colorimetric or fluorometric shifts, making them highly effective for both visual and instrumental analysis.⁴⁸ Other cyanide-detecting agents include Schiff bases, porphyrins, metal complexes (such as those incorporating copper or cobalt) and fluorescent probes like benzopyrylium salts and naphthalimides. These substances operate through diverse mechanisms, including nucleophilic addition, metal coordination and redox reactions, ensuring fast and precise cyanide detection in environmental and biological applications.^{49–53} However, some CN[−] chemosensors may exhibit interference with other anions including F[−], CH₃COO[−] and PO₄^{3−} ions. The inability of organic chromophores to work in aqueous medium further restrict their use to lab based studies only. Hence, there is a strong need to develop selective and sensitive chemosensors which are operative in aqueous medium.

In our previous work, we have synthesized a novel sensor for cyanide ion with partial aqueous workability.^{54,55} Therefore, in continuation of our research work and in order to achieve aqueous phase workability, we had reported a novel naked eye and UV-light visible probe synthesized by the Knoevenagel condensation reaction (**ADTI**) for the real time monitoring of toxic cyanide ion in 100% aqueous medium. **ADTI** exhibits remarkable sensitivity and selectivity for CN[−] ion in real samples with low LOD. Filter paper based test strips and smartphone based protocol were developed for on-site detection of cyanide ion.

Experimental

Materials and instrumentation

The organic building blocks including 2,3-dichloronaphthoquinone, pyridine, acetyl acetone, pyridine-4-carbaldehyde and indolium salt utilised in the present study were commercially procured from Sigma-Aldrich and used without further purification. Similarly, tetra-butylammonium salts for a range of anions, including cyanide, fluoride, chloride, iodide, nitrate, hydrogen sulfate, hexafluorophosphate, tetrafluoroborate, hydrogen sulphate and acetate were also purchased from Sigma-Aldrich. Both ¹H and ¹³C NMR were recorded on Avance Neo 400

MHz NMR spectrometer in DMSO-*d*₆. The FT-IR spectrometer from PerkinElmer (FL 6500) and Shimadzu LCMS-IT-TOF spectrometer were utilised for recording the infrared and mass spectra of the compound, respectively. All the absorption and emission studies were performed using Shimadzu UV-1900i and PerkinElmer 6500 spectrophotometer, respectively. The emission spectra of the compound were recorded at an excitation wavelength of 470 nm using a slit width and gain of 5 nm and 5, respectively. A test kit for the detection of cyanide was prepared using Whatman filter paper.

Synthesis of 12-acetyl-6,11-dioxo-6,11-dihydrobenzo[*f*]pyrido[1,2-*a*]indole-2-carbaldehyde (ADC)

The reaction was attempted in a round bottomed flask (100 mL) containing a mixture of 2,3-dichloronaphthoquinone (**1**) (0.022 mmol), pyridine-4-carbaldehyde (**2**) (0.11 mmol) and acetyl acetone (**3**) (0.11 mmol) in EtOH (10 mL). The content of the flask was refluxed for 5 h in a pre-heated oil bath placed over magnetic stirrer. Thin layer chromatography was performed on pre-coated aluminium plates (TLC Merck) using EA : hexane (2 : 8, v/v) to record the reaction progress (Scheme 1). After complete consumption of reactants, the heating source was turned off and the reaction flask was clamped and suspended over an oil bath to allow it to attain ambient temperature. After cooling, the resulting solid was filtered, washed with ethanol and air dried to afford a red coloured solid (**ADC**) in 90% yield. The final composition of the compound was confirmed using mass spectrometry (*m/z* calculated 318.0758 [M + H]⁺; found 318.0763 [M + H]⁺).


Synthetic protocol for (*E*)-2-(2-(12-acetyl-6,11-dioxo-6,11-dihydrobenzo[*f*]pyrido[1,2-*a*]indol-2-yl)vinyl)-1,3,3-trimethyl-3*H*-indol-1-ium iodide (**ADTI**)

To a stirred solution of **ADC** (0.05 mmol) in CH₃COOH (15 mL) contained in a round bottomed flask (100 mL), 1,2,3,3-tetramethyl-3*H*-indol-1-ium iodide (**4**) (0.05 mmol) was added. The content of the flask was heated to reflux for 8 h. The reaction was monitored using TLC (DCM/MeOH (20 : 1, v/v)). During the course of the reaction, a solid product gradually separates out (Scheme 1). After complete consumption of the reactants, the reaction flask was kept at ambient temperature and the separated solid was filtered, washed with methanol and air dried to afford (*E*)-2-(2-(12-acetyl-6,11-dioxo-6,11-dihydrobenzo[*f*]pyrido[1,2-*a*]indol-2-yl)vinyl)-1,3,3-trimethyl-3*H*-indol-1-ium iodide (**ADTI**) in 92% yield.

Spectral details

Red solid; yield: 92%; Mp: >300 °C; *R*_f (DCM/MeOH (20 : 1, v/v)) = 0.375; ¹H NMR (400 MHz, DMSO-*d*₆) δ_H 9.68 (d, *J* = 8 Hz, 1H), 8.91 (s, 1H), 8.55 (d, *J* = 16 Hz, 1H), 8.25 (dd, *J* = 8 & 2 Hz, 1H), 8.16 (m, 2H), 8.00–7.85 (m, 4H), 7.80 (d, *J* = 16 Hz, 1H), 7.68 (m, 2H), 4.24 (s, 3H), 2.81 (s, 3H), 1.84 (s, 6H); ¹³C NMR (100 MHz, DMSO-*d*₆) δ_C 196.83, 182.30, 181.87, 175.11, 149.36, 144.56, 142.06, 137.31, 134.98, 134.62, 134.46, 133.95, 133.92, 129.87, 129.28, 127.78, 127.44, 126.83, 126.48, 123.45, 123.35, 121.82, 118.10, 116.32, 115.45, 99.98, 52.98, 35.46, 31.91, 25.28; LCMS:





Scheme 1 Acetic acid catalysed synthesis of probe (ADTI).

m/z calculated 473.1860 $[M]^+$, found 473.1889 $[M]^+$; elemental analysis: calculated C, 78.63; H, 5.32; N, 5.92; O, 10.14; found: C, 78.60; H, 5.31; N, 5.91; O, 10.12.

Results and discussion

Idea behind the design

The rationale of the present work centres on the development of novel colorimetric and fluorometric probe with the potential to selectively and sensitively detect cyanide ions in the presence of various competing ions. Among different available strategies, the development of D–A framework with high degree of conjugation is one of the most recognised and widely accepted within the scientific community. The strategy for designing novel D–A framework began by structural modulation of 12-acetylbenzo[*f*]pyrido[1,2-*a*]indole-6,11-dione (API) with a formyl group resulting into 12-acetyl-6,11-dioxo-6,11-dihydrobenzo[*f*]pyrido[1,2-*a*]indole-2-carbaldehyde (ADC). The purpose of introducing

formyl group is to extend the conjugation as well as to modulate the electrophilicity of the existing framework (API). The natural orbital analysis, performed using a DFT-based method, revealed the presence of two electrophilic centres (C_{16} and C_{23} atoms), with the higher electrophilicity of C_{23} over C_{16} being established through local reactivity descriptor (Fukui function). The high nucleophilicity of cyanide ion encouraged us to enhance the electrophilicity and regioselectivity of ADC. To achieve this, it was condensed with indolium unit to yield indolium based conjugated framework (ADTI). The condensation reaction resulted into a new electrophilic centre (C_{25}) with enhanced electrophilicity, as established using Fukui function (Fig. 1). Moreover, a reduction in energy of FMO's with decreased HOMO–LUMO gap (2.29 eV) was noticed. This decrease in HOMO–LUMO gap resulted into a bathochromic shift in the absorption maxima of the compound. Within conjugated frameworks, chemodosimetric (reaction based approach) and H-abstraction strategies are widely utilised in cyanide ion




Fig. 1 Structural modulation and localisation of local and global descriptors.




Fig. 2 ^1H NMR, ^{13}C NMR, COSY and mass spectra.

recognition process. The reaction-based recognition of the cyanide ion involves a nucleophilic attack at the electrophilic site, resulting in the disruption of the ICT process thus leading to an observable variation in the spectral properties.

Considering all these facts, the designed compound (**ADTI**) has been synthesized *via* a two-step process and molecular geometry was established using various spectral techniques such as ^1H , ^{13}C NMR and mass spectrometry. The overall composition of **ADTI** was confirmed by mass spectra. After that the structure characterisation was performed using ^1H and COSY NMR. The ^1H NMR spectra of the compound exhibits two doublet at δ_{H} 8.55 and 7.80 ppm due to two protons with coupling constant 16 Hz which revealed the trans coupling of these protons (Fig. 2). The three proton attached to quaternized nitrogen resonates at δ_{H} 4.24 ppm as a singlet and the acetyl protons appeared at δ_{H} 2.81 ppm as singlet. The aromatic protons of the ring lie in the range of δ_{H} 9.68–7.68 ppm. Moreover, the peak at 196.83 ppm in ^{13}C NMR spectra reflects the presence of carbonyl group in compound. The aromatic carbons resonate between 115–150 ppm and signals at δ_{C} 52.98, 35.46 and 31.91 ppm correspond to methyl groups.

Photophysical studies

Donor-acceptor (D- π -A) framework based chromophores exhibit significant changes in their fluorescence spectra with change in solvent polarity.⁵⁶ Due to its push-pull structure, the probe (**ADTI**) serve as a classical example of a D- π -A system. An increase in solvent polarity enhances the dipole-dipole interactions between the probe and solvent molecules, leading to a decrease in the energy of the excited state. The photophysical parameters of **ADTI** were recorded in solvents of different

polarities at ambient temperature (Table S1[†]). The absorption spectra of **ADTI** displayed absorption maxima ranging between 476–503 nm (ESI Fig. S1[†]) which was assigned to the intra-molecular charge transfer process. Similarly, its emission maxima were recorded from 572 to 641 nm in the same set of solvents (ESI Fig. S1[†]). Since light absorption is a very fast process, the absorption spectra of the probe were not affected significantly by solvent polarity.⁵⁷ On the other hand, solvent polarity does have a significant effect on the excited state, as chromophores interact more with solvent molecules, resulting in larger dipole moments and a lowering of excited state energy. A significant variation in the emission spectra, accompanied by bathochromic shift of 69 nm, was noticed with increasing solvent polarity. A detailed insight of the solvatochromism behaviour of the designed probe can be gained by applying Lippert-Mataga equation on Stokes shift and solvent polarity function (Δf). A positive linear correlation ($R^2 = 0.98$) with solvent polarity was observed (ESI Fig. S2[†]). It is noteworthy to mention here that the excited state of **ADTI** is stabilised by polar solvent which further established that its emission arises from the ICT state.

Naked eye recognition and sensing studies

The selectivity of receptor towards a particular anion in a mixture of various competing anions is an important feature. Hence, the selectivity of **ADTI** towards cyanide ion was evaluated by visually examining the aqueous solution of probe in the presence of different anions. A visual color variation from yellow to colourless was observed only with CN^- ion, thus revealing the colorimetric selectivity of the developed probe towards cyanide ion (Fig. 3A). Additionally, the color change was






Fig. 3 (A) Absorption spectra of the probe in the presence of different anions; (B) titration profile of ADTI with increasing concentration of cyanide ion (0–16 μM); (C) emission spectra of probe in the presence of various anions; (D) influence of cyanide concentration on the emission spectra of the probe; (E) LOD plot; and (F) Benesi–Hildebrand plot.

also observed under UV illumination (365 nm) (ESI Fig. S3†). The red color of the aqueous solution of ADTI disappeared with the addition of CN⁻ ion, thus leading to a “turn-off” response (Fig. 3C). These outcomes finally establish the high selectivity of ADTI towards CN⁻ ion. Colorimetric responses in the field of sensing are always associated with absorption modulation. In the present study, a colorimetric response further directed us to perform detailed absorption based studies of ADTI with different anions. The effect of the presence of different analytes on the absorption spectra of ADTI probe was investigated using water as a medium. The purpose of selecting water as the medium for further recognition studies is to enhance the applicability of the developed probe in environmental monitoring and bio-sensing. The absorption maxima (470 nm) of probe remained unchanged in the presence of different anions except cyanide (Fig. 3A). The addition of cyanide ion reduced the absorption intensity of ADTI. These findings clearly demonstrate the strong selectivity of ADTI towards CN⁻ ions and align well with the results of the naked-eye detection experiment. The mode of interaction in a “ADTI–CN⁻” adduct was further established using absorption based titration experiment. A gradual addition of cyanide causes a consistent decrease in the intensity of absorption maxima due to the disruption of conjugation, which diminishes the ICT process (Fig. 3B). This type of colorimetric sensor offers robust and cost effective detection of cyanide ions without involving any sophisticated instrument.

The scope of the study was further extended by utilising emission based studies in cyanide recognition. The fluorescence response of ADTI was investigated in the presence and absence of anions using aqueous medium. Anions other than cyanide didn't influence the fluorescence emission of ADTI (Fig. 3C). However, the presence of cyanide ion resulted into a decrease in emission intensity with hypsochromic shift. The behaviour was likely observed due to the addition of cyanide ions, which disrupts conjugation, thereby producing a “turn-off” response. A gradual decrease in the emission intensity of ADTI with the incremental addition of cyanide ions to the solution was attributed to a complexation reaction (Fig. 3D). This variation in the fluorescence spectra of probe indicates the effective disruption of conjugation due to nucleophilic addition of cyanide ion, which ultimately obstructs the intramolecular charge transfer transition. The limit of detection (LOD) for cyanide recognition was calculated to be 3.78 nM, which is below the level recommended by the WHO (Fig. 3E).⁴⁸ The binding constant calculated for this complexation reaction using titration studies was found to be $8.9 \times 10^4 \text{ M}^{-1}$ (Fig. 3F). On the basis of the values of binding constant and detection limit, it has been concluded that the probe has the capability to recognise cyanide ion in environmental samples.

Further, to determine the selectivity of ADTI, the response of probe towards cyanide ion in the presence of other competing ions was investigated (Fig. 4A). It has been observed that the cyanide recognition ability of ADTI remained unaffected even in




Fig. 4 (A) Interference studies in the presence of other ion; (B) effect of pH on ADTI and "ADTI-CN⁻" adduct (C) response time; (D) photostability of ADTI and "ADTI-CN⁻" adduct; (E) thermal stability of ADTI and "ADTI-CN⁻" adduct; (F) Job's plot.

the presence of other competing ions. These outcomes reflect the highly selective nature of the probe.

Various environmental factors, such as pH, photostability and temperature play a significant role, as they can interfere with the recognition process and impact the overall sensing performance of the probe. Therefore, the influence of pH on the recognition potential of ADTI was explored (Fig. 4B). The outcome of this study revealed that the developed sensor has the capability to detect the CN⁻ ion within the pH range of 3–7. These outcomes reflect the potential ability of the probe to work under optimum pH range. In the basic pH, hydroxide ion competes with cyanide ion. After successfully establishing the workability of probe in optimum pH range, the photo stability of the probe and "probe-CN⁻" adduct was also assessed by irradiating their solution with UV light (254 nm) (Fig. 4C). The outcomes of the experiment revealed that the probe and "probe-CN⁻" adduct didn't exhibits photo-bleaching upon exposure to UV-irradiation. Similarly, the thermal stability for both was also assessed. It has been observed that the emission characteristic of both probe and "probe-CN⁻" adduct remained unaltered even at temperature greater than 50 °C. These findings revealed the high thermal stability of the probe and probe-CN⁻ adduct (Fig. 4D).

One of the major limitations of the already reported cyanide responsive probes is their inability to respond quickly to a μM amount of cyanide present in the solution. The response time of the probe towards CN⁻ ion was also evaluated by performing a time dependent study. It has been found that the probe displayed quick response within 5 s when exposed to CN⁻ ion (Fig. 4E). Moreover, the solution color changed from yellow to

colourless immediately. Further, the stoichiometry of the probe to CN⁻ ion was determined using Job's continuous variation method (Fig. 4F). The obtained graph displayed the maximum value at mole fraction of 0.5, revealing a 1 : 1 binding stoichiometry of receptor to cyanide ion.

Plausible sensing mechanism

The binding stoichiometry of the probe with the cyanide ion was confirmed using Job's plot. To gain a deeper insight into the probable mechanism of interaction and to support the experimental evidence from photophysical studies, density functional theory (DFT)-based calculations were performed for the probe and the "probe-CN⁻" adduct. The ground state optimized geometry of ADTI revealed a planarity in the molecular structure thus facilitating the intramolecular charge transfer process. However, in the case of the probe-CN⁻ adduct, the molecule loses its planarity due to the disruption of conjugation upon the nucleophilic addition of the cyanide ion. Additionally, a significant redistribution of electron density was observed, which accompanies an increase in the HOMO–LUMO gap (Fig. 5A). The results of DFT studies are in strong corroboration with the experimental findings and supports the disruption of conjugation with the nucleophilic addition of cyanide ion.

Furthermore, to establish the nucleophilic addition of cyanide ion to C=N⁺ of the indolium moiety, ¹H NMR titration was conducted in DMSO *d*₆. In the ¹H NMR spectra of ADTI, the two vinyl protons resonate as doublets at δ_H 8.55 and 7.80 ppm with a coupling constant 16 Hz. Upon cyanide addition, an upfield shift is observed in the NMR signals of these protons.






Fig. 5 (A) DFT studies; (B) ¹H NMR titration studies of ADTI in the presence of cyanide ion in DMSO-*d*₆ and (C) plausible mode of interaction between ADTI and cyanide ion.

Moreover, the signal corresponding to the aromatic protons also shifts to the upfield region. Based on the literature studies, the probable mechanism of sensing can be depicted as shown in Fig. 5B. From these findings, it can be concluded that the nucleophilic attack of the cyanide ion on the C=N⁺ group of the indolium moiety results in the disruption of conjugation, thereby inhibiting the flow of electrons from the donor to the acceptor part, resulting into a “turn-off” fluorescence response (Fig. 5C). Such behavior is commonly associated with the quenching of absorption and emission intensities.

Real time application

The rapid growth of industrial activities, coupled with inadequate waste disposal practices, significantly contributes to environmental degradation. Cyanide is commonly found as a contaminant in wastewaters from various industries including metal cleaning, plating, electroplating, metal processing, automobile parts manufacture, steel tempering, mining, photography, pharmaceuticals, coal coking, ore leaching and plastics. Therefore, there is an urgent need to develop an effective analytical strategy for qualitative and quantitative estimation of

cyanide ion in environmental samples. In order to demonstrate the practical utility of ADTI in cyanide recognition, real sample analysis was performed using tap water, Ganga river water and drinking water as different environmental matrices. The water samples are spiked with known concentration of cyanide ion and analysed using fluorescence technique. As can be seen from Table 1, the probe has the potential to detect the CN⁻ ion with excellent % detection (93.3–98.3%). Moreover, the developed method has been validated with the previously reported AgNO₃ method.⁵⁸ These findings revealed highly sensitive nature of probe, indicating its reliability and feasibility in cyanide recognition process.

On site recognition studies

To determine the practical application of the developed colorimetric and fluorometric probe for on-site detection of CN⁻ ions, the probe was impregnated onto a filter paper-based test strip. This approach facilitates easy and robust recognition without utilising sophisticated instrumentation.^{54,55} Initially, the strips were dipped into an ADTI solution (10⁻³ M) for impregnation. The solvent was then removed by vacuum drying the strips for 24 h. The dried strips, when visualized under visible and UV light, exhibited red and orange colors, respectively (Fig. 6A). Then, the strips were exposed to the aqueous solution of various analytes (anions). A discernible color change from red to colorless under visible light, and from orange to blue under UV light, was observed only with cyanide exposure. Fig. S4† represents the pictorial representation of the cyanide detection test strip. These outcomes demonstrated the practical applicability of the developed probe for on-site cyanide detection using ADTI-impregnated paper strips, thus extending the overall scope of the probe to remote areas.

In addition to probe impregnated paper strips, smartphone based monitoring of immediate color change is also one of the

Table 1 Real water sample analysis

Samples	[CN ⁻] (μM)		Detection (%)
	Spiked	Detected	
River water (Ganga)	0	ND	
River water (Ganga)	3	2.80	93.3
Tap water	0	ND	
Tap water	3	2.95	98.3
Drinking water	0	ND	
Drinking water	3	2.89	96.3
AgNO ₃ method	3	2.88	96.0






Fig. 6 (A) Filter paper based test strip for cyanide detection under naked eye and UV-lamp; (B) smartphone based sensing of CN⁻ in aqueous solution of sensor ADTI; (C) INHIBIT logic gate.

cheap and robust method for on-site detection.⁵⁸ In the present study, the immediate color change of a **ADTI** solution upon cyanide addition has been utilised for the development of smartphone based recognition process. In this experiment, the **ADTI** solution (10^{-5} M) was mixed with five different known concentrations of CN⁻ in sample vials, and all the solutions were placed in containers lined up next to a blank sample. A photograph was captured using smartphone camera (Samsung Galaxy A32) positioned approximately 50 centimetres from a white background. The digital image analysis provided RGB (Red, Green, and Blue) data, with values ranging from 0 to 255 for each channel. For example, a completely black color would yield RGB values of 0, 0, 0, while pure white is associated with values of 255, 255, 255.

The "Colour Assist App" a free android application, was used to monitor RGB values in real samples, significantly reducing the analysis time as compared to other commonly used methods for determining RGB values (such as Image J, Matlab and Adobe Photoshop) (Fig. 6B). This app was used to analyze the glass vials containing only **ADTI** and those with varying known concentrations of CN⁻ ion. Relying solely on visual observation is not sufficient for detailed analysis, so a calibration curve was established with RGB values at different CN⁻ concentrations. The resulting plot displayed a linear relationship with an $R^2 = 0.99$ for the R/G ratio against [CN⁻]. The fabrication of sensors for the development of electronic devices is fundamentally based on logic gates, which serve as the core building blocks for processing and interpreting data. These sensors can be integrated into complex systems to enable

automation, control, and decision-making processes, making them crucial for advancements in the field of environmental recognition and bio-sensing.⁵⁹ In the present study, the sensing behaviour of **ADTI** towards cyanide ion was also analysed using logic gates. The "turn off" fluorescence response of **ADTI** can be demonstrated by a molecular "INHIBIT" logic gate, considering **ADTI** and cyanide ion as two inputs and signal at 641 nm as output. The truth table has been represented in ESI Table S2.† In this case, the emission intensity of **ADTI** was high in the absence of cyanide ions, corresponding to the "ON" state with a logic value of "1" (Fig. 6C). The addition of cyanide ions to the solution led to a decrease in emission intensity, demonstrating the "OFF" state, which corresponds to a logic value of "0".

Comparison of the developed probe with previously reported probe

The performance of the synthesized probe has been compared with already available cyanide sensor. The probe displayed a distinct visual color change and selectively exhibit a fluorogenic "turn-off" response to cyanide ion. The presence of competing anions has minimal or no impact on cyanide recognition process. The probe displayed an excellent detection limit as low as 3.78 nM. The aqueous phase and optimum pH workability are some of the promising features of the developed probe. The probe displayed rapid response within 5 s. Generally, two approaches have been utilised in the detection of cyanide ion such as H-abstraction and chemodosimetric approach. The chemodosimetric approach based chemosensor



exhibits higher selectivity and sensitivity for the cyanide ion as this approach involves chemical bond formation between the probe and cyanide ion which is highly reliable. Table S3[†] summarises the comparison of various analytical attributes of already reported and our developed probe. From the Table S3,[†] it has been concluded that the present probe displayed superior LOD than the already available probe.

Current challenges and future prospects

Cyanide sensing is a critical area of research due to the compound's high toxicity and widespread use in industrial processes, posing significant environmental and health hazards. Current challenges in cyanide sensing include achieving high selectivity and sensitivity in complex environmental matrices, developing portable and cost-effective sensors for real-time monitoring, and for ensuring robust performance under diverse environmental conditions. Additionally, addressing the interference caused by other anions, broad pH applicability, low photo-stability and improving the response time of detection methods remain key obstacles. However, future prospects in this field are promising, with advancements in nanotechnology, material science and fluorescence-based techniques paving the way for highly efficient sensing platforms. Innovations such as colorimetric sensors array, multi-functional materials and integration with Internet of Things (IoT) systems could revolutionize cyanide monitoring, enhancing environmental safety and public health protection.

Conclusion

In summary, a novel colorimetric and “turn off” fluorescent probe (**ADTI**) has been developed by structural modulations with extension of conjugation. The designed molecule displayed solvatochromism, excellent selectivity and sensitivity towards CN[−] ion over other anions in aqueous medium. The LOD and binding constant was found to be 3.78 nM and $8.9 \times 10^4 \text{ M}^{-1}$, respectively. The recognition mechanism involved the nucleophilic attack of cyanide ion and resulted into the disruption of ICT. The interaction mechanism has been established using DFT and ¹H NMR studies. The designed probe has the potential to detect cyanide ion in environmental samples. Additionally, test strip impregnated with **ADTI** demonstrates its practical utility in the on-site monitoring of cyanide ion. Smartphone application based strategy also displayed its potential in cyanide recognition without using sophisticated instrumentation.

Data availability

Data will be made available on request.

Author contributions

Kiran – methodology, investigation, formal analysis, data curation, visualisation, writing – original draft; Anju Ranolia –

writing – reviewing and editing; Priyanka – writing – reviewing and editing; Anil Duhan – reviewing and editing; Rahul Kumar Dhaka – reviewing and editing; Snigdha Singh – reviewing and editing; Gaurav Joshi – writing – reviewing and editing; Parvin Kumar – reviewing and editing; Devender Singh – reviewing and editing; Muhammad Wahajuddin – conceptualization, resources, reviewing and editing; Jayant Sindhu – supervision, software, conceptualization, project administration, writing – original draft.

Conflicts of interest

The authors declare that they have no known competing financial interests.

Acknowledgements

Kiran, Priyanka and Anju Ranolia would like to express their gratitude to UGC for the SRF and JRF, financial assistance, respectively (45/CSIR-UGC NET (JRF) Dec-2018). Jayant Sindhu acknowledge the Department of Chemistry, CCS Haryana Agricultural University, Hisar for providing lab facility including instrumental support. Snigdha Singh acknowledge the IoE-FRP (Ref. No./IoE/2024–25/12/FRP) University of Delhi for providing financial support.

References

- 1 V. Kavitha, P. Viswanathamurthi, J. Haribabu and C. Echeverria, A new nitrile vinyl linked ultrafast receptor to track cyanide ions: Utilization on realistic samples and HeLa cell imaging, *Spectrochim. Acta, Part A*, 2023, **295**, 122607, DOI: [10.1016/j.saa.2023.122607](https://doi.org/10.1016/j.saa.2023.122607).
- 2 Y. Q. Xie, M. M. Han, Y. M. Zhang, H. Chen, H. B. Zhang, C. Y. Ren, L. Li, R. Wu, H. Yao, X. N. Shi, Q. Lin and T. B. Wei, A novel fluorescent probe with high sensitivity for sequential detection of CN[−] and Al³⁺ in highly aqueous medium and its applications in living cell bioimaging, *J. Photochem. Photobiol., A*, 2023, **437**, 114488, DOI: [10.1016/j.jphotochem.2022.114488](https://doi.org/10.1016/j.jphotochem.2022.114488).
- 3 C. lu Zhang, J. hao Guo, Y. Zhang and C. Liu, A naphthalimide-based bifunctional fluorescent probe for detecting CN[−] and alkaline pH and its application in food samples and living cells imaging, *Microchem. J.*, 2024, **206**, 111594, DOI: [10.1016/j.microc.2024.111594](https://doi.org/10.1016/j.microc.2024.111594).
- 4 S. Izadi, J. Tashkhourian and S. Alireza Hosseini Hafshejani, Biosynthesized silver nanoparticles based on bitter orange bloom extract and its application for the determination of cyanide ion in environmental samples, *J. Photochem. Photobiol., A*, 2024, **446**, 115173, DOI: [10.1016/j.jphotochem.2023.115173](https://doi.org/10.1016/j.jphotochem.2023.115173).
- 5 A. Kalavathi, K. Satheeshkumar, V. Dharaniprabha, K. N. Vennila and K. P. Elango, Multi-spectroscopic and TD-DFT studies on chromogenic and fluorogenic detection of cyanide in an aqueous solution, *J. Fluoresc.*, 2024, **34**(6), 2691–2705, DOI: [10.1007/s10895-023-03473-z](https://doi.org/10.1007/s10895-023-03473-z).



- 6 H. Wang, Z. Liu, T. Xie, L. Wang, Q. Xu, M. Chen, Z. Zhang, Z. Ma, S. Fan and H. Wu, Electron-deficient moiety regulated structure: An efficient strategy for the design of a highly sensitive cyanide “turn-on” fluorescent probe, *Sens. Actuators, B*, 2023, **379**, 133218, DOI: [10.1016/j.snb.2022.133218](#).
- 7 H. Zavvar Mousavi, A. Mohammadi, S. Yaghoubi and N. Azizi Khereshki, Design and synthesis of a novel optical chemoreceptor based on naphthalene azo dye for detecting cyanide ion in aqueous medium and real samples, *Chem. Pap.*, 2024, **78**, 397–407, DOI: [10.1007/s11696-023-03097-4](#).
- 8 A. Kalavathi, K. Satheeshkumar, V. Dharaniprabha, K. N. Vennila and K. P. Elango, Spectroscopic and theoretical studies on the selective detection of cyanide ions by a turn-on fluorescent chemo-dosimeter and its application in living cell imaging, *J. Fluoresc.*, 2025, **35**(1), 165–178, DOI: [10.1007/s10895-023-03509-4](#).
- 9 L. E. Towill, J. S. Drury, B. L. Whitfield, E. B. Lewis, E. L. Galyan and A. S. Hammons, *Reviews of the environmental effects of pollutants- V: Cyanide*, US Environmental Protection Agency, Cincinnati, Ohio, 1978.
- 10 A. Tigreros, M. C. Ríos, C. Serrano-Sterling, D. Becerra, J. C. Castillo and J. Portilla, Synthesis and photophysical properties of 3-aryl-2-cyanoacrylamides: Design of a turn-on fluorescent probe for cyanide ion detection, *J. Mol. Struct.*, 2025, **1321**, 139926, DOI: [10.1016/j.molstruc.2024.139926](#).
- 11 P. Purushothaman, P. S. Umabharathi, S. G. Sureanthiran, V. Harish and S. Karpagam, A potent ratiometric detection and estimation of cyanide in the mainstream smoke of tobacco products by spectroscopic technique, *Inorg. Chem. Commun.*, 2023, **158**, 111564, DOI: [10.1016/j.inoche.2023.111564](#).
- 12 A. Kumar, E. Jeong, Y. Noh and P. S. Chae, Fluorescence-based ratiometric sensors as emerging tools for CN[−] detection: Chemical structures, sensing mechanisms and applications, *Methods*, 2024, **222**, 57–80, DOI: [10.1016/j.ymeth.2024.01.001](#).
- 13 J. Isaad and A. El Achari, Colorimetric probe for sequential chemosensing of mercury(II) and cyanide ions in aqueous media, based on a benzoxadiazole-pyrazolin-5-one glycoconjugate with INHIBIT logic gate response, *J. Mol. Struct.*, 2023, **1271**, 134036, DOI: [10.1016/j.molstruc.2022.134036](#).
- 14 S. Soonthonhut, K. Silpcharu, M. Sukwattanasinitt and P. Rashatasakhon, Comparative study on indolium-fluorophore conjugates as dual sensors for cyanide ion in aqueous media, *J. Photochem. Photobiol., A*, 2023, **445**, 115042, DOI: [10.1016/j.jphotochem.2023.115042](#).
- 15 E. Jaszczak-Wilke, Ż. Polkowska, M. Koprowski, K. Owsianik, A. E. Mitchell and P. Balczewski, Amygdalin: Toxicity, anticancer activity and analytical procedures for its determination in plant seeds, *Molecules*, 2021, **26**(8), 2253, DOI: [10.3390/molecules26082253](#).
- 16 S. Mahernia, A. Amanlou, G. Kiaee and M. Amanlou, Determination of hydrogen cyanide concentration in mainstream smoke of tobacco products by polarography, *J. Environ. Health Sci. Eng.*, 2015, **13**, 57, DOI: [10.1186/s40201-015-0211-1](#).
- 17 L. Yadav, P. Yadav, H. Laddha, M. Sharma, P. Sharma, M. Agarwal and R. Gupta, Smartphone based CN[−] ion sensing in electroplating wastewater and food samples in a simple, economical and rapid manner, *Inorg. Chem. Commun.*, 2024, **161**, 112083, DOI: [10.1016/j.inoche.2024.112083](#).
- 18 A. Maji, A. Biswas, A. Das, S. Gharami, K. Aich and T. K. Mondal, A novel carbazole-benzothiazole-based chemodosimeter for the chromogenic and fluorogenic recognition of CN[−], *New J. Chem.*, 2023, **47**, 11557–11564, DOI: [10.1039/d3nj01157g](#).
- 19 L. Wu, H. Xu, M. Shen, Y. Li, Q. Yang and Y. Li, An innovative fluorescent probe for the detection of cyanide - enhanced sensitivity by controlling its electrostatic potential and suitable for applications such as cell imaging and food analysis, *Spectrochim. Acta, Part A*, 2024, **305**, 123443, DOI: [10.1016/j.saa.2023.123443](#).
- 20 P. S. Umabharathi, P. Purushothaman, S. Karpagam, T. Dasgupta and R. Tamizhselvi, A one-pot synthesized chemodosimetric chemosensor for highly sensitive cyanide ion detection in aqueous media and from Tobacco smoke, *ChemistrySelect*, 2024, **9**, e202303893, DOI: [10.1002/slct.202303893](#).
- 21 J. Palanisamy, R. Rajagopal and A. Alfarhan, Selective and effective sensing of cyanide ion with no interference in water by phenothiazine-indolium fused optical sensor, *J. Fluoresc.*, 2024, DOI: [10.1007/s10895-024-03715-8](#).
- 22 R. Tamizhselvi, R. Bhaskar, S. K. Ashok Kumar, S. Mohandoss, Y. R. Lee and A. A. Napoleon, Functionalized 2-hydrazinobenzothiazole with bithiophene as a colorimetric sensor for lethal cyanide ions and its application in food samples, *ACS Omega*, 2024, **9**, 11223–11231, DOI: [10.1021/acsomega.3c06057](#).
- 23 R. Khalid, S. A. Shahzad, M. A. Assiri, T. Javid, H. Irshad and M. Z. Ullah, AIE active fluorescent probe for the detection of cyanide ion through fluorescence enhancement in real samples: An integrated extensive fluorescence and DFT studies, *Microchem. J.*, 2024, **200**, 110264, DOI: [10.1016/j.microc.2024.110264](#).
- 24 S. D. Padghan, J. W. Hu, Y. H. Hsu, K. M. Liu and K. Y. Chen, Swift and quantitative naked-eye detection of cyanide: A dual-mode approach using an ICT-based colorimetric and ratiometric fluorescent chemodosimeter, *Dyes Pigm.*, 2024, **228**, 112229, DOI: [10.1016/j.dyepig.2024.112229](#).
- 25 K. Magesh, N. Vijay, S. P. Wu and S. Velmathi, Dual-responsive benzo-hemicyanine-based fluorescent probe for detection of cyanide and hydrogen sulfide: Real-time application in identification of food spoilage, *J. Agric. Food Chem.*, 2023, **71**, 1190–1200, DOI: [10.1021/acs.jafc.2c05567](#).
- 26 A. Ranolia, Kiran, J. Sindhu, P. Kumar and S. Kumar, Divulging indolium inspired cyanide sensors: Did it win the throne?, *Coord. Chem. Rev.*, 2024, **498**, 215463, DOI: [10.1016/j.ccr.2023.215463](#).
- 27 R. Rajamanikandan, K. Shanmugaraj, M. Ilanchelian and H. Ju, Cysteamine-decorated gold nanoparticles for



- plasmon-based colorimetric on-site sensors for detecting cyanide ions using the smart-phone color ratio and for catalytic reduction of 4-nitrophenol, *Chemosphere*, 2023, **316**, 137836, DOI: [10.1016/j.chemosphere.2023.137836](https://doi.org/10.1016/j.chemosphere.2023.137836).
- 28 A. Biswas, S. Gharami, A. Maji, S. Guha, G. Das, R. Naskar and T. K. Mondal, A distinctive and proficient fluorescent switch for ratiometric recognition of the menacing cyanide ion: Biological studies on MDA-MB-231 cells, *Anal. Methods*, 2024, **16**(46), 8010–8018, DOI: [10.1039/D4AY01676A](https://doi.org/10.1039/D4AY01676A).
 - 29 A. Biswas, R. Mukherjee, A. Maji, R. Naskar, K. Aich, N. Murmu and T. K. Mondal, A triphenylamine scaffold for fluorogenic sensing of noxious cyanide via the ICT mechanism and its bioimaging application, *Sens. Diagn.*, 2024, **3**, 1201–1211, DOI: [10.1039/d4sd00018h](https://doi.org/10.1039/d4sd00018h).
 - 30 K. Satheeshkumar, P. Saravanakumar, A. Kalavathi, K. N. Vennila and K. P. Elango, Spectroscopic and TD-DFT studies on the chromo-fluorogenic detection of cyanide ions in organic and aquo-organic media, *Spectrochim. Acta, Part A*, 2023, **302**, 123054, DOI: [10.1016/j.saa.2023.123054](https://doi.org/10.1016/j.saa.2023.123054).
 - 31 R. Manivannan and Y. A. Son, Blue light emitting fluorophore for the effective detection of cyanide ion and electronic test kit development for real time measurement, *Dyes Pigm.*, 2023, **210**, 110941, DOI: [10.1016/j.dyepig.2022.110941](https://doi.org/10.1016/j.dyepig.2022.110941).
 - 32 S. Uota, B. J. Hwang, R. Butcher, R. Mullins, J. Wachira, Y. Hijji and F. Abebe, A simple benzothiazolium-based sensor for cyanide detection: Applications in environmental analysis and bioimaging, *Spectrochim. Acta, Part A*, 2025, **326**, 125155, DOI: [10.1016/j.saa.2024.125155](https://doi.org/10.1016/j.saa.2024.125155).
 - 33 M. S. Sunitha and S. Sarveswari, The development of a simple imidazole-based probe for the selective detection of cyanide ion in real samples, *Inorg. Chim. Acta*, 2024, **562**, 121893, DOI: [10.1016/j.ica.2023.121893](https://doi.org/10.1016/j.ica.2023.121893).
 - 34 K. Kumar Yadav, A. Raina and T. Ghosh, Cyanide selective colorimetric and turn-on fluorimetric sensing by naphthohydrazide derivative and its application in real sample analysis, *Mater. Today: Proc.*, 2023, **78**, 108–113, DOI: [10.1016/j.matpr.2022.11.309](https://doi.org/10.1016/j.matpr.2022.11.309).
 - 35 C. Immanuel David and H. il Lee, Cutting-edge advances in colorimetric and fluorescent chemosensors for detecting lethal cyanide ion: A comprehensive review, *Microchem. J.*, 2024, **200**, 110359, DOI: [10.1016/j.microc.2024.110359](https://doi.org/10.1016/j.microc.2024.110359).
 - 36 K. Satheeshkumar, P. Saravanakumar, A. Kalavathi, K. N. Vennila, S. Ciattini, L. Chelazzi and K. P. Elango, A highly selective probe for fluorometric sensing of cyanide in an aqueous solution and its application in quantitative determination and living cell imaging, *Methods*, 2023, **215**, 1–9, DOI: [10.1016/j.ymeth.2023.05.002](https://doi.org/10.1016/j.ymeth.2023.05.002).
 - 37 A. Maji, K. Aich, A. Biswas, S. Gharami, B. Bera and T. K. Mondal, Efficient solid- and solution-state emissive reusable solvatochromic fluorophores for colorimetric and fluorometric detection of CN[−], *Analyst*, 2024, **149**, 1557–1570, DOI: [10.1039/d3an01697h](https://doi.org/10.1039/d3an01697h).
 - 38 K. Jamuna, S. Gayathri, S. Sivakumar and B. Ashokkumar, Design and development of a fluorometric and colorimetric sensor for toxic cyanide detection by pyridinium scaffolds: Live cell imaging and real sample analysis, *Sens. Diagn.*, 2023, **2**, 337–346, DOI: [10.1039/d2sd00163b](https://doi.org/10.1039/d2sd00163b).
 - 39 I. Kathuria, P. Sharma and S. Kumar, Colorimetric, fluorometric, and electrochemical sensing of cyanide ion in aqueous media using merocyanine-ferrocene conjugate, *J. Photochem. Photobiol., A*, 2023, **445**, 115081, DOI: [10.1016/j.jphotochem.2023.115081](https://doi.org/10.1016/j.jphotochem.2023.115081).
 - 40 N. Azizi, T. Hallaj and N. Samadi, A dual-mode fluorometric and smartphone-based colorimetric sensor for cyanide detection using tungsten disulfide quantum dots and silver nanoparticles, *J. Food Compos. Anal.*, 2024, **129**, 106081, DOI: [10.1016/j.jfca.2024.106081](https://doi.org/10.1016/j.jfca.2024.106081).
 - 41 Q. Meng, C. Bai, J. Yao, X. Wang, S. He, X. Liu, S. Wang, W. Xue, L. Zhang, B. Wei, H. Miao, C. Qu and R. Qiao, Dual-mode “turn-on” fluorescence and colorimetric probe toward cyanide based on ESIPT and its bioimaging applications, *J. Lumin.*, 2023, **263**, 120031, DOI: [10.1016/j.jlumin.2023.120031](https://doi.org/10.1016/j.jlumin.2023.120031).
 - 42 J. Palanisamy, A. M. Al-Mohaimed and W. A. Al-onazi, A quick, highly selective and sensitive colorimetric and fluorimetric sensor for cyanide ion, *J. Fluoresc.*, 2024, DOI: [10.1007/s10895-024-03638-4](https://doi.org/10.1007/s10895-024-03638-4).
 - 43 H. Lim, P. Jayasudha, R. Manivannan and Y. A. Son, Development of red to blue emissive fluorescent materials for cyanide ion sensing and single-component inks in digital printing application, *J. Photochem. Photobiol., A*, 2025, **458**, 115969, DOI: [10.1016/j.jphotochem.2024.115969](https://doi.org/10.1016/j.jphotochem.2024.115969).
 - 44 M. A. Belal, S. Hajra, S. Panda, K. R. Kaja, M. M. M. Abdo, A. Abd El-Moneim, D. Janas, Y. K. Mishra and H. J. Kim, Advances in gas sensors using screen printing, *J. Mater. Chem. A*, 2025, **13**, 5447–5497, DOI: [10.1039/D4TA06632D](https://doi.org/10.1039/D4TA06632D).
 - 45 K. R. Kaja, S. Hajra, S. Panda, M. A. Belal, U. Pharino, H. Khanbareh, N. Vittayakorn, V. Vivekananthan, C. Bowen and H. J. Kim, Exploring liquid-solid interface based triboelectrification, structures, and applications, *Nano Energy*, 2024, **131**, 110319, DOI: [10.1016/j.nanoen.2024.110319](https://doi.org/10.1016/j.nanoen.2024.110319).
 - 46 U. Pharino, K. Chaithawee, S. Pongampai, N. Chanlek, S. Kothan, J. Kaewkhao, S. Hajra, H. J. Kim, W. Vittayakorn, S. Sriphan and N. Vittayakorn, A highly sensitive disease pre-screening approach for glycosuria: Triboelectric sensing at the liquid-solid interface, *Chem. Eng. J.*, 2025, **508**, 160901, DOI: [10.1016/j.cej.2025.160901](https://doi.org/10.1016/j.cej.2025.160901).
 - 47 J. H. Park, R. Manivannan, P. Jayasudha and Y. A. Son, Selective detection of cyanide ion in 100% water by indolium based dual reactive binding site optical sensor, *J. Photochem. Photobiol., A*, 2020, **397**, 112571, DOI: [10.1016/j.jphotochem.2020.112571](https://doi.org/10.1016/j.jphotochem.2020.112571).
 - 48 Y. Morikawa, M. Hirabara, K. Nishiwaki, S. Suzuki and I. Nakanishi, A novel turn-on fluorescent sensor for cyanide ions based on the charge transfer transition of phenothiazine/indolium compounds, *Mater. Adv.*, 2021, **2**, 6104–6111, DOI: [10.1039/d1ma00608h](https://doi.org/10.1039/d1ma00608h).
 - 49 T. P. Le Nguyen, C. H. T. Nguyen, T. Q. Nguyen, C. D. Tran, T. H. Nguyen, L. T. T. Nguyen and H. T. Nguyen, A novel colorimetric and fluorometric dual-channel chemosensor



- based on a conjugated perylene-benzothiazole system for highly selective detection of cyanide in aqueous media, *ChemPlusChem*, 2023, **88**, e202300264, DOI: [10.1002/cplu.202300264](https://doi.org/10.1002/cplu.202300264).
- 50 A. Battal, S. B. Kassa, N. Altinolcek Gultekin, M. Tavasli and Y. Onganer, A reaction-based carbazole-dicyanovinyl conjugated colorimetric and ratiometric fluorescent probe for selective detection of cyanide ions, *Spectrochim. Acta, Part A*, 2024, **304**, 123350, DOI: [10.1016/j.saa.2023.123350](https://doi.org/10.1016/j.saa.2023.123350).
- 51 M. Acar, A. Daştan and R. Koçak, Fluorometric and colorimetric sensor for selective detection of cyanide anion by dibenzosuberone-based dihydropyridazine in aqueous solution, *Talanta*, 2024, **277**, 126241, DOI: [10.1016/j.talanta.2024.126241](https://doi.org/10.1016/j.talanta.2024.126241).
- 52 M. Singh, S. Nadendla and R. Kumar Kanaparthi, A highly sensitive colorimetric and fluorometric sensor for the detection of cyanide, *J. Photochem. Photobiol., A*, 2025, **458**, 115957, DOI: [10.1016/j.jphotochem.2024.115957](https://doi.org/10.1016/j.jphotochem.2024.115957).
- 53 S. Manoj Kumar, D. Jothi, S. Munusamy, S. Enbanathan and S. Kulathu Iyer, Imidazole-derived new colorimetric/fluorometric chemosensor for the sensitive recognition of CN⁻ ions: Real-time application in food samples and fluorescence bio-imaging, *J. Photochem. Photobiol., A*, 2023, **434**, 114269, DOI: [10.1016/j.jphotochem.2022.114269](https://doi.org/10.1016/j.jphotochem.2022.114269).
- 54 Kiran, A. Ranolia, Priyanka, I. Bala, J. Jangir, S. Singh, J. Sindhu, P. Kumar and D. Singh, Rationally designed dual channel reversible probe for cyanide recognition in aqueous medium with solid-state sensing abilities, *J. Photochem. Photobiol., A*, 2024, **453**, 115650, DOI: [10.1016/j.jphotochem.2024.115650](https://doi.org/10.1016/j.jphotochem.2024.115650).
- 55 Kiran, A. Ranolia, P. Rani, G. Joshi, R. Kumar, S. Kumar, P. Kumar, S. Singh and J. Sindhu, Precisely designed NIR based conjugated framework as a solid state emitter for selective recognition of cyanide ion in solid state and cancer cells, *J. Photochem. Photobiol., A*, 2024, **449**, 115373, DOI: [10.1016/j.jphotochem.2023.115373](https://doi.org/10.1016/j.jphotochem.2023.115373).
- 56 A. Azaid, M. Alaqarbeh, T. Abram, M. Raftani, R. Kacimi, Y. Khaddam, A. Sbai, T. Lakhliifi and M. Bouachrine, D- π -A push-pull chromophores based on N,N-Diethylaniline as a donor for NLO applications: Effects of structural modification of π -linkers, *J. Mol. Struct.*, 2024, **1295**, 136602, DOI: [10.1016/j.molstruc.2023.136602](https://doi.org/10.1016/j.molstruc.2023.136602).
- 57 J. Basavaraja, S. R. Inamdar and H. M. Suresh Kumar, Solvents effect on the absorption and fluorescence spectra of 7-diethylamino-3-thenoylcoumarin: Evaluation and correlation between solvatochromism and solvent polarity parameters, *Spectrochim. Acta, Part A*, 2015, **137**, 527–534, DOI: [10.1016/j.saa.2014.08.118](https://doi.org/10.1016/j.saa.2014.08.118).
- 58 J. Jangir, Kiran, A. Ranolia, Priyanka, S. Chahal, S. Singh, A. Duhan, R. K. Dhaka, D. Singh, P. Kumar and J. Sindhu, Colorimetric reversibility: Aqueous phase recognition of cyanide using smart phone-based device with real sample analysis, *Microchem. J.*, 2025, **208**, 112259, DOI: [10.1016/j.microc.2024.112259](https://doi.org/10.1016/j.microc.2024.112259).
- 59 L. Liu, P. Liu, L. Ga and J. Ai, Advances in applications of molecular Logic Gates, *ACS Omega*, 2021, **6**, 30189–30204, DOI: [10.1021/acsomega.1c02912](https://doi.org/10.1021/acsomega.1c02912).

