



## Cross polarization from $^1\text{H}$ to quadrupolar $^6\text{Li}$ nuclei for dissolution DNP

Journal:	<i>Physical Chemistry Chemical Physics</i>
Manuscript ID:	CP-ART-08-2014-003592.R1
Article Type:	Paper
Date Submitted by the Author:	23-Sep-2014
Complete List of Authors:	Perez Linde, Angel Joaquin; EPFL, LRMB Bornet, Aurélien; ÉCOLE POLYTECHNIQUE FÉDÉRALE DE LAUSANNE, Institut des Sciences et Ingénierie Chimique Milani, Jonas; ÉCOLE POLYTECHNIQUE FÉDÉRALE DE LAUSANNE, Institut des Sciences et Ingénierie Chimique Vuichoud, Basile; ÉCOLE POLYTECHNIQUE FÉDÉRALE DE LAUSANNE, Institut des Sciences et Ingénierie Chimique Melzi, Roberto; Bruker Italia, Jannin, Sami; ÉCOLE POLYTECHNIQUE FÉDÉRALE DE LAUSANNE, Institut des Sciences et Ingénierie Chimique; Bruker Biospin AG, Industriestrasse 26 Bodenhausen, Geoffrey; EPFL, Chemistry; ENS, ; Sorbonne Universités, UPMC Univ, ; CNRS UMR ,

## ARTICLE

Cross polarization from  $^1\text{H}$  to quadrupolar  $^6\text{Li}$  nuclei for dissolution DNP

Cite this: DOI: 10.1039/x0xx00000x

Received .....,  
Accepted .....

DOI: 10.1039/x0xx00000x

[www.rsc.org/](http://www.rsc.org/)Angel J. Perez Linde<sup>1</sup>, Aurélien Bornet<sup>1</sup>, Jonas Milani<sup>1</sup>, Basile Vuichoud<sup>1</sup>, Roberto Melzi<sup>1,2</sup>, Sami Jannin<sup>1,3</sup> and Geoffrey Bodenhausen<sup>1,4,5,6</sup>

[1] Institut des Sciences et Ingénierie Chimiques (ISIC), Ecole Polytechnique Fédérale de Lausanne (EPFL), 1015 Lausanne, Switzerland

[2] Bruker Italia S.r.l., Viale V. Lancetti 43, 20158 Milano, Italy

[3] Bruker BioSpin AG, Industriestrasse 26, 8117 Fällanden, Switzerland

[4] Ecole Normale Supérieure-PSL Research University, Département de Chimie  
24 rue Lhomond, 75005 Paris, France[5] Sorbonne Universités, UPMC Univ. Paris 06, LBM, 4 place Jussieu, F-75005,  
Paris, France

[6] CNRS UMR 7203 LBM, F-75005, Paris, France

to be submitted to PCCP

Cross polarization from protons to quadrupolar  $^6\text{Li}$  nuclei is combined with dynamic nuclear polarization of protons at 1.2 K and 6.7 T using TEMPOL as polarizing agent and followed by rapid dissolution. Compared to direct  $^6\text{Li}$  DNP without cross-polarization, a higher nuclear spin polarization  $P(^6\text{Li})$  can be obtained in a shorter time. A double resonance  $^1\text{H}$ - $^6\text{Li}$  probe was designed that is equipped for Longitudinally Detected Electron Spin Resonance.

**Keywords:**

Hyperpolarization, Dissolution Dynamic Nuclear Polarization, Cross Polarization, Lithium-6, LODSR

**Introduction.** Dynamic Nuclear Polarization (DNP) combined with rapid dissolution (D-DNP) [1] usually focuses on nuclear spins like  $^{13}\text{C}$  or  $^{15}\text{N}$  that have long longitudinal relaxation times  $T_1$  in moderate magnetic fields  $B_0 > 1$  T. Rather surprisingly, some quadrupolar nuclei such as  $^6\text{Li}$  can also have sufficiently long relaxation times  $T_1(^6\text{Li})$  in solution [1, 2]. Hyperpolarization of  $^6\text{Li}$  could provide new insights in clinical studies, bearing in mind that lithium salts are used to treat bipolar and depressive disorders [3]. In particle physics, DNP has been used to enhance the  $^6\text{Li}$  polarization of targets [4]. Polarizations as high as  $P(^6\text{Li}) = 71\%$  have been obtained in small  $^6\text{LiD}$  crystals at very low temperatures (a few hundreds of mK), although the polarization build-up times  $\tau_{\text{DNP}}$  can be several hours in  $B_0 = 2.5$  T. High energy irradiation can be used

In this Communication, we show how cross-polarization (CP) can also be used to obtain a high polarization  $P(^6\text{Li})$  in a short time by using a suitable double resonance probe with two radio-frequency saddle coils in quadrature and an audio-frequency Helmholtz coil for Longitudinally Detected Electron Spin Resonance (LODESR).

DNP was originally predicted by Albert Overhauser and demonstrated on  $^7\text{Li}$  [11, 12]. Various mechanisms of the transfer of polarization can be considered in solids, both in static samples [13-22] and in samples undergoing magic angle spinning (MAS) [23, 24]. In DNP with 50 mM TEMPOL at 6.7 T and  $1.2 < T < 4.2$  K, a combination of three mechanisms appears to be involved: the solid




Figure 1. Pulse sequence used for cross-polarization (CP) in solids at low temperatures [5-7]. Both protons and  $^6\text{Li}$  nuclei are initially saturated by a series of  $90^\circ$  pulses (typically  $n = 20$ ) spaced by 10 ms. Two nested loops ( $q$  inside  $m$ ) are used, the inner loop to measure the polarization of  $^6\text{Li}$  with small  $5^\circ$  flip angles, and the outer loop (typically  $m = 18$  with intervals of 10 s for samples at 4.2 K or 30 s at 1.2 K) to establish cross polarization contacts by simultaneously applying pairs of frequency-modulated inversion pulses. The EPR transitions are saturated throughout the experiment to induce DNP.

to produce paramagnetic defects that act as polarizing agents in such crystals. It was also demonstrated [2] that  $^6\text{Li}$  could be polarized in a glassy matrix using the free radical TEMPOL as polarizing agent in D-DNP experiments in much shorter build-up time  $\tau_{\text{DNP}} = 7.5$  minutes at  $T = 1.2$  K and  $B_0 = 3.35$  T, although the polarization  $P(^6\text{Li}) = 7\%$  was much lower than reported in  $^6\text{LiD}$  crystals. For carbon-13 spins, we have shown that high polarization levels can be obtained rapidly by first polarizing protons and using cross polarization (CP) in the manner of Hartmann and Hahn to transfer the polarization from protons to carbon-13 [5-10]. At  $T = 1.2$  K and  $B_0 = 6.7$  T we obtained  $P(^{13}\text{C}) = 71\%$  in 20 min with a horizontal solenoidal coil that precludes rapid dissolution [6], and  $P(^{13}\text{C}) = 45\%$  with vertical coils that are compatible with rapid dissolution [10].

effect, the cross effect and thermal mixing [25]. The combination of CP [26] with DNP can enhance the hyperpolarization of low gamma nuclei [27], including some quadrupolar nuclei [28].

Adiabatic cross-polarization (ACP) [29, 30] is highly suitable for D-DNP experiments since it suffices to use low  $rf$  amplitudes to minimize arcing [5-10]. Like other CP schemes [31-33], the final nuclear spin polarization must be aligned with the magnetic field prior to dissolution. CP contacts can be achieved by applying simultaneous pairs of adiabatic inversion pulses using frequency sweeps with so-called 'Wurst' amplitude profiles [34]. The magnetization vectors follow the effective fields so that they end up aligned with the magnetic field. After an even number of contacts, the magnetization vectors are parallel rather than antiparallel with respect to the magnetic field. Pairs of pulses can be

applied periodically to achieve a cumulative enhancement of the  ${}^6\text{Li}$  polarization (Figure 1).

**Experimental.** The experiments were performed using a home-built polarizer operating at  $B_0 = 6.7\text{ T}$  and  $T = 1.2\text{ K}$  [8]. A dedicated probe was built with a 188 GHz wave-guide and an *rf*

The probe comprises two orthogonal saddle coils made of copper wire and immersed in liquid helium (see (1) in Figure 2b.) The 6 mm inner diameter saddle coil for  ${}^6\text{Li}$  can be tuned and partly matched using a cryogenic trimmer capacitor with an extended shaft (2), in combination with American Technical Ceramics (ATC) chip capacitors. The trimmer capacitor shaft can be rotated from the




Figure 2. (a) Upper and (b) bottom parts of the probe. The LODESR coils (9) with the coil former (6), nut (10) and microwave cavity (20) are located below the NMR coils. (c) Circuit where the  ${}^1\text{H}$  side (shown in red) comprises a matching and two trimming capacitors for tuning (21). The  ${}^6\text{Li}$  part (shown in black) uses a Voltronics trimmer capacitor (2) in parallel with a chip capacitor suitable for low temperatures. Partial matching is obtained with a capacitor in series.

circuit for cross polarization from  ${}^1\text{H}$  to  ${}^6\text{Li}$  that is compatible with rapid dissolution.

outside with a rod (3) passing through a brass nut with a vacuum-tight O-ring (4). The nut was tightened at room temperature once the circuit was tuned. The  ${}^6\text{Li}$  channel has a bandwidth of *ca.* 10 MHz, centred on the  ${}^6\text{Li}$  resonance frequency of 41.97 MHz. The  ${}^6\text{Li}$  coil

presents two turns per wing of 0.4 mm diameter enamelled copper wire placed around a coil former of 6 mm diameter made of PTFE (5). The saddle coil for protons presents the same inner diameter and a length of 12 mm with one turn per wing made of 0.22 mm diameter enamelled copper wire. Partial matching was achieved with a 10 pF capacitor in series with this coil. The PTFE coil former (6) supports a coil for longitudinally detected electron spin resonance (LODESR) [35-37]. Lateral holes on this coil former (7) allowed the waveguide (8) to reach the central region where the NMR coils are placed (1). The Helmholtz coil (9) for LODESR [35] was made of 400 turns per wing of 0.10 mm diameter enamelled copper wire, placed on the LODESR coil former, and attached by means of a PTFE nut (10) to the NMR coil former. A brass platform (11) with a

nut (23). Vacuum was made possible by compressing the waveguide, cables and rod with a nut against a flange (24) with O-rings. The sample was loaded from the top through the Vetronite tube [8]. The sample holder, which has a cylindrical symmetry and is made of PEEK, has inner and outer diameters of 3 and 5 mm respectively.

The 6 mm thin-walled waveguide (8) made of non-magnetic steel was connected to a mirror (25) to reflect the microwaves generated by a frequency doubler (100 mW at 188 GHz) connected to an Elva source (400 mW at 94 GHz). The orthogonal coils have good *rf* homogeneity [40]. The probe design presented here is distinct from other probes that have been developed for cross polarization at low temperatures, either without [41, 42] or with




Figure 3. (a) Experimental build-up of  $P(^6\text{Li})$  observed by direct DNP without cross polarization at 6.7 T and 4.2 K (red) with mono-exponential fit (black) and experimental build-up of  $P(^6\text{Li})$  by indirect DNP combined with CP from  $^1\text{H}$  to  $^6\text{Li}$  (blue). (b) Decay of  $^6\text{Li}$  polarization to thermal equilibrium at 4.2 K in absence of irradiation (blue) with mono-exponential fit (black). The sample consisted of 4 beads of 10  $\mu\text{l}$  each of 7 M  $^6\text{Li}$ -enriched lithium chloride LiCl with 50 mM TEMPO in 40:40:20 (v:v:v) ethanol- $\text{d}_6$ : $\text{D}_2\text{O}$ : $\text{H}_2\text{O}$  with an extra bead of 10  $\mu\text{l}$  of 3 M ascorbic acid in  $\text{H}_2\text{O}$  to quench the radicals during dissolution.

central hole (12) to accept the sample holder (13) and outer holes for the waveguide (8), wires (14), and holding screws (15) was used to support the NMR coil former. Spacers (16) were placed at regular distances along the vertical axis of the probe, to maintain the structure, to receive the Vetronite tube (17), cables (18, 19) and the waveguide, and to allow the tuning rod to rotate freely. The NMR, and LODESR coils were placed in a brass cavity to confine the microwave irradiation (20). The sample volume is 50  $\mu\text{l}$ .

The tuning and matching box (21) was connected to the room-temperature end of the  $^1\text{H}$  cable (19) which is part of the tuning and matching system [38, 39]. The cable for  $^6\text{Li}$  (18) was connected directly to the filters before the preamplifier of a Bruker Avance II spectrometer. The isolation between the channels was -23 dB from  $^1\text{H}$  to  $^6\text{Li}$ , and -50 dB from  $^6\text{Li}$  to  $^1\text{H}$ . The flange at the top of the probe (22) was made of brass and held to the cryostat with a

rapid dissolution [5, 7, 9, 10].

Microwave frequency modulation with a sweep range of 50 MHz and a modulation frequency of 10 kHz was used to improve the nuclear polarization and shorten the DNP build-up times of proton [43, 44]. The centre of the frequency modulation was set to 188.270 GHz to obtain the maximum (negative) polarization  $P(^1\text{H})$ .

The adiabatic  $180^\circ$  Wurst inversion pulses [7, 10, 34] had bandwidths of 100 and 30 kHz for  $^1\text{H}$  and  $^6\text{Li}$ , respectively. The *rf* amplitude of the proton channel was set just below arcing (9 W, corresponding to an *rf* field of 10 kHz.) The *rf* amplitude of the  $^6\text{Li}$  channel was then optimized (14 W, *rf* field 10 kHz.) The lengths of both inversion pulses were 1.45 ms.

Short  $5^\circ$  pulses were used to measure the polarizations  $P(^1\text{H})$  and  $P(^6\text{Li})$  at low temperatures without causing significant losses. The DNP build-up times  $\tau_{\text{DNP}}(^1\text{H})$  and  $\tau_{\text{DNP}}(^6\text{Li})$  were measured at 4.2 and 1.2 K. The decay of the polarization  $P(^6\text{Li})$  towards thermal equilibrium due to  $T_1(^6\text{Li})$  relaxation was measured at 4.2 K and at room temperature after dissolution.

The sample consisted of 4 beads of 10  $\mu\text{l}$  each of 7 M lithium chloride LiCl enriched in  $^6\text{Li}$  with 50 mM TEMPO in 40:40:20 (v:v:v) ethanol- $\text{d}_6$ : $\text{D}_2\text{O}$ : $\text{H}_2\text{O}$ , with an extra 10  $\mu\text{l}$  bead of 3 M ascorbate to quench the radicals during and after dissolution [45].

600 s (10 minutes.) Dissolution does not cause significant losses of polarization  $P(^6\text{Li})$ .

When the temperature was lowered to 1.2 K the build up was significantly slowed down to  $\tau_{\text{DNP}}(^1\text{H}) = 215$  s and  $\tau_{\text{DNP}}(^6\text{Li}) = 2021$  s. Figure 4 (a) shows the experimental DNP-CP build up curve of  $P(^6\text{Li})$ . In this case the interval between CP contacts was 450 s. The polarization  $P(^6\text{Li})$  obtained with DNP and CP was again 1.7 times higher than could be obtained by direct DNP. Comparing  $^6\text{Li}$  signal intensities at thermal equilibrium at 4.2 K (with 32 pulses with  $5^\circ$  nutation angles each) with the maximum polarization obtained at 1.2 K, a maximum polarization  $P(^6\text{Li}) = 20 \pm 2\%$  was




Figure 4. (a) Experimental build-up of  $P(^6\text{Li})$  obtained with cross polarization (blue), with DNP only (red) and fit of the latter curve (black) at 6.7 T and 1.2 K. Since the proton build-up time constant was  $\tau_{\text{DNP}}(^1\text{H}) = 3.58$  min, intervals of 7.5 min between subsequent CP contacts were used as a compromise between the build-up of  $P(^1\text{H})$  and the decay of  $P(^6\text{Li})$ . After each CP contact, the polarization  $P(^6\text{Li})$  tends to return to the plateau observed by direct  $^6\text{Li}$  DNP. (b) Detail of the decay of  $P(^6\text{Li})$  in solution at 7 T and 300 K (blue), fitted to a mono-exponential function (black). The sample consisted of 4 beads of 10  $\mu\text{l}$  each of 7 M  $^6\text{Li}$ -enriched lithium chloride LiCl with 50 mM TEMPO in 40:40:20 (v:v:v) ethanol- $\text{d}_6$ : $\text{D}_2\text{O}$ : $\text{H}_2\text{O}$  with an extra bead of 10  $\mu\text{l}$  of 3 M ascorbic acid in  $\text{H}_2\text{O}$  to quench the radicals during dissolution. (c) Detail of two  $^6\text{Li}$  spectra taken at room temperature with a single flip angle of  $5^\circ$  (blue) and gain 16 and a single  $90^\circ$  flip angle and gain 1024. Spectra are scaled to fit in the graph.

**Polarisation in the solid.** The  $90^\circ$  pulse lengths were 7  $\mu\text{s}$  for protons using 100 W and 6.75  $\mu\text{s}$  for  $^6\text{Li}$  using 100 W. The build-up and relaxation times at 4.2 K were  $\tau_{\text{DNP}}(^1\text{H}) = 83$  s,  $\tau_{\text{DNP}}(^6\text{Li}) = 543$  s and  $T_1(^6\text{Li}) = 768$  s. In Figure 3(a) the build-up curves of  $P(^6\text{Li})$  are compared for direct DNP and indirect CP-DNP, both at 4.2 K. A single CP contact established 180 s after pre-saturation provided a larger spin polarization  $P(^6\text{Li})$  than could be achieved by direct DNP. A compromise delay of 180 s was chosen to allow the proton polarization to build up again before the next CP contact. After the third contact, the polarization  $P(^6\text{Li})$  was 1.7 times higher than could be obtained with direct DNP. A train of 32 short  $5^\circ$  pulses were applied to measure thermal equilibrium with a sufficient signal-to-noise ratio at 4.2 K. Afterwards, a pair of adiabatic inversion pulses were applied to establish a CP contact, and 32 pulses were applied to determine  $P(^6\text{Li})$ . At 4.2 K,  $T_1(^6\text{Li}) >$

measured. The efficiency of adiabatic CP at 1.2 K is approximately 24%. A proton polarization  $P(^1\text{H})$  76 % was reached at 1.2 K.

**Polarisation in solution.** Once the sample achieved its maximum polarization  $P(^6\text{Li})$  at 1.2 K, it was dissolved using 5 ml of overheated  $\text{D}_2\text{O}$  at 393 K at 1 MPa. The transfer time was 7 s using a magnetic tunnel which extended from the polarizer to a 7 T Bruker (300 MHz) spectrometer. Measurements were started 60 s after dissolution. With short  $5^\circ$  pulses applied every 10 s during the decay of the polarization  $P(^6\text{Li})$ , a  $T_1(^6\text{Li}) = 161$  s was measured in solution (Figure 4 b). By comparing the first spectrum taken after dissolution (one  $5^\circ$  pulse) with a spectrum taken in thermal equilibrium (64 scans with  $90^\circ$  pulses) at 7 T, the measured enhancement was  $\epsilon = 22000$  with respect to thermal equilibrium. The final concentration of LiCl was 56 mM.



**ESR response.** There are obvious advantages in monitoring the EPR response under the same conditions where DNP and CP are conducted [35-37, 46]. The two wires from the LODESER coil passed through a vacuum-sealed feed-through at the top of the probe, and were connected to a Stanford Research Systems low noise preamplifier (Model SR560). The presence of a Helmholtz coil surrounding the NMR coils did not significantly degrade the rf fields. A homemade Labview program controlled a NI USB-6211 DAQ, which was connected to the Elva microwave source for signal generation and also to the preamplifier for data acquisition [35-37]. An example of LODESER spectrum is given in Figure 5.




Figure 5. Longitudinally detected electron spin resonance (LODESER) spectrum of a sample of 12 mM DPPH in toluene, taken at 4.2 K. A TEMPOL spectrum obtained with this probe is shown in ref [34].

**Conclusions.** A high nuclear spin polarization  $P(^6\text{Li})$  was obtained rapidly by combining proton DNP using adiabatic cross-polarization (ACP) from  $^1\text{H}$  to  $^6\text{Li}$ . The proton polarization prior to CP was estimated to be  $P(^1\text{H}) = 76\%$  at 6.7 T and 1.2 K. A maximum polarization  $P(^6\text{Li}) = 20 \pm 2\%$  was obtained at 1.2 K. Although the efficiency of CP was not ideal (a gain of 1.7 on  $P(^6\text{Li})$  was obtained instead of the maximum which for  $^1\text{H}$ - $^6\text{Li}$  CP is 7, leaving an efficiency of  $1.7/7 \times 100 = 24\%$ ), it was possible with a single CP contact to exceed the plateau that can be achieved by direct  $^6\text{Li}$  DNP without CP, at both 4.2 and 1.2 K in a few minutes. With direct DNP using microwave frequency modulation but no CP, a polarization  $P(^6\text{Li}) = 11 \pm 1\%$  was achieved and 1.2 K. The relatively short  $T_1(^6\text{Li})$  relaxation time in solution remains a limitation for clinical applications. A double resonance NMR probe has been constructed for CP at low temperatures using two NMR saddle coils in quadrature, suitable for CP, despite some rf inhomogeneity. The probe presented an rf tuning bandwidth of 10 MHz on the  $^6\text{Li}$  channel, with sufficient rf field strengths for efficient CP, i.e., 10 kHz on each channel for several ms without arcing. The cross polarization sequence using double inversion was

found to be efficient for the quadrupolar  $^6\text{Li}$  nuclei. The choice of two coils in quadrature was a compromise to save space for some other features like LODESER. It has been demonstrated that despite the presence of a LODESER Helmholtz coil, an excellent proton polarization of 76 % could be achieved.

## Acknowledgements

The authors would like to thank Martial Rey, Bruker Biospin and the staff of the EPFL workshops André Fattet, Roger Mottier and Frédéric Gummy for assistance. This work was supported by the Swiss National Science Foundation (SNSF), the Ecole Polytechnique Fédérale de Lausanne (EPFL), the Swiss Commission for Technology and Innovation (CTI), the French CNRS, and the European Research Council (ERC) Advanced Grant 339754 "Dilute Para-Water".

## References

- 1 J. H. Ardenkjaer-Larsen, B. Fridlund, A. Gram, G. Hansson, M. Lerche, R. Servin, M. Thaning, K. Golman, *Proc. Natl. Acad. Sc. USA*, 2003, **100**, 10158
- 2 R. B. van Heeswijk, K. Uffmann, A. Comment, F. Kurdzesau, C. Perazzolo, C. Cudalbu, S. Jannin, J.A. Konter, P. Hautle, B. van den Brandt, G. Navon, J. J. van der Klink, R. Grueter, *Magn.. Reson. Med.*, 2009, **61**, 1489
- 3 J. R. Geddes, S. Burgess, K. Hawton, K. Jamison, G. M. Goodwin, *Am J. Psychiatry*, 2004, **161**, 217
- 4 V. Bouffard, Y. Roinel, P. Roubeau, A. Abragam, *J. Physique*, 1980, **41**, 1447
- 5 A.J. Perez Linde, Nottingham: Nottingham e-theses, 2010
- 6 S. Jannin, A. Bornet, S. Colombo, G. Bodenhausen, *Chem. Phys. Lett.* 2011, **517**, 234
- 7 A. Bornet, R. Melzi, S. Jannin, G. Bodenhausen, *Appl. Magn. Reson.*, 2012, **43**, 107
- 8 S. Jannin, A. Bornet, R. Melzi, G. Bodenhausen, *Chem.. Phys. Lett.*, 2012, **549**, 99
- 9 M. Batel, M. Krajewski, A. Däpp, A. Hunkeler, B. H. Meier, S. Kozerke, M. Ernst, *Chem. Phys. Lett.*, 2012, **554**, 72
- 10 A. Bornet, R. Melzi, A. J. Perez Linde, P. Hautle, B. Van den Brant, S. Jannin, G. Bodenhausen, *J. Phys.Chem. Lett.* 2013, **4**, 1117
- 11 A. Overhauser, *Phys. Rev.*, 1953, **92**, 411
- 12 T. R. Carver, C. P. Slichter, *Phys. Rev.* 1953, **92**, 212
- 13 W. Th. Wenckebach, T. J. B. Swaneburg, N. J. Poullis, *Phys. Rep.*, 1974, **14**, 181
- 14 A. Abragam, M. Goldman, *Rep. Prog. Phys.*, 1978, **41**
- 15 M. Borghini, *Phys. Rev. Lett.*, 1968, **20**, 419.
- 16 B. N. Provotorov, *Sov. Phys. JETP*, 1962, **14**, 1126.

- 17 V. A. Atsarkin, *Usp. Fiz. Nauk*, 1978, **126**, 3
- 18 K. N. Hu, G. T. Debelouchina, A. A. Smith, R. Griffin, *J. Chem. Phys.*, 2011, **134**, 125105
- 19 A. Karabanov, A. van der Drift, L. J. Edwards, I. Kuprov, W. Köckenberger, *Phys. Chem. Chem. Phys.*, 2012, **14**, 2658
- 20 S. Colombo Serra, A. Rosso, F. Tedoldi, *Phys. Chem. Chem. Phys.*, **15**, 2013, 8416
- 21 S. Jannin, A. Comment, J. J. van der Klink, *Appl. Magn. Reson.*, 2012, **43**, 59
- 22 B. Corcilius, A. Smith, R. Griffin, *J. Chem. Phys.*, 2012, 054201,
- 23 F. Mentink-Vigier, U. Akbey, Y. Hovav, S. Vega, H. Oschkinat, A. Feintuch, *J. Magn. Reson.*, 2012, **224**, 13
- 24 K. Thurber, R. Tycko, *J. Chem. Phys.*, 2012, **137**, 084508
- 25 Y. Hovav, O. Levinkron, A. Feintuch, S. Vega, *Appl. Magn. Reson.* 2012, **43**, 21
- 26 S. R. Hartmann, E. L. Hahn, *Phys. Rev.*, 1962, **128**, 2042.
- 27 V. S. Bajaj, C. T. Farrar, M. K. Hornstein, I. Mastovsky, J. Vieregg, J. Bryant, B. Eléna, K. E. Kreischer, R. J. Temkin, R. G. Griffin, *J. Magn. Reson.*, 2003, **160**, 85
- 28 V. Vitzhum, M. A. Caporini, G. Bodenhausen, *J. Magn. Reson.*, 2010, **205**, 177
- 29 K. J. Harris, A. Lupulescu, B. E. G. Lucier, L. Frydman, R. Shurko, *J. Magn. Reson.*, 2012, **224**, 38
- 30 S. Hediger, B. H. Meier, N. D. Kurur, G. Bodenhausen, R. R. Ernst, *Chem. Phys. Lett.*, 1994, **223**, 283
- 31 M. Fukuchi, A. Ramamoorthy, K. Takegoshi, *J. Magn. Reson.*, 2009, **196**, 105
- 32 P. Hodgkinson, C. Auger, L. Emsley, *J. Chem. Phys.*, 1998, **109**, 5
- 33 D. Rovnyak, *Concepts Magn. Reson.*, 2008, **32 A**, 4, 254
- 34 E. Kupce, R. Freeman, *J. Magn. Reson.* 1995, **115**, 273
- 35 J. Granwehr, J. Leggett, W. Köckenberger, *J. Magn. Reson.*, 2007, **187**, 266
- 36 M. Batel, M. Krajewski, K. Weiss, O. Width, A. Däpp, A. Hunkeler, M. Gimersky, K. P. Pruessmann, P. Boesiger, B. H. Meier, S. Kozerke, M. Ernst, *J. Magn. Reson.*, 2012, **214**, 166
- 37 J. Leggett, R. Hunter, J. Granwehr, R. Panek, A. J. Perez Linde, J. Horsewill, J. McMaster, G. Smith, W. Köckenberger, *Phys. Chem. Chem. Phys.*, 2010, **12**, 5883
- 38 V. D. Kodibaqkar, M. S. Conradi, *J. Magn. Reson.*, 2000, **144**, 53
- 39 A. Rath, *Magn. Reson. Med.*, 1990, **13**, 370-377
- 40 F. D. Doty, J. Kulkarni, C. Turner, G. Entzminger, A. Bielecki, J. Magn. Reson. 2006, **182**, 239
- 41 H. Cho, J. Baugh, C. A. Ryan, D. G. Cory, C. Ramanathan, *J. Magn. Reson.*, 2007, **187**, 242
- 42 C. Rienstra, Cambridge, M. A. MIT, PhD thesis, 1999
- 43 Y. Hovav, A. Feintuch, S. Vega, D. Goldfarb, *J. Magn. Reson.*, 2014, **238**, 94
- 44 A. Bornet, J. Milani, B. Vuichoud, A. J. Perez Linde, G. Bodenhausen, S. Jannin, *Chem. Phys. Lett.*, 2014, **602**, 63-67
- 45 P. Miéville, P. Ahuja, R. Sarkar, S. Jannin, P. Vasos, S. Gerber-Lemaire, M. Mishkovsky, A. Comment, R. Gruetter, O. Ouari, P. Tordo, G. Bodenhausen, *Angew. Chem. Int. Ed. Eng.*, 2010, **49**, 6182
- 46 T.R. Eichhorn, M. Haag, B. van den Brandt, P. Hautle, W.Th. Wenckebach, S. Jannin, J.J. van der Klink, A. Comment, *J. Magn. Reson.* 2013, **234**, 58