



Analyst

## A Computer Vision Enhanced Smart Phone Platform for Microfluidic Urine Glucometry

Journal:	<i>Analyst</i>
Manuscript ID	AN-ART-08-2023-001356.R1
Article Type:	Paper
Date Submitted by the Author:	25-Oct-2023
Complete List of Authors:	<p>Meng, Zhuolun; Rutgers University Department of Electrical and Computer Engineering</p> <p>Tayyab, Muhammad; Rutgers The State University of New Jersey, Electrical and Computer Engineering</p> <p>Lin, Zhongtian; Rutgers University, Electrical and Computer Engineering</p> <p>Raji, Hassan; Rutgers University Department of Electrical and Computer Engineering</p> <p>Javanmard, Mehdi; Rutgers University Department of Electrical and Computer Engineering</p>

SCHOLARONE™  
Manuscripts

Cite this: DOI: 00.0000/xxxxxxxxxx

A Computer Vision Enhanced Smart Phone Platform for Microfluidic Urine Glucometry

Zhuolun Meng<sup>a</sup>, Muhammad Tayyab<sup>a</sup>, Zhongtian Lin<sup>a</sup>, Hassan Raji<sup>a</sup>, and Mehdi Javanmard<sup>a</sup>

Received Date  
Accepted Date

DOI: 00.0000/xxxxxxxxxx

Glucose is an important biomarker for diagnosing and prognosing various diseases, including diabetes and hypoglycemia, which can have severe side effects, symptoms, and even death in patients. As a result, there is a need for quick and economical glucose level measurement to help identify those at potential risk. With the increase in smartphone users, portable smartphone glucose sensors are becoming popular. In this paper, we present a disposable microfluidic glucose sensor that accurately and rapidly quantifies glucose levels in human urine using a combination of colorimetric analysis and computer vision. This glucose sensor implements a disposable microfluidic device based on medical-grade tapes and glucose analysis strips on a glass slide integrated with a custom-made PolyDimethylSiloxane (PDMS) micropump that accelerates capillary flow, making it economical, convenient, rapid, and equipment-free. After absorbing the target solution, the disposable device is slid into the 3D-printed main chassis and illuminated with exclusive Light Emitting Diode (LED) illumination, which is pivotal to color-sensitive experiments. After collecting images, the images are imported into the algorithm to measure the glucose levels using computer vision and average RGB values measurement. This article illustrates the impressive accuracy and consistency of the glucose sensor in quantifying glucose in sucrose water. This is evidenced by the close agreement between the computer vision method used by the sensor and the traditional method of measuring method in the biology field, as well as the small variation observed between different sensor performances. The exponential regression curve used in the study further confirms the strong relationship between glucose concentrations and average RGB values, with an R-square value of 0.997 indicating a high degree of correlation between these variables. The article also emphasized the potential transferability of the solution described to other types of assays and smartphone-based sensors.

1 Introduction

Glucose, also known as blood sugar, is a crucial component of the human body and is the primary source of energy for cells. The blood carries glucose to all the human body's cells for energy. When glucose levels become abnormal, they can lead to various diseases, including diabetes. This condition occurs when blood sugar levels are too high and can cause serious long-term health problems<sup>1,2</sup>. According to the Centers for Disease Control and Prevention, there are currently 37.3 million diabetes patients in the US, with an additional 96 million people aged 18 years or older having prediabetes which takes 38% of the adult US population<sup>3</sup>. Treating diabetes is costly, with an estimated total cost of 327 billion USD in the US in 2017, due to the many health problems associated with high blood glucose, such as heart disease, stroke, kidney disease, and eye problems<sup>4</sup>. Even individuals without diabetes may experience issues with blood sugar that is

too low or high. Maintaining a regular schedule of eating, physical activity, and taking medicines can help to manage blood sugar levels which in a normal range for glucose vary generally from 70 to 99 mg/dL. The cost of glucose tests is essential for individuals at risk, and with the increasing popularity, convenience, and user-friendly nature of smartphones, they are becoming an essential tool for monitoring and diagnosing glucose levels in healthcare.

In the last few decades, biomedical devices and sensors have been rapidly developing to meet the requirements and challenges of modern society. These devices and sensors can utilize small volumes of fluids for medical detection and diagnosis<sup>5-12</sup>. Specifically, microfabrication technology has had a profound effect on the types of biosensors and has been introduced into biological assays<sup>13-20</sup>. However, microfabrication techniques technically require a cleanroom environment to fabricate devices, which makes it difficult to operate equipment and reduce the price. As a result, PDMS was explored as a replacement for microfabrication, thanks to its advantages such as being low-cost, biocompatible, flexible, and inert elastomers. PDMS is commonly used as the material to fabricate soft microfluidic channels, which can adjust their di-

<sup>a</sup> Electrical and Computer Engineering, Rutgers University-New Brunswick, 94 Brett Road, Piscataway, NJ, USA.

mensions and shapes according to the design of molds. Cured PDMS is then treated by baking or oxide plasma to bond onto substrates and form microfluidic channels. However, bonding issues during the bonding of PDMS to other substrates are the rule of thumb<sup>21–26</sup>. Additionally, PDMS needs time and temperature to cure and reach a solid state. Different molds are also required to fabricate microfluidic channels with various dimensions and shapes to obtain the expected PDMS channels.

In this paper, we present an alternative approach for fabricating biomedical devices using polymer materials and paper<sup>27–29</sup>. These devices have desirable features such as disposability, ease of fabrication, and low risk of mutual contamination, making them ideal for use in the biomedical field<sup>30–33</sup>. Specifically, we describe the design and fabrication of a glucose diagnostic sensor that utilizes colorimetric measurement based on a smartphone camera and computer vision algorithm; making it completely user-friendly and eliminating the need for large and expensive equipment for detection<sup>10,34</sup>. Compared with the methods discussed in Yanyan Xia's review paper<sup>33</sup>, we improved the accuracy of this sensor by utilizing external light sources and implementing background illumination control. These enhancements positively impacted the sensor's performance, in addition to incorporating a 3D-printed portable phone chassis<sup>34</sup>. This approach satisfies the ASSURED criteria (Affordable, Sensitive, Specific, User-friendly, Rapid and Robust, Equipment-free, and Deliverable) established by the WHO<sup>35</sup>. The design and fabrication process of the sensor is simple and easy. AutoCAD was used to design the sensor features, and a laser cutter was used to cut out the features on medical tapes, enabling easy shaping of the channels. The cost of each piece can be lowered to 0.24 USD per sensor, which is more economical than microfabrication and commercial glucose test kits. Compared to other similar approaches, this solution offers greater flexibility for application in a various assays and contexts. The laser-cut channels can be conveniently reshaped to match the dimensions and shapes of the substrates<sup>36,37</sup>. To enhance the flow speed and reaction rate in the channel and avoid solution waste, we used PDMS micropumps to control the volume and speed of the solution directly<sup>37,38</sup>. To process a large number of pictures in a short time automatically<sup>39,40</sup>, we utilized a computer vision method, and the automatic algorithm has been verified to replace the manual traditional measurement methods. This automatic customized detector has potential to be utilized into wide range of assays effortlessly, providing advantages over the phone application and more convenient transferability over the cluster algorithm described in Uddin Jalal and Marzia Hoque Tania's papers<sup>35,36</sup>. The regression model obtained from titration experiments can show the relationship between glucose concentrations and average RGB values<sup>15,38,41</sup>.

## 2 Methods and Materials

### 2.1 Hardware components

The hardware used for the smartphone-based glucose measurement platform includes a lens band (Easy-Macro Lens Band, Easy-Marco, MA, USA)<sup>42</sup>, which can reduce the depth of focus of the smartphone camera and further decrease the size of the

main chassis, especially its height. Two LED diodes are used as illumination sources instead of smartphone flash or other external sources in the purpose of eliminating the external light noise and consistently controlling the illumination magnitude. A customized 3D-printed chassis is used to house all components (weight: 178 g including LEDs, two coin batteries, one battery holder, chassis; size:  $15.4 \times 8.5 \times 5.9$  cm)<sup>38,41,43,44</sup>. An iPhone 12 (Apple, Inc., Cupertino, CA, USA) is utilized as the primary smartphone camera to conduct all experiments. The 3D-printed chassis was designed in AutoCAD and 3D printed using Ultimaker S5 3D printer and Ultimaker Polylacticacid (PLA) as filament material (Ultimaker, Utrecht, Netherlands). It is designed to fit the iPhone 12 (Apple, Inc., Cupertino, CA, USA) and acts as a whole black opaque phone holder, which can block the outside light noise that fluctuates with different environments<sup>34,38,41,43</sup>.

In Fig. 1A, the main chassis is introduced, and a strips-window is designed to slide standard microscope glass slides (AmScope, United Scope LLC., CA, USA) where the glucose sensor is bonded. After absorbing the sucrose solution, glucose sensors are slid into the main chassis through the strip window which is shown in Fig. 1B<sup>38</sup>.




Fig. 1 Smartphone-based glucose detection platform. a) 3D-printed smartphone box. b) Actual image was taken after sliding the microfluidics device into the platform.

### 2.2 Microfluidic device fabrication

For this glucose sensor, there are totally three components consisting of a microfluidic channel, PDMS micropump, and PDMS extension part. The microfluidic channel was created using two medical-grade tapes. The double-sided adhesive tape (3M 9965 Diagnostic Microfluidics tape, 3M, MN, USA) was engraved using a laser cutter (Boss LS1420, MinPower (%) -1: 14, MaxPower (%) -1: 14, Speed (mm/s): 9, Boss Laser, FL, USA) to form the structure of the microfluidic channel. The 2D model was designed on AutoCAD. The glucose test strip (ACCUTEST Urine Reagent Strip, Accutest, CA, USA) was placed into the channel with the help of

a hydrophilic cover (3M 9962 Diagnostic Microfluidics tape, 3M, MN, USA) that was designed on AutoCAD and engraved using a laser cutter (Boss LS1420, MinPower (%) -1: 14, MaxPower (%) -1: 14, Speed(mm/s): 9, Boss Laser, FL, USA) to assist the liquid flow. The actual microfluidic channel is shown in Fig. 2A. The inlet of the channels is located at the end of the device, and the height of the inlet is equal to the height of double-sided adhesive tapes, which is 86  $\mu\text{m}$ .




Fig. 2 Actual pictures of a microfluidic channel and molds for the PDMS components. a) Microfluidic channel. b) PDMS micropumps and 3D-printed molds. c) PDMS extension part, capillary and 3D-printed molds. d) Microfluidic device after assemble of all components.

To accelerate the reaction of glucose strips through faster flow velocity, the PDMS micropump was used as an extremely powerful active solution. The disposable extension part is also clean and convenient for absorbing the solution into the sensor<sup>37,38</sup>. To fabricate PDMS micropumps and extension part, 3D-printed molds were used. However, the rough surface of the 3D-printed molds could potentially interfere with the PDMS curing process. To address this issue, acrylic lacquer (Acrylic lacquer, Rust-oleum) was sprayed onto the surface after 3D printing, forming a thin smooth acrylic layer over the coarse 3D-printed molds surface. This isolates the liquid PDMS from coarse 3D-printed surfaces. The protocol for forming the acrylic layer involves three steps:

first, spray acrylic to cover all surfaces and bake for 30 minutes; second, spray acrylic to cover all surfaces again and bake for 30 minutes; and third, let the mold bake for 60 minutes baking is needed after spraying to ensure the solidity of the acrylic. To fabricate PDMS (Dow chemical company, Midland, MI, USA) structures that can feasibly support the application of micropump and extension part, the ratio of elastomer and curing agent was adjusted from the normal 10:1 to 20:1. This made the PDMS soft enough to press with enough strength to support the mechanical structure. Consequently, the molds, PDMS micropump, and extension part are shown in Fig. 2B and C.

In Fig. S2, different configurations for diameters and heights of PDMS micropumps are shown, with the purpose of analyzing the relationship between theoretical and actual volume of micropumps can suck up and the effect of diameters and heights. For this purpose, theoretical volumes of the micropumps were calculated and compared to the actual volume that the micropumps can introduce to the sensor. Regarding height, three dimensions of micropumps were fabricated with different heights of 1 cm, 0.7 cm, and 0.4 cm, all with a diameter of 2 cm, as shown in Fig. S2A. From Fig. S2C, it is evident that the 0.7 cm micropump was the ideal height for this application. Additionally, micropumps with different diameters of 2 cm, 1.5 cm, and 0.5 cm, all with a height of 0.7 cm were fabricated, as shown in Fig. S2B. The density of dyed DI (Deionized) water was measured by weighing the solution with a known volume to obtain the density, which is 1 g/cm<sup>3</sup>. The glucose sensor was weighed before and after micropumping to calculate the weight difference in terms of the micropumps' capability.

Fig. S2 demonstrates that the effect from diameter is larger than that from height on the actual micropumps' capability. To investigate the actual volume for micropumps, three PDMS micropumps were fabricated utilizing the same molds for each dimension. Three glucose sensors were assembled to introduce the dyed DI water into the devices. The weights were measured before and after sucking water into the device, and the actual volume of micropumps was calculated by dividing weight difference by the density. Nevertheless, the micropump is dysfunctional if the height of micropumps is too small since the structure cannot support the pressing force, which is shown in Fig. S2C, and the actual volume for 0.4 cm height is 0. In Fig. S2C, the ratio of theoretical and actual volume is about 3:1, which is larger than that for micropumps with different diameters. To make the sensor more convenient to use, the extension part was fabricated as shown in Fig. 2C. The glass capillary is utilized as the guidance for the solution, which increases the flexibility of this device. In Fig. 2D, the final biomedical device is shown after assembling all components.

### 2.3 Experimental protocol

The first step to utilize this glucose sensor is to mount PDMS micropump and PDMS extension onto the microfluidic channel. The assembled glucose sensor is shown conceptually in Fig. 3A. To operate and utilize this glucose sensor, the PDMS micropump is pressed to create a negative pressure environment inside the



whole sensor. Next, the glass capillary is immersed into the container for sucrose solution or potential urine as shown in Fig. 3B. In Fig. 3C, the PDMS micropump is thumb-pressed to vent the air before dipping glass capillary into solution. Once the thumb is released, the solution is vacuumed into the channel to react with the glucose test strips. After absorbing enough solution, the device is slid into the phone chassis which as shown in Fig. 3D, and pictures are taken for the next step of colorimetric measurement.




Fig. 3 Conceptual pictures of the principle of a microfluidic channel. a) Microfluidic device. b) Dip the microfluidic device into a container of solution. c) Apply pressure down the micropump to create a negative pressure environment. d) Microfluidic device in 3D printed phone holder after adsorption.

## 2.4 Automatic glucose measurement algorithm

To facilitate the convenience and allow accurate measurement of glucose sensor results in the R, G, and B channels, a combination of 3D printing technology and computer vision was implemented. Frames from the videos taken using an iPhone 12 positioned on the chassis were captured for glucose sensors following the recommended reading time of 30 seconds from the user manual. For automated glucose sensor measurement, an automated computer vision method, shown in Fig. 4, was used as a replacement for traditional manual methods. The computer vision workflow utilized YOLOv4, one of the most mature, accurate, and popular computer vision models available, to detect the target feature in one image and produce the metadata associated with the object which is glucose analysis strips in this application, such as locate target area with the coordinates of a rectangular box. Sample images of glucose analysis strips were picked diversly, manually annotated using Roboflow<sup>45</sup>, and photos are prepared for training the detector, which can locate the glucose sensor from background and thereafter and provide the average RGB values for glucose test

strips. To train the model fits for the photos, 40 sample photos were captured for training and validating. Then the target areas were manually selected, and the information was transferred into digital information, which is the relative YOLOv4 format coordinate in Roboflow. This provides the dataset for customized detector training in Google Colab. The wights. file was exported after training and was implemented into the TensorFlow platform on the desktop, which can automatically detect, crop, and send to the next step to provide measurement results.

To quantitatively measure the colorimetric values, the average RGB value was selected as the parameter to reflect the concentration of glucose for glucose test strips measurement. According to the user manual, the color of reacted analysis strips will shift from light to dark color as the concentration of glucose increases. Thus, the equation is  $\text{avg RGB} = (R+G+B)/3$  which combines the red, green, and blue channels and interprets the inverse proportional relationship between average RGB value and glucose concentrations. This function was implemented using MATLAB.




Fig. 4 Flowcharts illustrate the steps of training and testing the glucose detection algorithm. a) Training flowchart of glucose detector. b) Testing flowchart of glucose detector.

## 3 Results

The comparison between manual (ImageJ) versus automatic (YOLOv4) was shown in Fig. 5. The glucose measurement algorithm was able to automatically select the target area with 100% accuracy, as demonstrated as Glucosestrip 1.00 in the bottom picture. To quantify the relationship between ImageJ and YOLOv4 measurement results, 30 frames were captured from recorded videos with a 10-second interval over a 3-minute period. The plots in Fig. 6 show high coherence for glucose concentrations of 50, 100, 250, 500, 1000 and 2000 mg/dL. Consequently, the computer vision method can be an efficient replacement of the conventional ImageJ method. The reason for deviation of Fig. 6E




Fig. 5 Comparison between ImageJ and glucose detector. a) Manually selected target area through ImageJ software. b) Automatically detected target area by glucose detector.

is because the position of the test strip is off from the horizontal line. Since the YOLOv4 can only square the target area out, it will detect small part of the boundary which affect the results a little.

The titration curve was obtained by plotting the glucose concentrations obtained from the glucometer (ANKOVO TD-4627, Shenzhen, China) against the average RGB values of the glucose sensors obtained from the automatic algorithm. Titration experiments were performed at six concentrations, and the average RGB values were measured at 30 seconds after reacting, which is the recommended reading time. The result shows an exponential relationship between glucose concentrations and the average RGB values, and the function is  $Y=72.143+76.261*\exp(-0.00409X)$  and the R-Square equals to 0.997, as shown in Fig. 7.

4 Statistical Performance

In this section, the main results of the computer vision-related tests are presented statistically. Fig. 8a shows the correlation curve for YOLOv4 and ImageJ. The black curve represents the diagonal line with the formula  $Y = X$ . The regression line close association ( $R\text{-square} = 0.99166$ ) with the artificial diagonal line validates the utility and accuracy of the computer vision method. In Fig. 8b, almost all data points are around the mean line, demonstrating the agreement between these two different methods. However, a few data points are around the -1.96 SD line due to the 1000 mg/dL glucose concentration shown in Fig. 6e. This error can be easily eliminated in the future. The statistical variation of the glucose detector over the mean average values is shown in Fig. 8c. Most percentage changes over average RGB values are lower than 5%. For the 2000 mg/dL concentration, the percentage changes are slightly higher, about 8%. The plot's small variation can prove that this glucose sensor and the auto-




Fig. 6 Correlation between time and measured average RGB values and the comparison between ImageJ and Glucose detector measurements. a) 50 mg/dL glucose concentration. b) 100 mg/dL glucose concentration. c) 250 mg/dL glucose concentration. d) 500 mg/dL glucose concentration. e) 1000 mg/dL glucose concentration. f) 2000 mg/dL glucose concentration.




Fig. 7 Regression curve based on Glucose detector measurements

matic detector are steady and consistent. The disposable glucose sensor, when combined with the automatic measurement method, has shown to be a reliable and accurate tool for measuring glucose levels. Therefore, this approach is a promising candidate for glucose monitoring in various settings.




Fig. 8 Statistical Performance. a) Correlation curve of YOLOv4 and ImageJ. b) The Bland-Altman Plot. c) Variation for replicates of different concentrations.

## 5 Conclusions and Discussion

The outcome presented in this paper illustrates the ability to use disposable tapes, PDMS, and glass slides to explore and fabricate innovative, accurate, and economical type of glucose sensors. Combining the PDMS extension part and micropump can bring advantages to assist the usage of glucose sensors. This glucose sensor is uncomplicated and flexible, even people without experi-

ence can operate perfectly. The price of one glucose sensor can be lowered to 0.24 dollar per piece, which is even lower than the glucose test kit. The retail price for one sensor is \$0.24 USD. Based on bulk production, the estimated price can be \$0.19 USD per sensor for large orders. However, the final price is subject to negotiations with the vendor and may vary based on factors such as shipping costs, taxes, and any other applicable fees. The improvement of control of light environment from 3D printed chassis and LED diodes can provide consistent and unnoisy light source for measurement. The results presented in this paper demonstrate the ability of the YOLOv4 automatic measurement algorithm to replace the traditional manual measurement method (in this paper, is ImageJ). Computer vision can accelerate the measurement process, and in the future, this algorithm will be transferred onto portable electronics through TensorFlow Lite. The results presented in this paper demonstrate the fundamental research and feasibility. One phone model was tested since it describes the methodology and algorithm using the average RGB value, which is independent of the phone models. For future work, ultimately, if and when this device gets used by users of various smartphone models and cameras, each different phone model will require its own custom phone chassis design (like any smartphone product and its respective accessories) and also will need to be recalibrated for its image response, as every camera model differs from another. The scope of this paper was fundamental research and demonstrating feasibility rather than showing universality across all smartphone models. Finally, the regression curve was given by different concentrations of glucose with 0.997 R-Square which demonstrates the confidence of this model.

Scalable manufacturability and flexibility of substrate material are key strengths of our device. By utilizing a laser cutter to create the channels, the device can be easily scaled to fit the dimensions of various substrates, which allows for efficient and cost-effective manufacturing. In addition, the device is compatible with a variety of sterile substrate materials, which provides flexibility for producers to use whatever substrate they have available. Strong transferability is a key strength of our device, thanks to our flexible manufacturing methods. The device can be easily redesigned to fit other assay candidates and smartphone candidates, which provides versatility and adaptability for users and manufacturers. In addition, the use of low-cost assay test strips makes our sensor a cost-effective diagnosis method that can be adapted to various assays. However, it's important to note that the sensor is color sensitive and requires the assay reaction to bring about obvious color changes for detection. Additionally, the recommended reading time for specific assays may need to be recalibrated. These potential challenges should be taken into account when considering the use of our device for different assay candidates. Overall, the combination of scalable manufacturability, strong transferability, and cost-effectiveness of our device makes it a highly adaptable and valuable solution for a range of applications.

## Glucose sensor measurement algorithm

To train the YOLOv4 model used in the test strip analysis workflow (shown in Fig. 4), we analyzed 40 test strips (in the GitHub shown below) that were manually labeled

(using Roboflow [https://roboflow.com]; Annotate—Roboflow) with the positions of each glucose test strip (i.e., each test was represented/encapsulated by an ‘information area’); each area was labeled according to what type of test it captured (i.e., the “group information” of each information area). These annotations were transformed into coordinates and used to train the YOLOv4 detector. The custom detectors were trained in Google colab with a clone of the YOLOv4 darknet from Github (https://www.github.com/AlexeyAB/darknet), implemented with the NVIDIA CUDA GPU acceleration toolkit. After training, we obtained a weights file which includes all information about the custom detector and can predict the target areas with 99–100% accuracy. The areas covering the glucose strips that are extracted from each image are further processed by Tensorflow (https://www.tensorflow.org), which is an open-source software library for machine learning and artificial intelligence, to detect the glucose analysis strips. To utilize the weights files/custom detector in a convenient manner on desktop computers, we created the Tensorflow GPU environment using the CUDA toolkit and the Anaconda environment (https://www.anaconda.com) for processing graphical input. To implement YOLOv4 using TensorFlow, we first converted the weights into the corresponding TensorFlow model files and then ran the custom models on desktop computers. We used computer terminal commands to crop the detected areas and saved these as new images for subsequent steps in image processing. The target areas extracted from each image by the trained YOLOv4 detector (i.e., the glucose strip from background) are imported into the colorimetric measurement software in MATLAB (https://www.mathworks.com/) to generate the average RGB values produced by the workflow.

### Data availability

The glucose strips dataset and workflow are available from <http://github.com/zhuolunmeng/A-Smartphone-Based-Disposable-Microfluidic-Glucose-Sensor-Utilizing-Computer-Vision>

### Author Contributions

Zhuolun Meng: conceptualization, methodology, formal analysis, investigation, data curation, writing – original draft, writing - review & editing, software, validation, visualization. Muhammad Tayyab: validation, writing - review & editing. Zhongtian Lin: validation, writing - review & editing. Hassan Raji: validation, writing - review & editing. Mehdi Javanmard: conceptualization, methodology, funding acquisition, project administration, supervision, writing - review & editing.

### Conflicts of interest

There are no conflicts to declare.

### Acknowledgements

This work was supported in part by the National Science Foundation under Grant 1711165, 1556253 and 1846740.

### Notes and references

1 Blood sugar | blood glucose | diabetes, <https://medlineplus.gov/bloodsugar.html>.

2 Blood glucose (sugar) test: Levels and amp; What They mean, <https://my.clevelandclinic.org/health/diagnostics/12363-blood-glucose-test>.

3 National Diabetes Statistics Report, 2022, <https://www.cdc.gov/diabetes/data/statistics-report/index.html>.

4 B. Zhou, Y. Lu, K. Hajifathalian, J. Benthall, M. Di Cesare, G. Danaei, H. Bixby, M. J. Cowan, M. K. Ali, C. Taddei *et al.*, *The lancet*, 2016, **387**, 1513–1530.

5 T. G. Drummond, M. G. Hill and J. K. Barton, *Nature biotechnology*, 2003, **21**, 1192–1199.

6 A. E. Kamholz, B. H. Weigl, B. A. Finlayson and P. Yager, *Analytical chemistry*, 1999, **71**, 5340–5347.

7 K. S. Kim and J.-K. Park, *Lab on a Chip*, 2005, **5**, 657–664.

8 G. M. Whitesides, *nature*, 2006, **442**, 368–373.

9 P. S. Dittrich, K. Tachikawa and A. Manz, *Analytical chemistry*, 2006, **78**, 3887–3908.

10 D. R. Reyes, D. Iossifidis, P.-A. Auroux and A. Manz, *Analytical chemistry*, 2002, **74**, 2623–2636.

11 S. R. Mahmoodi, P. Xie, D. P. Zachs, E. J. Peterson, R. S. Graham, C. R. Kaiser, H. H. Lim, M. G. Allen and M. Javanmard, *Science Advances*, 2021, **7**, eabf4401.

12 J. Mok, M. N. Mindrinos, R. W. Davis and M. Javanmard, *Proceedings of the National Academy of Sciences*, 2014, **111**, 2110–2115.

13 Z. Huang, X. Li, M. Martins-Green and Y. Liu, *Biomedical microdevices*, 2012, **14**, 873–883.

14 M. E. Piyasena and S. W. Graves, *Lab on a Chip*, 2014, **14**, 1044–1059.

15 D. B. Weibel, W. R. DiLuzio and G. M. Whitesides, *Nature Reviews Microbiology*, 2007, **5**, 209–218.

16 Z. Lin, S.-Y. Lin, P. Xie, C.-Y. Lin, G. M. Rather, J. R. Bertino and M. Javanmard, *Scientific reports*, 2020, **10**, 3015.

17 J. Sui, F. Foflonker, D. Bhattacharya and M. Javanmard, *Scientific reports*, 2020, **10**, 1251.

18 J. Sui, P. Xie, Z. Lin and M. Javanmard, *Talanta*, 2020, **215**, 120791.

19 M. Tayyab, P. Xie, M. A. Sami, H. Raji, Z. Lin, Z. Meng, S. R. Mahmoodi and M. Javanmard, *Scientific reports*, 2022, **12**, 20119.

20 P. Xie, N. Song, W. Shen, M. Allen and M. Javanmard, *Biomedical Microdevices*, 2020, **22**, 1–9.

21 W. W. Y. Chow, K. F. Lei, G. Shi, W. J. Li and Q. Huang, *Smart materials and structures*, 2005, **15**, S112.

22 J. El-Ali, P. K. Sorger and K. F. Jensen, *Nature*, 2006, **442**, 403–411.

23 S. R. Quake and A. Scherer, *science*, 2000, **290**, 1536–1540.

24 A. R. Abate, D. Lee, T. Do, C. Holtze and D. A. Weitz, *Lab on a Chip*, 2008, **8**, 516–518.

25 D. C. Duffy, J. C. McDonald, O. J. Schueller and G. M. Whitesides, *Analytical chemistry*, 1998, **70**, 4974–4984.

26 C. Iliescu, H. Taylor, M. Avram, J. Miao and S. Franssila, *Biomicrofluidics*, 2012, **6**, 016505.



- 27 P. Yager, T. Edwards, E. Fu, K. Helton, K. Nelson, M. R. Tam and B. H. Weigl, *Nature*, 2006, **442**, 412–418.
- 28 H. Yu, Y. Lu, Y.-g. Zhou, F.-b. Wang, F.-y. He and X.-h. Xia, *Lab on a Chip*, 2008, **8**, 1496–1501.
- 29 J. Y. Zhang, J. Do, W. R. Premasiri, L. D. Ziegler and C. M. Klapperich, *Lab on a Chip*, 2010, **10**, 3265–3270.
- 30 H. Cho, J. Kim, C.-W. Jeon and K.-H. Han, *Lab on a Chip*, 2017, **17**, 4113–4123.
- 31 U. Eletxigerra, J. Martinez-Perdiguero and S. Merino, *Sensors and Actuators B: Chemical*, 2015, **221**, 1406–1411.
- 32 G. S. Fiorini and D. T. Chiu, *BioTechniques*, 2005, **38**, 429–446.
- 33 Y. Xia, J. Si and Z. Li, *Biosensors and Bioelectronics*, 2016, **77**, 774–789.
- 34 G. T. Smith, N. Dwork, S. A. Khan, M. Millet, K. Magar, M. Javanmard and A. K. E. Bowden, *Lab on a Chip*, 2016, **16**, 2069–2078.
- 35 M. H. Tania, K. T. Lwin, A. M. Shabut, M. Najlah, J. Chin and M. A. Hossain, *Expert Systems with Applications*, 2020, **139**, 112843.
- 36 U. M. Jalal, G. J. Jin and J. S. Shim, *Analytical chemistry*, 2017, **89**, 13160–13166.
- 37 M. K. Kanakasabapathy, M. Sadasivam, A. Singh, C. Preston, P. Thirumalaraju, M. Venkataraman, C. L. Bormann, M. S. Draz, J. C. Petrozza and H. Shafiee, *Science translational medicine*, 2017, **9**, eaai7863.
- 38 Z. Meng, M. Tayyab, Z. Lin, H. Raji and M. Javanmard, *Sensors*, 2022, **23**, 394.
- 39 S. Jiang, M. Jafari, M. Kharbeche, M. Jalayer and K. N. Al-Khalifa, *IEEE Transactions on Intelligent Transportation Systems*, 2020, **23**, 3169–3179.
- 40 S. Jiang, Y. Zhang, R. Liu, M. Jafari and M. Kharbeche, *IEEE Transactions on Intelligent Transportation Systems*, 2022, **23**, 18237–18252.
- 41 Z. Meng, A. Williams, P. Liau, T. G. Stephens, C. Drury, E. N. Chiles, X. Su, M. Javanmard and D. Bhattacharya, *Scientific reports*, 2022, **12**, 14398.
- 42 Easy-Macro, *Macro: Macro lens for iPhone, Android, or any mobile device*, <https://www.easy-macro.com/>.
- 43 S. Talebian and M. Javanmard, *Talanta*, 2021, **228**, 122244.
- 44 Q. Mei, H. Jing, Y. Li, W. Yisibashaer, J. Chen, B. N. Li and Y. Zhang, *Biosensors and Bioelectronics*, 2016, **75**, 427–432.
- 45 Give your software the power to see objects in images and video, <https://roboflow.com/>.